

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080621\
 Data File : BN015820.D
 Acq On : 06 Aug 2021 23:09
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN080621

Quant Time: Aug 07 04:47:33 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 07 04:36:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.161	0.168	-4.3	112	0.00
3	n-Nitrosodimethylamine	0.379	0.365	3.7	113	0.00
4 S	2-Fluorophenol	1.178	1.176	0.2	114	0.00
5 S	Phenol-d6	1.622	1.620	0.1	115	0.00
6	bis(2-Chloroethyl)ether	1.154	1.188	-2.9	115	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
8 S	Nitrobenzene-d5	0.301	0.284	5.6	117	-0.01
9	Naphthalene	1.049	1.043	0.6	114	-0.01
10	Hexachlorobutadiene	0.191	0.191	0.0	114	0.00
11 SURR	2-Methylnaphthalene-d10	0.614	0.616	-0.3	114	0.00
12	2-Methylnaphthalene	0.712	0.703	1.3	114	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
14 S	2,4,6-Tribromophenol	0.209	0.156	25.4#	95	0.00
15 S	2-Fluorobiphenyl	1.493	1.423	4.7	113	0.00
16	Acenaphthylene	1.863	1.783	4.3	112	0.00
17	Acenaphthene	1.171	1.130	3.5	113	0.00
18	Fluorene	1.421	1.385	2.5	114	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
20	4,6-Dinitro-2-methylphenol	0.031	0.026	16.1	122	-0.01
21	4-Bromophenyl-phenylether	0.214	0.181	15.4	98	0.00
22	Hexachlorobenzene	0.264	0.256	3.0	114	0.00
23	Atrazine	0.219	0.201	8.2	109	0.00
24	Pentachlorophenol	0.100	0.088	12.0	119	0.00
25	Phenanthrene	1.141	1.116	2.2	116	0.00
26	Anthracene	1.108	1.062	4.2	114	0.00
27 SURR	Fluoranthene-d10	1.121	1.100	1.9	116	0.00
28	Fluoranthene	1.277	1.269	0.6	118	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	119	-0.01
30	Pyrene	1.417	1.355	4.4	116	0.00
31 S	Terphenyl-d14	1.233	1.124	8.8	110	0.00
32	Benzo(a)anthracene	1.365	1.299	4.8	117	0.00
33	Chrysene	1.290	1.267	1.8	120	0.00
34	Bis(2-ethylhexyl)phthalate	1.007	0.891	11.5	112	0.00
35 I	Perylene-d12	1.000	1.000	0.0	117	0.00
36	Indeno(1,2,3-cd)pyrene	1.510	1.465	3.0	120	-0.01
37	Benzo(b)fluoranthene	1.401	1.400	0.1	122	0.00
38	Benzo(k)fluoranthene	1.320	1.291	2.2	116	0.00
39 C	Benzo(a)pyrene	1.240	1.203	3.0	119	0.00
40	Dibenzo(a,h)anthracene	1.236	1.206	2.4	120	-0.01
41	Benzo(g,h,i)perylene	1.305	1.260	3.4	118	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN080621