

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080624\
 Data File : BN033275.D
 Acq On : 06 Aug 2024 16:09
 Operator : MA/JU
 Sample : P3458-03
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MLS-15-224-239

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/07/2024
 Supervised By :mohammad ahmed 08/08/2024

Quant Time: Aug 06 17:15:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 05 23:57:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.503	152	969	0.400	ng	0.00
7) Naphthalene-d8	10.244	136	3343	0.400	ng	-0.07
13) Acenaphthene-d10	14.126	164	2195m	0.400	ng	0.00
19) Phenanthrene-d10	16.908	188	5787	0.400	ng	-0.01
29) Chrysene-d12	21.116	240	6667m	0.400	ng	0.00
35) Perylene-d12	23.291	264	7752	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.177	112	331m	0.138	ng	-0.12
5) Phenol-d6	6.838	99	297m	0.090	ng	-0.15
8) Nitrobenzene-d5	8.803	82	867	0.323	ng	-0.12
11) 2-Methylnaphthalene-d10	11.886	152	1611	0.325	ng	-0.04
14) 2,4,6-Tribromophenol	15.691	330	364	0.445	ng	-0.02
15) 2-Fluorobiphenyl	12.757	172	2588	0.345	ng	-0.04
27) Fluoranthene-d10	18.941	212	5803	0.358	ng	0.00
31) Terphenyl-d14	19.526	244	6116	0.531	ng	-0.02

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080624\
Data File : BN033275.D
Acq On : 06 Aug 2024 16:09
Operator : MA/JU
Sample : P3458-03
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MLS-15-224-239

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/07/2024
Supervised By :mohammad ahmed 08/08/2024

Quant Time: Aug 06 17:15:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 05 23:57:48 2024
Response via : Initial Calibration

