

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080624\
 Data File : BN033278.D
 Acq On : 06 Aug 2024 17:59
 Operator : MA/JU
 Sample : P3467-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 919-J-WS-080224

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/07/2024
 Supervised By :mohammad ahmed 08/08/2024

Quant Time: Aug 06 18:32:18 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 05 23:57:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.553	152	968	0.400	ng	0.04
7) Naphthalene-d8	10.287	136	3824	0.400	ng	-0.03
13) Acenaphthene-d10	14.111	164	2488	0.400	ng	-0.01
19) Phenanthrene-d10	16.890	188	6420	0.400	ng	-0.03
29) Chrysene-d12	21.104	240	7473	0.400	ng	-0.02
35) Perylene-d12	23.271	264	8396	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.278	112	458m	0.191	ng	-0.01
5) Phenol-d6	7.019	99	458m	0.139	ng	0.03
8) Nitrobenzene-d5	8.856	82	1050	0.341	ng	-0.06
11) 2-Methylnaphthalene-d10	11.871	152	2050	0.361	ng	-0.06
14) 2,4,6-Tribromophenol	15.661	330	457	0.493	ng	-0.06
15) 2-Fluorobiphenyl	12.743	172	3101	0.364	ng	-0.06
27) Fluoranthene-d10	18.929	212	8257	0.460	ng	-0.02
31) Terphenyl-d14	19.519	244	7156	0.554	ng	-0.03

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080624\
Data File : BN033278.D
Acq On : 06 Aug 2024 17:59
Operator : MA/JU
Sample : P3467-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
919-J-WS-080224

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/07/2024
Supervised By :mohammad ahmed 08/08/2024

Quant Time: Aug 06 18:32:18 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 05 23:57:48 2024
Response via : Initial Calibration

