

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080624\
 Data File : BN033280.D
 Acq On : 06 Aug 2024 19:13
 Operator : MA/JU
 Sample : P3469-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
ClientSampleId :
 MLS-15-513-528

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/08/2024
 Supervised By :mohammad ahmed 08/08/2024

Quant Time: Aug 07 02:04:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 05 23:57:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.568	152	915m	0.400	ng	0.06
7) Naphthalene-d8	10.265	136	3252	0.400	ng	-0.05
13) Acenaphthene-d10	14.126	164	2153m	0.400	ng	0.00
19) Phenanthrene-d10	16.908	188	5462	0.400	ng	-0.01
29) Chrysene-d12	21.124	240	6159	0.400	ng	0.00
35) Perylene-d12	23.291	264	7078	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.184	112	306m	0.135	ng	-0.11
5) Phenol-d6	7.033	99	181	0.058	ng	0.04
8) Nitrobenzene-d5	8.909	82	874m	0.334	ng	-0.01
11) 2-Methylnaphthalene-d10	11.882	152	1632	0.338	ng	-0.05
14) 2,4,6-Tribromophenol	15.691	330	355	0.442	ng	-0.02
15) 2-Fluorobiphenyl	12.757	172	2373	0.322	ng	-0.04
27) Fluoranthene-d10	18.941	212	7067	0.462	ng	0.00
31) Terphenyl-d14	19.526	244	5617	0.528	ng	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080624\
 Data File : BN033280.D
 Acq On : 06 Aug 2024 19:13
 Operator : MA/JU
 Sample : P3469-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MLS-15-513-528

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/08/2024
 Supervised By :mohammad ahmed 08/08/2024

Quant Time: Aug 07 02:04:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 05 23:57:48 2024
 Response via : Initial Calibration

