

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN080720\
 Data File : BN011392.D
 Acq On : 07 Aug 2020 10:40
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 07 11:52:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN072820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 31 12:17:46 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2	1,4-Dioxane	0.563	0.553	1.8	87	0.00
3	n-Nitrosodimethylamine	0.289	0.242	16.3	90	0.00
4 S	2-Fluorophenol	1.051	0.992	5.6	96	0.00
5 S	Phenol-d6	1.278	1.117	12.6	78	0.00
6	bis(2-Chloroethyl)ether	1.064	0.855	19.6	73	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	94	-0.01
8 S	Nitrobenzene-d5	0.357	0.297	16.8	85	0.00
9	Naphthalene	1.186	1.051	11.4	87	-0.01
10	Hexachlorobutadiene	0.287	0.281	2.1	95	-0.01
11 SURR	2-Methylnaphthalene-d10	0.730	0.613	16.0	85	-0.01
12	2-Methylnaphthalene	0.841	0.702	16.5	85	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	77	-0.01
14 S	2,4,6-Tribromophenol	0.185	0.147	20.5	73	0.00
15 S	2-Fluorobiphenyl	1.727	1.631	5.6	84	-0.01
16	Acenaphthylene	2.059	1.781	13.5	82	-0.01
17	Acenaphthene	1.343	1.200	10.6	76	-0.01
18	Fluorene	1.723	1.407	18.3	73	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	71	-0.01
20	4,6-Dinitro-2-methylphenol	0.055	0.029	47.3#	55	0.00
21	4-Bromophenyl-phenylether	0.265	0.228	14.0	73	-0.01
22	Hexachlorobenzene	0.301	0.270	10.3	75	-0.01
23	Atrazine	0.206	0.176	14.6	76	-0.01
24	Pentachlorophenol	0.093	0.083	10.8	86	-0.01
25	Phenanthrene	1.302	1.146	12.0	68	-0.01
26	Anthracene	1.082	0.952	12.0	77	-0.01
27 SURR	Fluoranthene-d10	1.221	1.225	-0.3	89	0.00
28	Fluoranthene	1.480	1.475	0.3	90	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	103	-0.01
30	Pyrene	1.803	1.493	17.2	91	0.00
31 S	Terphenyl-d14	1.211	0.972	19.7	89	0.00
32	Benzo(a)anthracene	1.341	1.233	8.1	97	-0.01
33	Chrysene	1.504	1.474	2.0	101	-0.01
34	Bis(2-ethylhexyl)phthalate	0.719	0.477	33.7#	69	-0.01
35	Indeno(1,2,3-cd)pyrene	1.469	1.762	-19.9	132	-0.02
36 I	Perylene-d12	1.000	1.000	0.0	128	-0.02
37	Benzo(b)fluoranthene	1.595	1.395	12.5	129	-0.01
38	Benzo(k)fluoranthene	1.756	1.463	16.7	114	-0.01
39 C	Benzo(a)pyrene	1.384	1.209	12.6	124	-0.02
40	Dibenzo(a,h)anthracene	1.432	1.347	5.9	139	-0.02
41	Benzo(g,h,i)perylene	1.588	1.494	5.9	132	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN080720\
Data File : BN011392.D
Acq On : 07 Aug 2020 10:40
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 07 11:52:46 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN072820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 31 12:17:46 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0