

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN080720\
 Data File : BN011397.D
 Acq On : 07 Aug 2020 14:38
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 07 15:12:00 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN072820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 31 12:17:46 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	70	-0.02
2	1,4-Dioxane	0.400	0.380	5.0	65	0.00
3	n-Nitrosodimethylamine	0.400	0.281	29.8#	58	0.00
4 S	2-Fluorophenol	0.400	0.342	14.5	67	-0.02
5 S	Phenol-d6	0.400	0.310	22.5	53	0.00
6	bis(2-Chloroethyl)ether	0.400	0.305	23.8	53	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	68	-0.01
8 S	Nitrobenzene-d5	0.400	0.263	34.3#	49	0.00
9	Naphthalene	0.400	0.345	13.8	61	-0.01
10	Hexachlorobutadiene	0.400	0.377	5.8	66	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.314	21.5	58	-0.01
12	2-Methylnaphthalene	0.400	0.298	25.5#	55	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	53	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.319	20.3	51	0.00
15 S	2-Fluorobiphenyl	0.400	0.357	10.8	55	-0.01
16	Acenaphthylene	0.400	0.359	10.3	59	-0.01
17	Acenaphthene	0.400	0.348	13.0	51	-0.01
18	Fluorene	0.400	0.345	13.8	54	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	60	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.327	18.3	45	0.01
21	4-Bromophenyl-phenylether	0.400	0.310	22.5	56	-0.01
22	Hexachlorobenzene	0.400	0.318	20.5	56	-0.01
23	Atrazine	0.400	0.284	29.0#	54	-0.01
24	Pentachlorophenol	0.400	0.290	27.5#	59	-0.01
25	Phenanthrene	0.400	0.361	9.8	59	-0.01
26	Anthracene	0.400	0.344	14.0	63	0.00
27 SURR	Fluoranthene-d10	0.400	0.315	21.3	59	0.00
28	Fluoranthene	0.400	0.312	22.0	59	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	82	-0.01
30	Pyrene	0.400	0.269	32.8#	60	0.00
31 S	Terphenyl-d14	0.400	0.283	29.3#	63	-0.01
32	Benzo(a)anthracene	0.400	0.376	6.0	80	-0.01
33	Chrysene	0.400	0.368	8.0	76	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.274	31.5#	57	-0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.456	-14.0	101	-0.02
36 I	Perylene-d12	0.400	0.400	0.0	92	-0.01
37	Benzo(b)fluoranthene	0.400	0.343	14.2	90	-0.01
38	Benzo(k)fluoranthene	0.400	0.334	16.5	82	-0.01
39 C	Benzo(a)pyrene	0.400	0.349	12.8	89	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.403	-0.8	107	-0.02
41	Benzo(g,h,i)perylene	0.400	0.368	8.0	92	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN080720\
Data File : BN011397.D
Acq On : 07 Aug 2020 14:38
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 07 15:12:00 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN072820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 31 12:17:46 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0