

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080922\
 Data File : BN021088.D
 Acq On : 09 Aug 2022 16:49
 Operator : CG/JU
 Sample : N4069-19
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GBJM3

Quant Time: Aug 09 23:12:52 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 09 23:10:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.788	152	3703	0.400	ng/u1	0.00
4) Naphthalene-d8	10.585	136	12032	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.443	164	7066	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.188	188	14410	0.400	ng/u1	0.00
17) Chrysene-d12	21.395	240	11414	0.400	ng/u1	0.00
23) Perylene-d12	23.721	264	9948	0.400	ng/u1	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.194	96	15757	3.326	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.185	152	4837	0.274	ng/u1	0.00
18) Fluoranthene-d10	19.226	212	12691	0.360	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080922\
Data File : BN021088.D
Acq On : 09 Aug 2022 16:49
Operator : CG/JU
Sample : N4069-19
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GBJM3

Quant Time: Aug 09 23:12:52 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 09 23:10:06 2022
Response via : Initial Calibration

