

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080922\
 Data File : BN021090.D
 Acq On : 09 Aug 2022 18:05
 Operator : CG/JU
 Sample : N4069-15
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GBJL9

Quant Time: Aug 09 23:13:18 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 09 23:10:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.789	152	3433	0.400	ng/ul	0.00
4) Naphthalene-d8	10.584	136	11508	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.442	164	6781	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.192	188	13635	0.400	ng/ul	0.00
17) Chrysene-d12	21.398	240	10600	0.400	ng/ul	0.00
23) Perylene-d12	23.722	264	10004	0.400	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.194	96	17955	4.088	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.190	152	5774	0.342	ng/ul	0.01
18) Fluoranthene-d10	19.230	212	14485	0.442	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.228	88	261	0.058	ng/ul#	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080922\
Data File : BN021090.D
Acq On : 09 Aug 2022 18:05
Operator : CG/JU
Sample : N4069-15
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GBJL9

Quant Time: Aug 09 23:13:18 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 09 23:10:06 2022
Response via : Initial Calibration

