

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080924\
 Data File : BN033320.D
 Acq On : 09 Aug 2024 16:24
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/12/2024
 Supervised By :mohammad ahmed 08/12/2024

Quant Time: Aug 09 16:49:38 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 08 04:39:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.580	152	4387	0.400	ng	0.00
7) Naphthalene-d8	10.349	136	12338	0.400	ng	# 0.00
13) Acenaphthene-d10	14.212	164	6802	0.400	ng	0.00
19) Phenanthrene-d10	16.970	188	16003	0.400	ng	# 0.00
29) Chrysene-d12	21.174	240	13800	0.400	ng	# 0.00
35) Perylene-d12	23.366	264	15453	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.204	112	3954	0.342	ng	0.00
5) Phenol-d6	6.756	99	5333	0.338	ng	0.00
8) Nitrobenzene-d5	8.715	82	3420	0.454	ng	0.00
11) 2-Methylnaphthalene-d10	11.947	152	7152	0.372	ng	0.00
14) 2,4,6-Tribromophenol	15.717	330	955	0.347	ng	0.00
15) 2-Fluorobiphenyl	12.844	172	11692	0.456	ng	0.00
27) Fluoranthene-d10	19.007	212	16209	0.368	ng	0.00
31) Terphenyl-d14	19.629	244	9426	0.396	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.196	88	1948	0.401	ng	98
3) n-Nitrosodimethylamine	3.485	42	2652	0.429	ng	# 88
6) bis(2-Chloroethyl)ether	7.016	93	5336	0.395	ng	98
9) Naphthalene	10.391	128	14052	0.406	ng	98
10) Hexachlorobutadiene	10.701	225	2785	0.438	ng	# 98
12) 2-Methylnaphthalene	12.019	142	9228	0.387	ng	100
16) Acenaphthylene	13.934	152	13081	0.394	ng	100
17) Acenaphthene	14.276	154	8973	0.426	ng	97
18) Fluorene	15.271	166	11829	0.420	ng	100
20) 4,6-Dinitro-2-methylph...	15.345	198	637	0.557	ng	# 37
21) 4-Bromophenyl-phenylether	16.176	248	3944	0.406	ng	# 80
22) Hexachlorobenzene	16.275	284	4532	0.444	ng	95
23) Atrazine	16.449	200	2551	0.318	ng	# 93
24) Pentachlorophenol	16.623	266	1274	0.356	ng	98
25) Phenanthrene	17.008	178	19378	0.418	ng	99
26) Anthracene	17.095	178	16525	0.367	ng	100
28) Fluoranthene	19.039	202	22934	0.393	ng	99
30) Pyrene	19.402	202	23366	0.430	ng	100
32) Benzo(a)anthracene	21.165	228	20108	0.375	ng	99
33) Chrysene	21.210	228	22549	0.435	ng	98
34) Bis(2-ethylhexyl)phtha...	21.139	149	7602	0.274	ng	98
36) Indeno(1,2,3-cd)pyrene	25.545	276	24716	0.366	ng	98
37) Benzo(b)fluoranthene	22.714	252	21752	0.379	ng	97
38) Benzo(k)fluoranthene	22.758	252	24717m	0.443	ng	
39) Benzo(a)pyrene	23.270	252	18053	0.354	ng	98
40) Dibenzo(a,h)anthracene	25.568	278	19648	0.368	ng	98
41) Benzo(g,h,i)perylene	26.199	276	22815	0.393	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080924\
Data File : BN033320.D
Acq On : 09 Aug 2024 16:24
Operator : MA/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/12/2024
Supervised By :mohammad ahmed 08/12/2024

Quant Time: Aug 09 16:49:38 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 08 04:39:01 2024
Response via : Initial Calibration






#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.580 min Scan# 61
Delta R.T. -0.009 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC






Tgt Ion:152 Resp: 438
Ion Ratio Lower Upper
152 100
150 155.6 120.7 181.1
115 63.1 55.8 83.8

Manual Integrations

APPROVED




Reviewed By :Yogesh Patel 08/12/2024


Supervised By :mohammad ahmed 08/12/2024



#2
1,4-Dioxane
Concen: 0.401 ng
RT: 3.196 min Scan# 16
Delta R.T. -0.002 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Tgt Ion: 88 Resp: 1948
Ion Ratio Lower Upper
88 100
43 44.4 37.0 55.6
58 82.5 65.1 97.7






#3
n-Nitrosodimethylamine
Concen: 0.429 ng
RT: 3.485 min Scan# 50
Delta R.T. -0.010 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC






Tgt Ion: 42 Resp: 2652
Ion Ratio Lower Upper
42 100
74 106.1 94.9 142.3
44 0.0 7.0 10.4

Manual Integrations

APPROVED




Reviewed By :Yogesh Patel 08/12/2024


Supervised By :mohammad ahmed 08/12/2024



#4
2-Fluorophenol
Concen: 0.342 ng
RT: 5.204 min Scan# 294
Delta R.T. -0.002 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Tgt Ion:112 Resp: 3954
Ion Ratio Lower Upper
112 100
64 60.9 48.7 73.1
63 32.1 25.4 38.2









#5
Phenol-d6
Concen: 0.338 ng
RT: 6.756 min Scan# 509
Delta R.T. -0.002 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion: 99 Resp: 5338
Ion Ratio Lower Upper
99 100
42 24.6 19.3 28.9
71 33.6 27.6 41.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/12/2024
Supervised By :mohammad ahmed 08/12/2024



#6
bis(2-Chloroethyl)ether
Concen: 0.395 ng
RT: 7.016 min Scan# 545
Delta R.T. -0.009 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Tgt Ion: 93 Resp: 5336
Ion Ratio Lower Upper
93 100
63 83.0 64.5 96.7
95 32.7 26.6 39.8





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.349 min Scan# 915
Delta R.T. -0.002 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Instrument :

BNA_N





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:136 Resp: 12338
Ion Ratio Lower Upper
136 100
137 11.7 9.9 14.9
54 8.7 11.4 17.2
68 7.0 9.5 14.3#




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/12/2024
Supervised By :mohammad ahmed 08/12/2024



#8
Nitrobenzene-d5
Concen: 0.454 ng
RT: 8.715 min Scan# 762
Delta R.T. -0.002 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Tgt Ion: 82 Resp: 3420
Ion Ratio Lower Upper
82 100
128 42.0 29.4 44.2
54 56.4 50.0 75.0









#9
Naphthalene
Concen: 0.406 ng
RT: 10.391 min Scan# 919
Delta R.T. -0.013 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:128 Resp: 1405
Ion Ratio Lower Upper
128 100
129 11.4 10.1 15.1
127 13.7 11.4 17.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/12/2024
Supervised By :mohammad ahmed 08/12/2024



#10
Hexachlorobutadiene
Concen: 0.438 ng
RT: 10.701 min Scan# 948
Delta R.T. -0.002 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24


Tgt Ion:225 Resp: 2785
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.1 51.8 77.6





#11
2-Methylnaphthalene-d10
Concen: 0.372 ng
RT: 11.947 min Scan# 1152
Delta R.T. -0.007 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:152 Resp: 7152
Ion Ratio Lower Upper
152 100
151 21.2 16.7 25.1

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/12/2024
Supervised By :mohammad ahmed 08/12/2024




#12
2-Methylnaphthalene
Concen: 0.387 ng
RT: 12.019 min Scan# 1158
Delta R.T. -0.007 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24




Tgt Ion:142 Resp: 9228
Ion Ratio Lower Upper
142 100
141 88.3 70.5 105.7
115 37.3 30.3 45.5





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.212 min Scan# 1494
Delta R.T. -0.010 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:164 Resp: 680
Ion Ratio Lower Upper
164 100
162 105.5 82.6 124.0
160 56.1 38.2 57.4

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/12/2024
Supervised By :mohammad ahmed 08/12/2024




#14
2,4,6-Tribromophenol
Concen: 0.347 ng
RT: 15.717 min Scan# 1602
Delta R.T. 0.001 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24



Tgt Ion:330 Resp: 955
Ion Ratio Lower Upper
330 100
332 94.6 75.6 113.4
141 45.3 31.0 46.4





#15
2-Fluorobiphenyl
Concen: 0.456 ng
RT: 12.844 min Scan# 1169
Delta R.T. 0.001 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Instrument :

BNA_N

ClientSampleId :





SSTDCCC0.4EC

Tgt Ion:172 Resp: 1169
Ion Ratio Lower Upper
172 100
171 34.0 28.3 42.5
170 23.4 19.4 29.2

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 08/12/2024
Supervised By :mohammad ahmed 08/12/2024



#16
Acenaphthylene
Concen: 0.394 ng
RT: 13.934 min Scan# 1438
Delta R.T. 0.001 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24


Tgt Ion:152 Resp: 13081
Ion Ratio Lower Upper
152 100
151 19.6 15.7 23.5
153 13.1 10.3 15.5





#17
Acenaphthene
Concen: 0.426 ng
RT: 14.276 min Scan# 1470
Delta R.T. -0.010 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:154 Resp: 897.0
Ion Ratio Lower Upper
154 100
153 112.6 89.5 134.3
152 61.0 43.6 65.4

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/12/2024
Supervised By :mohammad ahmed 08/12/2024




#18
Fluorene
Concen: 0.420 ng
RT: 15.271 min Scan# 1563
Delta R.T. -0.010 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24




Tgt Ion:166 Resp: 11829
Ion Ratio Lower Upper
166 100
165 98.2 78.9 118.3
167 13.6 10.6 16.0





#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.970 min Scan# 11
Delta R.T. 0.001 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:188 Resp: 1600
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.8 10.5 15.7




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/12/2024
Supervised By :mohammad ahmed 08/12/2024



#20
4,6-Dinitro-2-methylphenol
Concen: 0.557 ng
RT: 15.345 min Scan# 1570
Delta R.T. 0.001 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24


Tgt Ion:198 Resp: 637
Ion Ratio Lower Upper
198 100
51 83.8 127.9 191.9#
105 56.1 112.3 168.5#





#21
4-Bromophenyl-phenylether
Concen: 0.406 ng
RT: 16.176 min Scan# 1639
Delta R.T. 0.001 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:248 Resp: 3944
Ion Ratio Lower Upper
248 100
250 99.0 74.6 112.0
141 45.4 62.2 93.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/12/2024
Supervised By :mohammad ahmed 08/12/2024



#22
Hexachlorobenzene
Concen: 0.444 ng
RT: 16.275 min Scan# 1647
Delta R.T. -0.012 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24


Tgt Ion:284 Resp: 4532
Ion Ratio Lower Upper
284 100
142 42.4 30.6 45.8
249 32.1 24.6 36.8





#23
Atrazine
Concen: 0.318 ng
RT: 16.449 min Scan# 1661
Delta R.T. -0.012 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:200 Resp: 255
Ion Ratio Lower Upper
200 100
173 30.5 18.9 28.3
215 45.6 38.6 57.8




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/12/2024
Supervised By :mohammad ahmed 08/12/2024



#24
Pentachlorophenol
Concen: 0.356 ng
RT: 16.623 min Scan# 1675
Delta R.T. 0.001 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24


Tgt Ion:266 Resp: 1274
Ion Ratio Lower Upper
266 100
264 60.6 50.3 75.5
268 62.9 51.0 76.4





#25
Phenanthrene
Concen: 0.418 ng
RT: 17.008 min Scan# 1713
Delta R.T. 0.001 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:178 Resp: 19378
Ion Ratio Lower Upper
178 100
176 19.4 15.2 22.8
179 15.3 12.0 18.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/12/2024
Supervised By :mohammad ahmed 08/12/2024



#26
Anthracene
Concen: 0.367 ng
RT: 17.095 min Scan# 1713
Delta R.T. -0.012 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Tgt Ion:178 Resp: 16525
Ion Ratio Lower Upper
178 100
176 18.5 14.8 22.2
179 14.9 12.2 18.4





#27
Fluoranthene-d10
Concen: 0.368 ng
RT: 19.007 min Scan# 1940
Delta R.T. -0.004 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Instrument :

BNA_N

ClientSampleId :





SSTDCCC0.4EC

Tgt Ion:212 Resp: 16209
Ion Ratio Lower Upper
212 100
106 16.3 12.2 18.4
104 9.1 6.9 10.3

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 08/12/2024
Supervised By :mohammad ahmed 08/12/2024



#28
Fluoranthene
Concen: 0.393 ng
RT: 19.039 min Scan# 1941
Delta R.T. -0.004 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Tgt Ion:202 Resp: 22934
Ion Ratio Lower Upper
202 100
101 12.3 9.4 14.0
203 17.1 13.8 20.6






#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.174 min Scan# 2117
Delta R.T. -0.004 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC






Tgt Ion:240 Resp: 13800
Ion Ratio Lower Upper
240 100
120 17.1 9.0 13.6
236 29.2 22.4 33.6

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 08/12/2024



Supervised By :mohammad ahmed 08/12/2024




#30
Pyrene
Concen: 0.430 ng
RT: 19.402 min Scan# 2019
Delta R.T. 0.001 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24



Tgt Ion:202 Resp: 23366
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 17.8 14.2 21.4





#31
 Terphenyl-d14
 Concen: 0.396 ng
 RT: 19.629 min Scan# 2098
 Delta R.T. -0.004 min
 Lab File: BN033320.D
 Acq: 09 Aug 2024 16:24

Instrument :

BNA_N

ClientSampleId :





SSTDCCC0.4EC

Tgt Ion:244 Resp: 9420
 Ion Ratio Lower Upper
 244 100
 212 9.7 7.2 10.8
 122 16.9 12.7 19.1

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 08/12/2024
 Supervised By :mohammad ahmed 08/12/2024



#32
 Benzo(a)anthracene
 Concen: 0.375 ng
 RT: 21.165 min Scan# 2382
 Delta R.T. 0.005 min
 Lab File: BN033320.D
 Acq: 09 Aug 2024 16:24


Tgt Ion:228 Resp: 20108
 Ion Ratio Lower Upper
 228 100
 226 26.9 22.1 33.1
 229 20.1 15.6 23.4





#33
Chrysene
Concen: 0.435 ng
RT: 21.210 min Scan# 2120
Delta R.T. -0.004 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:228 Resp: 22549
Ion Ratio Lower Upper
228 100
226 31.2 23.8 35.8
229 19.4 15.8 23.8

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/12/2024
Supervised By :mohammad ahmed 08/12/2024




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.274 ng
RT: 21.139 min Scan# 2379
Delta R.T. -0.013 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24




Tgt Ion:149 Resp: 7602
Ion Ratio Lower Upper
149 100
167 26.0 21.6 32.4
279 3.0 2.6 4.0





#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.366 min Scan# 21
 Delta R.T. 0.005 min
 Lab File: BN033320.D
 Acq: 09 Aug 2024 16:24




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC



Tgt Ion:264 Resp: 1545
 Ion Ratio Lower Upper
 264 100
 260 26.8 21.4 32.0
 265 35.6 22.4 33.6




Manual Integrations
 APPROVED


Reviewed By :Yogesh Patel 08/12/2024
 Supervised By :mohammad ahmed 08/12/2024



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.366 ng
 RT: 25.545 min Scan# 3593
 Delta R.T. -0.001 min
 Lab File: BN033320.D
 Acq: 09 Aug 2024 16:24


Tgt Ion:276 Resp: 24716
 Ion Ratio Lower Upper
 276 100
 138 29.9 22.7 34.1
 277 24.8 19.8 29.8





#37
Benzo(b)fluoranthene
Concen: 0.379 ng
RT: 22.714 min Scan# 2640
Delta R.T. -0.001 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:252 Resp: 21752
Ion Ratio Lower Upper
252 100
253 23.1 17.8 26.6
125 14.1 10.0 15.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/12/2024
Supervised By :mohammad ahmed 08/12/2024



#38
Benzo(k)fluoranthene
Concen: 0.443 ng m
RT: 22.758 min Scan# 2640
Delta R.T. -0.001 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Tgt Ion:252 Resp: 24717
Ion Ratio Lower Upper
252 100
253 22.4 18.3 27.5
125 13.7 10.2 15.4






#39
Benzo(a)pyrene
Concen: 0.354 ng
RT: 23.270 min Scan# 21
Delta R.T. 0.002 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Instrument :

BN033320.D

ClientSampleId :

SSTDCCC0.4EC






Tgt Ion:252 Resp: 1805
Ion Ratio Lower Upper
252 100
253 23.1 18.2 27.4
125 16.0 11.4 17.2

Manual Integrations

APPROVED




Reviewed By :Yogesh Patel 08/12/2024


Supervised By :mohammad ahmed 08/12/2024



#40
Dibenzo(a,h)anthracene
Concen: 0.368 ng
RT: 25.568 min Scan# 3601
Delta R.T. -0.004 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24


Tgt Ion:278 Resp: 19648
Ion Ratio Lower Upper
278 100
139 21.7 16.3 24.5
279 25.2 19.4 29.2





#41
Benzo(g,h,i)perylene
Concen: 0.393 ng
RT: 26.199 min Scan# 3817
Delta R.T. -0.007 min
Lab File: BN033320.D
Acq: 09 Aug 2024 16:24

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion: 276 Resp: 22815
Ion Ratio Lower Upper
276 100
277 23.8 19.0 28.4
138 27.3 20.8 31.2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/12/2024
Supervised By :mohammad ahmed 08/12/2024

