

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081021\
 Data File : BN015842.D
 Acq On : 10 Aug 2021 17:41
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 10 18:19:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 09 19:13:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	101	0.00
2	1,4-Dioxane	0.400	0.370	7.5	81	0.00
3	n-Nitrosodimethylamine	0.400	0.369	7.8	99	0.00
4 S	2-Fluorophenol	0.400	0.383	4.3	97	0.00
5 S	Phenol-d6	0.400	0.382	4.5	98	0.00
6	bis(2-Chloroethyl)ether	0.400	0.400	0.0	98	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	102	0.00
8 S	Nitrobenzene-d5	0.400	0.366	8.5	107	0.00
9	Naphthalene	0.400	0.382	4.5	100	0.00
10	Hexachlorobutadiene	0.400	0.398	0.5	105	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.393	1.8	102	0.00
12	2-Methylnaphthalene	0.400	0.404	-1.0	107	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.400	0.336	16.0	100	0.00
15 S	2-Fluorobiphenyl	0.400	0.371	7.3	101	0.00
16	Acenaphthylene	0.400	0.352	12.0	97	0.00
17	Acenaphthene	0.400	0.372	7.0	100	0.00
18	Fluorene	0.400	0.371	7.3	99	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	101	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.407	-1.7	95	0.00
21	4-Bromophenyl-phenylether	0.400	0.376	6.0	100	0.00
22	Hexachlorobenzene	0.400	0.391	2.3	104	0.00
23	Atrazine	0.400	0.334	16.5	97	0.00
24	Pentachlorophenol	0.400	0.311	22.3	100	0.00
25	Phenanthrene	0.400	0.373	6.8	99	0.00
26	Anthracene	0.400	0.360	10.0	98	0.01
27 SURR	Fluoranthene-d10	0.400	0.384	4.0	104	0.00
28	Fluoranthene	0.400	0.380	5.0	101	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	109	0.01
30	Pyrene	0.400	0.361	9.8	102	0.00
31 S	Terphenyl-d14	0.400	0.389	2.8	113	0.00
32	Benzo(a)anthracene	0.400	0.380	5.0	108	0.00
33	Chrysene	0.400	0.366	8.5	102	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.299	25.3#	97	0.00
35 I	Perylene-d12	0.400	0.400	0.0	107	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.364	9.0	107	0.00
37	Benzo(b)fluoranthene	0.400	0.376	6.0	108	0.00
38	Benzo(k)fluoranthene	0.400	0.381	4.8	105	0.00
39 C	Benzo(a)pyrene	0.400	0.367	8.3	105	0.00
40	Dibenzo(a,h)anthracene	0.400	0.371	7.3	110	0.00
41	Benzo(g,h,i)perylene	0.400	0.359	10.3	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4