

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081022\
 Data File : BN021130.D
 Acq On : 10 Aug 2022 23:22
 Operator : CG/JU
 Sample : N4128-12
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BGBT9

Quant Time: Aug 11 02:20:49 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 09 23:10:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.788	152	3137	0.400	ng/ul	0.00
4) Naphthalene-d8	10.578	136	10409	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.437	164	6043	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.187	188	11936	0.400	ng/ul	0.00
17) Chrysene-d12	21.392	240	8840	0.400	ng/ul	-0.01
23) Perylene-d12	23.713	264	8392	0.400	ng/ul	#-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.194	96	16695	4.160	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.190	152	4778	0.313	ng/ul	0.01
18) Fluoranthene-d10	19.225	212	12110	0.444	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.232	88	772	0.187	ng/ul#	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081022\
Data File : BN021130.D
Acq On : 10 Aug 2022 23:22
Operator : CG/JU
Sample : N4128-12
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BGBT9

Quant Time: Aug 11 02:20:49 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 09 23:10:06 2022
Response via : Initial Calibration

