

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081022\
 Data File : BN021137.D
 Acq On : 11 Aug 2022 03:42
 Operator : CG/JU
 Sample : N4128-19
 Misc :
 ALS Vial : 30 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 BGC01

Quant Time: Aug 11 04:11:52 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 09 23:10:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.784	152	3293	0.400	ng/ul	0.00
4) Naphthalene-d8	10.580	136	10941	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.438	164	6468	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.189	188	12781	0.400	ng/ul	0.00
17) Chrysene-d12	21.395	240	9593	0.400	ng/ul	0.00
23) Perylene-d12	23.719	264	9131	0.400	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.194	96	17019	4.040	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.186	152	5614	0.350	ng/ul	0.00
18) Fluoranthene-d10	19.226	212	13508	0.456	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.232	88	503	0.116	ng/ul#	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 0.116 ng/ul
 RT: 3.232 min Scan# 59
 Delta R.T. 0.004 min
 Lab File: BN021137.D
 Acq: 11 Aug 2022 03:42

Instrument :
 BNA_N
 ClientSampleId :
 BGC01

Tgt Ion:	88	Resp:	503
Ion	Ratio	Lower	Upper
88	100		
43	29.6	31.4	47.0#
58	23.6	49.0	73.6#

