

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081121\
 Data File : BN015846.D
 Acq On : 11 Aug 2021 09:51
 Operator : CG/JU
 Sample : PB138187BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB138187BL

Quant Time: Aug 11 10:45:56 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 09 19:13:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	9787	0.400	ng	0.00
7) Naphthalene-d8	10.723	136	48022	0.400	ng	0.00
13) Acenaphthene-d10	14.561	164	27280	0.400	ng	0.00
19) Phenanthrene-d10	17.310	188	55128	0.400	ng	0.00
29) Chrysene-d12	21.493	240	57430	0.400	ng	0.00
35) Perylene-d12	23.929	264	53950	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.494	112	9166	0.360	ng	0.00
5) Phenol-d6	7.080	99	12048	0.349	ng	0.00
8) Nitrobenzene-d5	9.076	82	13741	0.365	ng	0.00
11) 2-Methylnaphthalene-d10	12.314	152	32565	0.422	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	3336	0.274	ng	0.00
15) 2-Fluorobiphenyl	13.190	172	42660	0.389	ng	0.00
27) Fluoranthene-d10	19.341	212	73720	0.440	ng	0.00
31) Terphenyl-d14	19.937	244	77464	0.459	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081121\
Data File : BN015846.D
Acq On : 11 Aug 2021 09:51
Operator : CG/JU
Sample : PB138187BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB138187BL

Quant Time: Aug 11 10:45:56 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 09 19:13:08 2021
Response via : Initial Calibration

