

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081121\
 Data File : BN015845.D
 Acq On : 11 Aug 2021 08:49
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 11 10:45:35 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 09 19:13:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	96	0.00
2	1,4-Dioxane	0.400	0.419	-4.7	88	0.00
3	n-Nitrosodimethylamine	0.400	0.381	4.8	98	0.00
4 S	2-Fluorophenol	0.400	0.376	6.0	91	0.00
5 S	Phenol-d6	0.400	0.369	7.8	91	0.00
6	bis(2-Chloroethyl)ether	0.400	0.397	0.8	93	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	97	0.00
8 S	Nitrobenzene-d5	0.400	0.369	7.8	102	0.00
9	Naphthalene	0.400	0.384	4.0	95	0.00
10	Hexachlorobutadiene	0.400	0.404	-1.0	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.393	1.8	97	0.00
12	2-Methylnaphthalene	0.400	0.378	5.5	95	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	95	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.337	15.8	93	0.00
15 S	2-Fluorobiphenyl	0.400	0.377	5.8	95	-0.01
16	Acenaphthylene	0.400	0.357	10.8	91	-0.01
17	Acenaphthene	0.400	0.375	6.3	94	-0.01
18	Fluorene	0.400	0.374	6.5	93	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	95	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.435	-8.7	106	0.00
21	4-Bromophenyl-phenylether	0.400	0.384	4.0	95	0.00
22	Hexachlorobenzene	0.400	0.391	2.3	96	0.00
23	Atrazine	0.400	0.331	17.3	90	0.00
24	Pentachlorophenol	0.400	0.333	16.8	100	0.00
25	Phenanthrene	0.400	0.376	6.0	93	0.00
26	Anthracene	0.400	0.355	11.3	90	0.00
27 SURR	Fluoranthene-d10	0.400	0.378	5.5	95	0.00
28	Fluoranthene	0.400	0.378	5.5	93	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	97	0.00
30	Pyrene	0.400	0.365	8.8	92	0.00
31 S	Terphenyl-d14	0.400	0.378	5.5	98	0.00
32	Benzo(a)anthracene	0.400	0.362	9.5	93	-0.01
33	Chrysene	0.400	0.375	6.3	94	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.325	18.8	95	-0.01
35 I	Perylene-d12	0.400	0.400	0.0	97	-0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.365	8.8	97	-0.01
37	Benzo(b)fluoranthene	0.400	0.372	7.0	96	-0.01
38	Benzo(k)fluoranthene	0.400	0.394	1.5	98	-0.01
39 C	Benzo(a)pyrene	0.400	0.363	9.3	94	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.368	8.0	99	-0.01
41	Benzo(g,h,i)perylene	0.400	0.364	9.0	95	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4