

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081121\
 Data File : BN015852.D
 Acq On : 11 Aug 2021 14:57
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 11 15:26:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 11 12:37:07 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.167	0.151	9.6	78	0.00
3	n-Nitrosodimethylamine	0.408	0.370	9.3	96	0.00
4 S	2-Fluorophenol	1.041	0.989	5.0	94	0.00
5 S	Phenol-d6	1.411	1.344	4.7	96	0.00
6	bis(2-Chloroethyl)ether	1.098	1.086	1.1	95	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.314	0.295	6.1	108	0.00
9	Naphthalene	1.050	1.006	4.2	98	0.00
10	Hexachlorobutadiene	0.271	0.273	-0.7	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.642	0.634	1.2	101	0.00
12	2-Methylnaphthalene	0.729	0.691	5.2	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.179	0.153	14.5	100	0.00
15 S	2-Fluorobiphenyl	1.607	1.479	8.0	99	0.00
16	Acenaphthylene	1.704	1.516	11.0	96	0.00
17	Acenaphthene	1.130	1.048	7.3	98	0.00
18	Fluorene	1.424	1.328	6.7	98	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
20	4,6-Dinitro-2-methylphenol	0.039	0.025	35.9#	96	0.00
21	4-Bromophenyl-phenylether	0.215	0.205	4.7	102	0.00
22	Hexachlorobenzene	0.278	0.268	3.6	103	0.00
23	Atrazine	0.189	0.157	16.9	98	0.00
24	Pentachlorophenol	0.106	0.083	21.7	101	0.00
25	Phenanthrene	1.120	1.057	5.6	101	0.00
26	Anthracene	1.014	0.903	10.9	98	0.01
27 SURR	Fluoranthene-d10	1.215	1.175	3.3	105	0.00
28	Fluoranthene	1.343	1.286	4.2	102	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	111	0.01
30	Pyrene	1.262	1.112	11.9	101	0.00
31 S	Terphenyl-d14	1.175	1.127	4.1	114	0.00
32	Benzo(a)anthracene	1.300	1.219	6.2	109	0.00
33	Chrysene	1.303	1.178	9.6	103	0.00
34	Bis(2-ethylhexyl)phthalate	0.592	0.438	26.0#	98	0.00
35 I	Perylene-d12	1.000	1.000	0.0	110	0.00
36	Indeno(1,2,3-cd)pyrene	1.372	1.242	9.5	109	0.00
37	Benzo(b)fluoranthene	1.398	1.288	7.9	108	0.00
38	Benzo(k)fluoranthene	1.296	1.219	5.9	106	0.00
39 C	Benzo(a)pyrene	1.211	1.109	8.4	107	0.00
40	Dibenzo(a,h)anthracene	1.114	1.024	8.1	111	0.01
41	Benzo(g,h,i)perylene	1.213	1.077	11.2	105	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4