

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081121\
 Data File : BN015852.D
 Acq On : 11 Aug 2021 14:57
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 11 15:26:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 11 12:37:07 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	99	0.00
2	1,4-Dioxane	0.400	0.361	9.8	78	0.00
3	n-Nitrosodimethylamine	0.400	0.363	9.3	96	0.00
4 S	2-Fluorophenol	0.400	0.380	5.0	94	0.00
5 S	Phenol-d6	0.400	0.381	4.8	96	0.00
6	bis(2-Chloroethyl)ether	0.400	0.396	1.0	95	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	100	0.00
8 S	Nitrobenzene-d5	0.400	0.376	6.0	108	0.00
9	Naphthalene	0.400	0.383	4.3	98	0.00
10	Hexachlorobutadiene	0.400	0.402	-0.5	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.395	1.3	101	0.00
12	2-Methylnaphthalene	0.400	0.379	5.3	99	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.400	0.341	14.8	100	0.00
15 S	2-Fluorobiphenyl	0.400	0.368	8.0	99	0.00
16	Acenaphthylene	0.400	0.356	11.0	96	0.00
17	Acenaphthene	0.400	0.371	7.3	98	0.00
18	Fluorene	0.400	0.373	6.8	98	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	102	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.251	37.3#	96	0.00
21	4-Bromophenyl-phenylether	0.400	0.382	4.5	102	0.00
22	Hexachlorobenzene	0.400	0.387	3.3	103	0.00
23	Atrazine	0.400	0.334	16.5	98	0.00
24	Pentachlorophenol	0.400	0.311	22.3	101	0.00
25	Phenanthrene	0.400	0.378	5.5	101	0.00
26	Anthracene	0.400	0.356	11.0	98	0.01
27 SURR	Fluoranthene-d10	0.400	0.387	3.3	105	0.00
28	Fluoranthene	0.400	0.383	4.3	102	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	111	0.01
30	Pyrene	0.400	0.352	12.0	101	0.00
31 S	Terphenyl-d14	0.400	0.384	4.0	114	0.00
32	Benzo(a)anthracene	0.400	0.375	6.3	109	0.00
33	Chrysene	0.400	0.362	9.5	103	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.296	26.0#	98	0.00
35 I	Perylene-d12	0.400	0.400	0.0	110	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.362	9.5	109	0.00
37	Benzo(b)fluoranthene	0.400	0.368	8.0	108	0.00
38	Benzo(k)fluoranthene	0.400	0.376	6.0	106	0.00
39 C	Benzo(a)pyrene	0.400	0.366	8.5	107	0.00
40	Dibenzo(a,h)anthracene	0.400	0.367	8.3	111	0.01
41	Benzo(g,h,i)perylene	0.400	0.355	11.3	105	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4