

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN081219\
 Data File : BN007128.D
 Acq On : 13 Aug 2019 12:55
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 13 13:57:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN081019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 12:12:02 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	155	0.00
2	n-Nitrosodimethylamine	0.400	0.373	6.8	184	0.00
3 S	2-Fluorophenol	0.400	0.426	-6.5	172	0.00
4 S	Phenol-d6	0.400	0.437	-9.2	182	0.00
5	bis(2-Chloroethyl)ether	0.400	0.392	2.0	159	0.00
6 I	Naphthalene-d8	0.400	0.400	0.0	167	0.00
7 S	Nitrobenzene-d5	0.400	0.410	-2.5	187	0.00
8	Naphthalene	0.400	0.376	6.0	168	0.00
9	Hexachlorobutadiene	0.400	0.379	5.3	166	-0.01
10 SURR	2-Methylnaphthalene-d10	0.400	0.370	7.5	167	0.00
11	2-Methylnaphthalene	0.400	0.372	7.0	168	0.00
12 I	Acenaphthene-d10	0.400	0.400	0.0	162	0.00
13 S	2,4,6-Tribromophenol	0.400	0.344	14.0	191	0.00
14 S	2-Fluorobiphenyl	0.400	0.230	42.5#	85	0.00
15	Acenaphthylene	0.400	0.487	-21.7	210	0.00
16	Acenaphthene	0.400	0.385	3.8	167	-0.01
17	Fluorene	0.400	0.352	12.0	172	0.00
18 I	Phenanthrene-d10	0.400	0.400	0.0	146	0.00
19	4-Bromophenyl-phenylether	0.400	0.394	1.5	146	-0.01
20	Hexachlorobenzene	0.400	0.402	-0.5	149	0.00
21	Pentachlorophenol	0.400	0.385	3.8	163	0.00
22	Phenanthrene	0.400	0.390	2.5	148	0.00
23	Anthracene	0.400	0.402	-0.5	158	-0.01
24 SURR	Fluoranthene-d10	0.400	0.357	10.8	137	0.00
25	Fluoranthene	0.400	0.364	9.0	138	0.00
26 I	Chrysene-d12	0.400	0.400	0.0	131	-0.01
27	Pvrene	0.400	0.423	-5.7	143	0.00
28 S	Terphenyl-d14	0.400	0.413	-3.2	140	-0.01
29	Benzo(a)anthracene	0.400	0.400	0.0	134	0.00
30	Chrysene	0.400	0.399	0.3	136	-0.01
31	Indeno(1,2,3-cd)pyrene	0.400	0.429	-7.2	149	-0.01
32 I	Perylene-d12	0.400	0.400	0.0	146	-0.01
33	Benzo(b)fluoranthene	0.400	0.390	2.5	145	-0.01
34	Benzo(k)fluoranthene	0.400	0.373	6.8	136	0.00
35 C	Benzo(a)pyrene	0.400	0.397	0.8	151	-0.01
36	Dibenzo(a,h)anthracene	0.400	0.394	1.5	150	-0.01
37	Benzo(g,h,i)perylene	0.400	0.389	2.8	144	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0