

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN081219\
 Data File : BN007127.D
 Acq On : 13 Aug 2019 12:19
 Operator : HP/JU
 Sample : K4240-03
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PST-035-B244-080619

Quant Time: Aug 13 12:58:48 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN081019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 12:12:02 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.43	152	9372	0.40	ng	0.00
6) Naphthalene-d8	10.19	136	36764	0.40	ng	0.00
12) Acenaphthene-d10	14.08	164	20676	0.40	ng	0.00
18) Phenanthrene-d10	16.82	188	48580	0.40	ng	0.00
26) Chrysene-d12	21.05	240	47654	0.40	ng	-0.01
32) Perylene-d12	23.16	264	52476	0.40	ng	0.00
System Monitoring Compounds						
3) 2-Fluorophenol	5.06	112	8752	0.39	ng	0.00
4) Phenol-d6	6.62	99	11662	0.42	ng	0.00
7) Nitrobenzene-d5	8.56	82	9166	0.31	ng	0.00
10) 2-Methylnaphthalene-d10	11.79	152	20356	0.30	ng	0.00
13) 2,4,6-Tribromophenol	15.57	330	258	0.02	ng	0.00
14) 2-Fluorobiphenyl	12.71	172	5143	0.17	ng	0.00
24) Fluoranthene-d10	18.87	212	50226	0.29	ng	0.00
28) Terphenyl-d14	19.49	244	38494	0.30	ng	-0.01
Target Compounds						Qvalue
15) Acenaphthylene	13.79	152	11453	0.106	ng	96
22) Phenanthrene	16.87	178	3010	0.021	ng	# 90
27) Pyrene	19.26	202	3417	0.020	ng	# 86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN081219\
Data File : BN007127.D
Acq On : 13 Aug 2019 12:19
Operator : HP/JU
Sample : K4240-03
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PST-035-B244-080619

Quant Time: Aug 13 12:58:48 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN081019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 13 12:12:02 2019
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.43 min Scan# 547
Delta R.T. -0.00 min
Lab File: BN007127.D
Acq: 13 Aug 2019 12:19

Instrument :
BNA_N
ClientSampleId :
PST-035-B244-080619

Tot Ion:152 Resp: 9372
Ion Ratio Lower Upper
152 100
150 149.0 121.7 182.5
115 57.9 46.2 69.4




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#3


2-Fluorophenol
Concen: 0.389 ng
RT: 5.06 min Scan# 252
Delta R.T. -0.00 min
Lab File: BN007127.D
Acq: 13 Aug 2019 12:19

Tgt Ion:112 Resp: 8752
Ion Ratio Lower Upper
112 100
64 52.4 42.6 63.8
63 27.7 22.2 33.2



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BNC
Ion 63.00 (62.70 to 63.70): BNC





#4

Phenol-d6

Concen: 0.418 ng

RT: 6.62 min Scan# 446

Delta R.T. 0.01 min

Lab File: BN007127.D

Acq: 13 Aug 2019 12:19

Instrument :

BNA_N

ClientSampleId :

PST-035-B244-080619


Tot Ion: 99 Resp: 11662

Ion Ratio Lower Upper

99 100

42 28.4 15.8 23.6#




71 33.7 27.0 40.4



Abundance Ion 99.00 (98.70 to 99.70): BN007110.D (-439) (-)

Ion 42.00 (41.70 to 42.70): BN007110.D (-439) (-)

Ion 71.00 (70.70 to 71.70): BN007110.D (-439) (-)



#6

Naphthalene-d8

Concen: 0.400 ng

RT: 10.19 min Scan# 813

Delta R.T. -0.00 min

Lab File: BN007127.D

Acq: 13 Aug 2019 12:19

Tgt Ion: 136 Resp: 36764


Ion Ratio Lower Upper

136 100

137 11.1 9.4 14.0

54 6.0 6.6 9.8#

68 5.6 6.0 9.0#






Abundance Ion 136.00 (135.70 to 136.70): BN007110.D (-806) (-)

Ion 137.00 (136.70 to 137.70): BN007110.D (-806) (-)

Ion 54.00 (53.70 to 54.70): BN007110.D (-806) (-)

Ion 68.00 (67.70 to 68.70): BN007110.D (-806) (-)





#7

Nitrobenzene-d5

Concen: 0.309 ng

RT: 8.56 min Scan# 685

Delta R.T. -0.00 min

Lab File: BN007127.D

Acq: 13 Aug 2019 12:19

Instrument :

BNA_N

ClientSampleId :

PST-035-B244-080619


Tot Ion: 82 Resp: 9166

Ion Ratio Lower Upper

82 100

128 50.7 34.6 51.8




54 42.9 36.2 54.4



Abundance Ion 82.00 (81.70 to 82.70): BNC

Ion 128.00 (127.70 to 128.70): BNC

Ion 54.00 (53.70 to 54.70): BNC



#10

2-Methylnaphthalene-d10

Concen: 0.297 ng

RT: 11.79 min Scan# 981

Delta R.T. -0.00 min

Lab File: BN007127.D


Acq: 13 Aug 2019 12:19

Tgt Ion: 152 Resp: 20356

Ion Ratio Lower Upper



152 100


151 19.4 15.5 23.3



Abundance Ion 152.00 (151.70 to 152.70): BNC

Ion 151.00 (150.70 to 151.70): BNC





#12

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.08 min Scan# 1310

Delta R.T. -0.00 min

Lab File: BN007127.D

Acq: 13 Aug 2019 12:19

Instrument :

BNA_N

ClientSampleId :

PST-035-B244-080619


Tot Ion:164 Resp: 20676

Ion Ratio Lower Upper

164 100

162 101.6 82.2 123.4




160 46.6 38.3 57.5



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#13

2,4,6-Tribromophenol

Concen: 0.023 ng

RT: 15.57 min Scan# 1444

Delta R.T. -0.00 min

Lab File: BN007127.D

Acq: 13 Aug 2019 12:19


Tgt Ion:330 Resp: 258

Ion Ratio Lower Upper

330 100

332 0.0 0.0 0.0



141 42.6 33.6 50.4




Abundance Ion 330.00 (329.70 to 330.70): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E





#14

2-Fluorobiphenyl

Concen: 0.171 ng

RT: 12.71 min Scan# 1187

Delta R.T. -0.00 min

Lab File: BN007127.D

Acq: 13 Aug 2019 12:19

Instrument :

BNA_N

ClientSampleId :

PST-035-B244-080619


Tot Ion:172 Resp: 5143

Ion Ratio Lower Upper

172 100

171 32.9 26.3 39.5




170 20.8 17.0 25.4



Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#15

Acenaphthylene

Concen: 0.106 ng

RT: 13.79 min Scan# 1284

Delta R.T. -0.00 min

Lab File: BN007127.D

Acq: 13 Aug 2019 12:19


Tgt Ion:152 Resp: 11453

Ion Ratio Lower Upper

152 100

151 18.5 15.8 23.6



153 10.9 10.4 15.6




Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E





#18

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.82 min Scan# 1560

Delta R.T. -0.00 min

Lab File: BN007127.D

Acq: 13 Aug 2019 12:19

Instrument :

BNA_N

ClientSampleId :

PST-035-B244-080619


Tot Ion:188 Resp: 48580

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0




80 8.8 7.8 11.6



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BNC

Ion 80.00 (79.70 to 80.70): BNC



#22

Phenanthrene

Concen: 0.021 ng

RT: 16.87 min Scan# 1564

Delta R.T. -0.00 min

Lab File: BN007127.D

Acq: 13 Aug 2019 12:19


Tgt Ion:178 Resp: 3010

Ion Ratio Lower Upper

178 100

176 19.3 15.2 22.8



179 24.6 12.3 18.5#




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#24

Fluoranthene-d10

Concen: 0.292 ng

RT: 18.87 min Scan# 1844

Delta R.T. -0.01 min

Lab File: BN007127.D

Acq: 13 Aug 2019 12:19





Instrument :

BNA_N

ClientSampleId :

PST-035-B244-080619

Tot Ion:212	Resp:	50226
Ion Ratio	Lower	Upper
212	100	
106	10.4	8.2 12.2
104	5.8	4.5 6.7



#26

Chrysene-d12

Concen: 0.400 ng




RT: 21.05 min Scan# 2171


Delta R.T. -0.01 min

Lab File: BN007127.D

Acq: 13 Aug 2019 12:19

Tgt Ion:240	Resp:	47654
Ion Ratio	Lower	Upper
240	100	
120	8.7	4.0 6.0#
236	27.8	21.3 31.9









#27
Pvrene
Concen: 0.020 ng
RT: 19.26 min Scan# 1924
Delta R.T. -0.01 min
Lab File: BN007127.D
Acq: 13 Aug 2019 12:19

Instrument :
BNA_N
ClientSampleId :
PST-035-B244-080619

Tot Ion:202 Resp: 3417
Ion Ratio Lower Upper
202 100
200 21.9 16.6 24.8
203 29.3 14.0 21.0#




Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E








#28
Terphenyl-d14
Concen: 0.296 ng
RT: 19.49 min Scan# 1969
Delta R.T. -0.01 min
Lab File: BN007127.D
Acq: 13 Aug 2019 12:19

Tgt Ion:244 Resp: 38494
Ion Ratio Lower Upper
244 100
212 8.1 6.2 9.2
122 10.8 7.4 11.2



Abundance Ion 244.00 (243.70 to 244.70): E
Ion 212.00 (211.70 to 212.70): E
Ion 122.00 (121.70 to 122.70): E





#32
 Pervlene-d12
 Concen: 0.400 ng
 RT: 23.16 min Scan# 2521
 Delta R.T. -0.01 min
 Lab File: BN007127.D
 Acq: 13 Aug 2019 12:19

Instrument :
 BNA_N
 ClientSampleId :
 PST-035-B244-080619

Tot Ion: 264 Resp: 52476
 Ion Ratio Lower Upper
 264 100
 260 25.0 19.8 29.8
 265 33.1 26.2 39.4

