

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN081219\
 Data File : BN007128.D
 Acq On : 13 Aug 2019 12:55
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Aug 13 13:57:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN081019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 12:12:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.43	152	9540	0.40	ng	0.00
6) Naphthalene-d8	10.19	136	36515	0.40	ng	0.00
12) Acenaphthene-d10	14.08	164	20416	0.40	ng	0.00
18) Phenanthrene-d10	16.82	188	47430	0.40	ng	0.00
26) Chrysene-d12	21.05	240	46189	0.40	ng	-0.01
32) Perylene-d12	23.16	264	51529	0.40	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 2-Fluorophenol	5.06	112	9762	0.43	ng	0.00
4) Phenol-d6	6.62	99	12402	0.44	ng	0.00
7) Nitrobenzene-d5	8.56	82	12080	0.41	ng	0.00
10) 2-Methylnaphthalene-d10	11.79	152	25218	0.37	ng	0.00
13) 2,4,6-Tribromophenol	15.57	330	3777	0.34	ng	0.00
14) 2-Fluorobiphenyl	12.71	172	6833	0.23	ng	0.00
24) Fluoranthene-d10	18.87	212	60066	0.36	ng	0.00
28) Terphenyl-d14	19.49	244	51973	0.41	ng	-0.01


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) n-Nitrosodimethylamine	3.36	42	2341	0.373	ng	# 97
5) bis(2-Chloroethyl)ether	6.87	93	9847	0.392	ng	99
8) Naphthalene	10.24	128	38548	0.376	ng	99
9) Hexachlorobutadiene	10.54	225	8277	0.379	ng	# 100
11) 2-Methylnaphthalene	11.87	142	27020	0.372	ng	98
15) Acenaphthylene	13.79	152	52005	0.487	ng	100
16) Acenaphthene	14.13	154	26322	0.385	ng	99
17) Fluorene	15.13	166	36001	0.352	ng	95
19) 4-Bromophenyl-phenylether	16.03	248	11567	0.394	ng	# 84
20) Hexachlorobenzene	16.14	284	11340	0.402	ng	99
21) Pentachlorophenol	16.48	266	4053	0.385	ng	99
22) Phenanthrene	16.87	178	55401	0.390	ng	100
23) Anthracene	16.95	178	52150	0.402	ng	99
25) Fluoranthene	18.90	202	66122	0.364	ng	100
27) Pyrene	19.26	202	68593	0.423	ng	99
29) Benzo(a)anthracene	21.03	228	67576	0.400	ng	93
30) Chrysene	21.08	228	65539	0.399	ng	97
31) Indeno(1,2,3-cd)pyrene	25.24	276	77647	0.429	ng	98
33) Benzo(b)fluoranthene	22.54	252	68990	0.390	ng	98
34) Benzo(k)fluoranthene	22.58	252	65073	0.373	ng	97
35) Benzo(a)pyrene	23.07	252	63525	0.397	ng	98
36) Dibenzo(a,h)anthracene	25.26	278	63025	0.394	ng	# 95
37) Benzo(g,h,i)perylene	25.87	276	63968	0.389	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN081219\
Data File : BN007128.D
Acq On : 13 Aug 2019 12:55
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Aug 13 13:57:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN081019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 13 12:12:02 2019
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.43 min Scan# 547
Delta R.T. -0.00 min
Lab File: BN007128.D
Acq: 13 Aug 2019 12:55

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tot Ion:152 Resp: 9540
Ion Ratio Lower Upper
152 100
150 151.7 121.7 182.5
115 58.1 46.2 69.4




#2

n-Nitrosodimethylamine
Concen: 0.373 ng
RT: 3.36 min Scan# 41
Delta R.T. 0.00 min
Lab File: BN007128.D
Acq: 13 Aug 2019 12:55

Tgt Ion: 42 Resp: 2341
Ion Ratio Lower Upper
42 100
74 135.2 110.6 166.0
44 11.7 0.0 0.0#





#3

2-Fluorophenol

Concen: 0.426 ng

RT: 5.06 min Scan# 252

Delta R.T. -0.00 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC


Tot Ion:112 Resp: 9762

Ion Ratio Lower Upper

112 100

64 52.4 42.6 63.8




63 27.8 22.2 33.2



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): BNO

Ion 63.00 (62.70 to 63.70): BNO



#4

Phenol-d6

Concen: 0.437 ng

RT: 6.62 min Scan# 446

Delta R.T. 0.01 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55


Tgt Ion: 99 Resp: 12402

Ion Ratio Lower Upper

99 100

42 18.6 15.8 23.6



71 34.6 27.0 40.4




Abundance Ion 99.00 (98.70 to 99.70): BNO

Ion 42.00 (41.70 to 42.70): BNO

Ion 71.00 (70.70 to 71.70): BNO





#5

bis(2-Chloroethyl)ether

Concen: 0.392 ng

RT: 6.87 min Scan# 477

Delta R.T. -0.00 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC


Tot Ion: 93 Resp: 9847

Ion Ratio Lower Upper

93 100

63 70.0 55.6 83.4




95 32.8 26.7 40.1



Abundance Ion 93.00 (92.70 to 93.70): BN007110.D (-472) (-)

Ion 63.00 (62.70 to 63.70): BN007110.D (-472) (-)

Ion 95.00 (94.70 to 95.70): BN007110.D (-472) (-)



#6

Naphthalene-d8

Concen: 0.400 ng

RT: 10.19 min Scan# 813

Delta R.T. -0.00 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55

Tgt Ion: 136 Resp: 36515


Ion Ratio Lower Upper

136 100

137 11.1 9.4 14.0

54 6.1 6.6 9.8#

68 5.6 6.0 9.0#






Abundance Ion 136.00 (135.70 to 136.70): BN007110.D (-806) (-)

Ion 137.00 (136.70 to 137.70): BN007110.D (-806) (-)

Ion 54.00 (53.70 to 54.70): BN007110.D (-806) (-)

Ion 68.00 (67.70 to 68.70): BN007110.D (-806) (-)





#7

Nitrobenzene-d5

Concen: 0.410 ng

RT: 8.56 min Scan# 685

Delta R.T. -0.00 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC


Tot Ion: 82 Resp: 12080

Ion Ratio Lower Upper

82 100

128 49.1 34.6 51.8




54 40.9 36.2 54.4



Abundance Ion 82.00 (81.70 to 82.70): BN007128.D (-679) (-)

Ion 128.00 (127.70 to 128.70): BN007128.D (-679) (-)

Ion 54.00 (53.70 to 54.70): BN007128.D (-679) (-)



#8

Naphthalene

Concen: 0.376 ng

RT: 10.24 min Scan# 817

Delta R.T. -0.00 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55


Tgt Ion: 128 Resp: 38548

Ion Ratio Lower Upper

128 100

129 11.6 9.4 14.2



127 13.1 11.0 16.6




Abundance Ion 128.00 (127.70 to 128.70): BN007128.D (-809) (-)

Ion 129.00 (128.70 to 129.70): BN007128.D (-809) (-)

Ion 127.00 (126.70 to 127.70): BN007128.D (-809) (-)









#9
Hexachlorobutadiene
Concen: 0.379 ng
RT: 10.54 min Scan# 841
Delta R.T. -0.01 min
Lab File: BN007128.D
Acq: 13 Aug 2019 12:55

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tot Ion:225 Resp: 8277
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 51.1 76.7




Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E






#10
2-Methylnaphthalene-d10
Concen: 0.370 ng
RT: 11.79 min Scan# 981
Delta R.T. -0.00 min
Lab File: BN007128.D
Acq: 13 Aug 2019 12:55

Tgt Ion:152 Resp: 25218
Ion Ratio Lower Upper
152 100
151 19.1 15.5 23.3



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E





#11

2-Methylnaphthalene

Concen: 0.372 ng

RT: 11.87 min Scan# 998

Delta R.T. -0.00 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC


Tot Ion:142 Resp: 27020

Ion Ratio Lower Upper

142 100

141 87.9 69.0 103.6




115 34.6 27.1 40.7



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#12

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.08 min Scan# 1310

Delta R.T. -0.00 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55


Tgt Ion:164 Resp: 20416

Ion Ratio Lower Upper

164 100

162 99.6 82.2 123.4



160 44.9 38.3 57.5




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#13
 2,4,6-Tribromophenol
 Concen: 0.344 ng
 RT: 15.57 min Scan# 1444
 Delta R.T. -0.00 min
 Lab File: BN007128.D
 Acq: 13 Aug 2019 12:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tot Ion:330 Resp: 3777
 Ion Ratio Lower Upper
 330 100
 332 0.0 0.0 0.0
 141 40.2 33.6 50.4



#14
 2-Fluorobiphenyl
 Concen: 0.230 ng
 RT: 12.71 min Scan# 1187
 Delta R.T. -0.00 min
 Lab File: BN007128.D
 Acq: 13 Aug 2019 12:55

Tgt Ion:172 Resp: 6833
 Ion Ratio Lower Upper
 172 100
 171 32.0 26.3 39.5
 170 19.9 17.0 25.4





#15

Acenaphthylene

Concen: 0.487 ng

RT: 13.79 min Scan# 1284

Delta R.T. -0.00 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55

Instrument :


BNA_N

ClientSampleId :

SSTDCCC0.4EC

Tot Ion:152 Resp: 52005

Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.8	23.6
153	12.9	10.4	15.6






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#16

Acenaphthene

Concen: 0.385 ng

RT: 14.13 min Scan# 1315


Delta R.T. -0.01 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55

Tgt Ion:154 Resp: 26322

Ion	Ratio	Lower	Upper
154	100		
153	112.9	89.5	134.3
152	55.1	44.8	67.2






Abundance

Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E





#17

Fluorene

Concen: 0.352 ng

RT: 15.13 min Scan# 1405

Delta R.T. -0.00 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC


Tot Ion:166 Resp: 36001

Ion Ratio Lower Upper

166 100

165 102.6 77.9 116.9




167 12.8 10.9 16.3



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#18

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.82 min Scan# 1560

Delta R.T. -0.00 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55


Tgt Ion:188 Resp: 47430

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0



80 8.1 7.8 11.6




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BNC

Ion 80.00 (79.70 to 80.70): BNC





#19

4-Bromophenyl-phenylether

Concen: 0.394 ng

RT: 16.03 min Scan# 1485

Delta R.T. -0.01 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC


Tot Ion:248 Resp: 11567

Ion Ratio Lower Upper

248 100

250 91.1 78.9 118.3




141 75.7 43.3 64.9#



Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E



#20

Hexachlorobenzene

Concen: 0.402 ng

RT: 16.14 min Scan# 1496

Delta R.T. -0.00 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55


Tgt Ion:284 Resp: 11340

Ion Ratio Lower Upper

284 100

142 37.6 30.5 45.7



249 33.0 26.7 40.1




Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E





#21

Pentachlorophenol

Concen: 0.385 ng

RT: 16.48 min Scan# 1528

Delta R.T. -0.00 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC


Tot Ion:266 Resp: 4053

Ion Ratio Lower Upper

266 100

264 63.9 52.3 78.5




268 64.5 51.5 77.3



Abundance Ion 266.00 (265.70 to 266.70): E

Ion 264.00 (263.70 to 264.70): E

Ion 268.00 (267.70 to 268.70): E



#22

Phenanthrene

Concen: 0.390 ng

RT: 16.87 min Scan# 1564

Delta R.T. -0.00 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55


Tgt Ion:178 Resp: 55401

Ion Ratio Lower Upper

178 100

176 19.1 15.2 22.8



179 15.5 12.3 18.5




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#23

Anthracene

Concen: 0.402 ng

RT: 16.95 min Scan# 1572

Delta R.T. -0.01 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC


Tot Ion:178 Resp: 52150

Ion Ratio Lower Upper

178 100

176 18.3 14.6 21.8




179 15.1 12.5 18.7



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#24

Fluoranthene-d10

Concen: 0.357 ng

RT: 18.87 min Scan# 1844

Delta R.T. -0.01 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55


Tgt Ion:212 Resp: 60066

Ion Ratio Lower Upper

212 100

106 10.4 8.2 12.2



104 5.7 4.5 6.7




Abundance Ion 212.00 (211.70 to 212.70): E

Ion 106.00 (105.70 to 106.70): E

Ion 104.00 (103.70 to 104.70): E





#25

Fluoranthene

Concen: 0.364 ng

RT: 18.90 min Scan# 1850

Delta R.T. -0.01 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC


Tot Ion:202 Resp: 66122

Ion Ratio Lower Upper

202 100

101 8.9 7.0 10.4




203 17.2 13.8 20.8



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#26

Chrysene-d12

Concen: 0.400 ng

RT: 21.05 min Scan# 2171

Delta R.T. -0.01 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55


Tgt Ion:240 Resp: 46189

Ion Ratio Lower Upper

240 100

120 8.9 4.0 6.0#



236 27.2 21.3 31.9




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#27
Pvrene
Concen: 0.423 ng
RT: 19.26 min Scan# 1924
Delta R.T. -0.01 min
Lab File: BN007128.D
Acq: 13 Aug 2019 12:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tot Ion:202 Resp: 68593
Ion Ratio Lower Upper
202 100
200 20.8 16.6 24.8
203 18.7 14.0 21.0



#28
Terphenyl-d14
Concen: 0.413 ng
RT: 19.49 min Scan# 1969
Delta R.T. -0.01 min
Lab File: BN007128.D
Acq: 13 Aug 2019 12:55

Tgt Ion:244 Resp: 51973
Ion Ratio Lower Upper
244 100
212 7.8 6.2 9.2
122 10.6 7.4 11.2





#29

Benzo(a)anthracene

Concen: 0.400 ng

RT: 21.03 min Scan# 2170

Delta R.T. -0.00 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55

Instrument :





BNA_N

ClientSampleId :

SSTDCCC0.4EC

Tot Ion:228 Resp: 67576

Ion	Ratio	Lower	Upper
228	100		
226	32.0	21.8	32.6
229	20.8	15.6	23.4



#30

Chrysene

Concen: 0.399 ng

RT: 21.08 min Scan# 2174




Delta R.T. -0.01 min


Lab File: BN007128.D

Acq: 13 Aug 2019 12:55

Tgt Ion:228 Resp: 65539

Ion	Ratio	Lower	Upper
228	100		
226	31.8	23.8	35.6
229	19.8	15.7	23.5









#31
Indeno(1,2,3-cd)pyrene
Concen: 0.429 ng
RT: 25.24 min Scan# 3127
Delta R.T. -0.01 min
Lab File: BN007128.D
Acq: 13 Aug 2019 12:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tot Ion:276 Resp: 77647
Ion Ratio Lower Upper
276 100
138 21.8 16.0 24.0
277 24.7 19.9 29.9



#32
Perylene-d12
Concen: 0.400 ng
RT: 23.16 min Scan# 2520
Delta R.T. -0.01 min
Lab File: BN007128.D
Acq: 13 Aug 2019 12:55

Tgt Ion:264 Resp: 51529
Ion Ratio Lower Upper
264 100
260 25.0 19.8 29.8
265 32.7 26.2 39.4









#33
Benzo(b)fluoranthene
Concen: 0.390 ng
RT: 22.54 min Scan# 2338
Delta R.T. -0.01 min
Lab File: BN007128.D
Acq: 13 Aug 2019 12:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tot Ion:252 Resp: 68990
Ion Ratio Lower Upper
252 100
253 23.5 18.2 27.2
125 10.5 7.4 11.0



#34
Benzo(k)fluoranthene
Concen: 0.373 ng
RT: 22.58 min Scan# 2351
Delta R.T. -0.01 min
Lab File: BN007128.D
Acq: 13 Aug 2019 12:55

Tgt Ion:252 Resp: 65073
Ion Ratio Lower Upper
252 100
253 23.7 18.1 27.1
125 10.8 7.4 11.0





#35

Benzo(a)pyrene

Concen: 0.397 ng

RT: 23.07 min Scan# 2493

Delta R.T. -0.01 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC


Tot Ion:252 Resp: 63525

Ion Ratio Lower Upper

252 100

253 23.5 18.5 27.7

125 11.5 8.2 12.4






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#36

Dibenzo(a,h)anthracene

Concen: 0.394 ng

RT: 25.26 min Scan# 3133

Delta R.T. -0.01 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55


Tgt Ion:278 Resp: 63025

Ion Ratio Lower Upper

278 100

139 17.5 11.3 16.9#

279 25.3 19.3 28.9








Abundance

Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#37

Benzo(a,h,i)perylene

Concen: 0.389 ng

RT: 25.87 min Scan# 3312

Delta R.T. -0.01 min

Lab File: BN007128.D

Acq: 13 Aug 2019 12:55

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tot Ion:	276	Resp:	63968
Ion Ratio	Lower	Upper	
276	100		
277	25.5	19.8	29.8
138	19.6	14.6	22.0

