

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN081220\
 Data File : BN011418.D
 Acq On : 12 Aug 2020 13:13
 Operator : CG/JU
 Sample : SSTD3.201
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.201

Quant Time: Aug 12 14:11:11 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN081220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 12 11:51:18 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.45	152	788	0.40	ng/ul	0.00
2) Naphthalene-d8	10.19	136	2225	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.07	164	1179	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.83	188	2876	0.40	ng/ul	0.00
16) Chrysene-d12	21.05	240	3377	0.40	ng/ul	0.00
20) Perylene-d12	23.17	264	2339	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	0.00	152	0d	0.00	ng/ul	
14) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
11) Pentachlorophenol	16.49	266	2251	3.528	ng/ul	Ovalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN081220\
Data File : BN011418.D
Acq On : 12 Aug 2020 13:13
Operator : CG/JU
Sample : SSTD3.201
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.201

Quant Time: Aug 12 14:11:11 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN081220.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 12 11:51:18 2020
Response via : Initial Calibration

