


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081220\
 Data File : BN011412.D
 Acq On : 12 Aug 2020 09:37
 Operator : CG/JU
 Sample : DFTPP64
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP64

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN081220.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Aug 12 14:05:32 2020



AutoFind: Scans 2433, 2434, 2435; Background Corrected with Scan 2427

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	20.2	49330	PASS
68	69	0.00	2	1.1	1004	PASS
69	198	0.00	100	36.1	88154	PASS
70	69	0.00	2	0.6	530	PASS
127	198	10	80	47.0	114557	PASS
197	198	0.00	2	0.6	1384	PASS
198	198	100	100	100.0	243968	PASS
199	198	5	9	7.1	17433	PASS
275	198	10	60	26.5	64704	PASS
365	198	1	100	5.1	12362	PASS
441	198	0.01	100	9.7	23633	PASS
442	198	50	100	64.7	157885	PASS
443	442	15	24	19.6	31018	PASS

Scan 2427 (17.263 min): BN011412.D
DFTPP64

m/z	Abundance
39.90	305.0
44.00	761.0
73.00	647.0
147.10	369.0
207.10	1111.0
208.00	324.0
281.00	415.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

m/z	Abundance
38.00	942.0
39.00	3022.0
40.00	415.0
44.00	491.0
49.10	677.0
50.00	15438.0
51.00	46120.0
52.00	2771.0
54.80	346.0
56.00	1907.0
57.00	2946.0
61.00	809.0
61.90	1402.0
63.00	2470.0
64.00	804.0
64.90	941.0
67.90	1401.0
68.90	80640.0
69.90	415.0
73.00	1066.0
74.00	11119.0
75.00	15228.0
76.10	4674.0
77.00	100376.0
78.00	6989.0
78.90	7382.0
80.00	5569.0
80.90	7667.0
81.90	1807.0
83.00	1211.0
84.90	1195.0
85.90	3090.0
86.90	971.0
88.10	363.0
90.90	1792.0
91.90	2337.0
93.00	15446.0
93.90	1033.0
96.90	401.0
98.00	12442.0
99.00	9309.0
100.00	964.0
101.00	3672.0
102.90	1095.0
103.90	3463.0
104.90	2900.0
106.10	936.0
107.00	27752.0
107.90	4681.0
108.90	1288.0
110.00	49064.0
110.90	7489.0
112.00	1138.0
112.80	808.0
115.90	1754.0
117.00	43824.0
118.00	3797.0
118.80	350.0
120.10	461.0
122.00	2034.0
123.00	3604.0
123.90	1684.0
124.90	1202.0
127.00	92984.0
128.00	7695.0
129.00	42392.0
130.00	3904.0
130.90	706.0
133.00	378.0
133.90	1034.0
135.00	3526.0
135.90	1526.0
137.00	2996.0
138.00	496.0
138.80	337.0
140.10	578.0
141.00	6246.0
142.10	1690.0
142.90	1088.0
144.00	355.0
144.80	462.0
146.00	1143.0
146.90	3407.0
147.90	9032.0
149.00	2252.0
150.00	708.0
151.00	831.0
152.90	1451.0
153.80	1136.0
155.00	2809.0
156.00	5304.0
157.20	877.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

157.80	894.0
158.90	1022.0
160.00	1909.0
161.00	2517.0
161.90	990.0
164.00	354.0
165.00	2253.0
165.90	2083.0
167.00	16720.0
168.00	11226.0
169.00	1648.0
170.00	329.0
170.90	876.0
171.90	1089.0
173.00	1585.0
174.00	2743.0
175.00	6137.0
176.00	1007.0
176.90	2122.0
178.00	664.0
179.00	9315.0
180.00	6772.0
181.00	3446.0
181.90	656.0
182.90	471.0
183.80	852.0
185.00	3975.0
186.00	31664.0
186.90	10662.0
188.00	1451.0
188.90	2070.0
190.10	316.0
191.00	1370.0
192.00	2459.0
193.00	3272.0
194.10	1261.0
196.00	6713.0
197.00	1587.0
197.90	186240.0
199.00	14019.0
200.00	835.0
201.50	509.0
202.90	1454.0
204.00	8623.0
205.00	15495.0
206.00	52768.0
207.00	8510.0
207.90	2628.0
208.90	1092.0
209.90	1024.0
211.00	2230.0
211.90	517.0
213.10	366.0
215.00	810.0
215.90	1288.0
216.90	18200.0
217.90	2416.0
220.20	451.0
221.00	7009.0
221.90	2458.0
222.90	3386.0
224.00	30840.0
225.00	8788.0
226.00	1197.0
227.00	17152.0
228.00	2322.0
229.00	3215.0
229.80	388.0
230.90	1001.0
232.90	321.0
233.90	1237.0
234.90	903.0
235.80	834.0
237.00	1940.0
238.00	593.0
238.90	806.0
239.90	875.0
241.00	682.0
242.00	2101.0
243.00	2024.0
244.00	23904.0
245.00	3108.0
245.90	5764.0
247.00	1036.0
248.90	1449.0
249.90	374.0
253.00	365.0
255.00	119056.0
256.00	17672.0
257.00	2005.0
258.00	8012.0
259.00	1061.0
260.80	316.0
264.90	4068.0
265.90	830.0
270.80	354.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

272.10	319.0
273.00	3358.0
273.90	8327.0
274.90	41632.0
276.00	7141.0
277.00	5458.0
277.80	898.0
281.10	384.0
283.00	715.0
284.00	303.0
285.00	722.0
292.90	902.0
295.00	682.0
295.90	16704.0
297.00	2263.0
303.00	2092.0
304.00	538.0
313.90	1032.0
314.90	2545.0
315.80	779.0
320.80	312.0
322.00	447.0
323.00	3994.0
323.90	547.0
326.90	993.0
327.80	360.0
332.90	552.0
334.10	2548.0
334.90	504.0
341.10	788.0
345.90	1159.0
346.90	354.0
352.00	1689.0
353.00	730.0
354.00	970.0
354.90	564.0
364.90	6619.0
366.10	1494.0
372.00	2050.0
373.00	677.0
382.80	582.0
401.90	732.0
402.90	1706.0
403.90	336.0
421.00	907.0
422.10	746.0
423.00	4187.0
424.00	1957.0
441.00	12172.0
442.00	85368.0
443.00	17784.0
443.90	1740.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

m/z	Abundance
38.00	1151.0
39.00	4895.0
39.90	558.0
44.00	571.0
48.90	525.0
50.00	18096.0
51.00	60480.0
52.00	4084.0
54.90	505.0
56.00	1916.0
57.00	4056.0
61.00	1356.0
62.00	1632.0
63.00	4418.0
63.90	364.0
65.00	1643.0
68.10	2476.0
68.90	107464.0
69.90	598.0
73.00	1603.0
74.00	14058.0
75.00	20680.0
76.10	7567.0
77.00	134464.0
78.00	8662.0
79.00	10555.0
80.00	7763.0
81.00	9316.0
82.00	2603.0
83.00	2311.0
84.90	1550.0
86.00	3077.0
87.00	1309.0
87.80	346.0
91.00	2211.0
92.00	2783.0
92.90	22352.0
94.00	1099.0
94.80	406.0
95.90	884.0
97.00	530.0
98.00	16073.0
99.00	13626.0
99.90	1046.0
101.00	5927.0
102.90	1690.0
104.00	4050.0
104.90	3569.0
106.00	1250.0
107.00	46648.0
107.90	6143.0
109.00	1198.0
110.00	67536.0
111.00	10777.0
111.90	1276.0
113.00	774.0
115.00	367.0
115.80	2776.0
117.00	56712.0
118.00	4517.0
119.00	628.0
119.90	726.0
121.90	3708.0
122.90	4422.0
123.90	2105.0
124.90	1857.0
125.90	532.0
126.10	533.0
127.00	141376.0
128.00	11622.0
129.00	62384.0
130.00	5487.0
130.90	904.0
132.00	530.0
132.90	408.0
134.00	1729.0
134.90	4721.0
135.90	2119.0
137.00	3401.0
138.00	690.0
139.00	544.0
140.00	1445.0
140.90	9734.0
142.00	2464.0
142.90	1524.0
143.80	371.0
144.10	402.0
145.00	470.0
146.00	1995.0
146.90	4961.0
148.00	14536.0
148.90	2999.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

149.90	857.0
150.90	1036.0
151.80	589.0
152.80	2838.0
153.90	1750.0
155.00	4916.0
156.00	7162.0
157.00	1127.0
157.90	1969.0
159.10	1144.0
160.00	2437.0
161.00	5100.0
162.00	1528.0
163.10	545.0
163.90	583.0
165.00	3710.0
166.00	2386.0
167.00	23152.0
168.00	15874.0
169.00	2264.0
170.00	726.0
170.70	400.0
171.10	558.0
172.00	2276.0
173.00	1914.0
174.00	4169.0
174.90	7755.0
176.00	1817.0
177.00	3690.0
178.00	1097.0
178.90	15184.0
180.00	9946.0
180.90	4968.0
182.00	863.0
182.80	537.0
183.80	963.0
185.00	6059.0
186.00	47288.0
187.00	14264.0
187.90	1522.0
189.00	4063.0
189.80	565.0
190.90	1373.0
191.90	4935.0
192.90	6872.0
194.00	829.0
194.50	357.0
195.10	412.0
196.00	10341.0
196.80	2567.0
197.90	295104.0
198.90	20832.0
200.00	1848.0
201.40	1238.0
202.90	3333.0
204.00	14607.0
205.00	24224.0
206.00	91736.0
207.00	12880.0
207.90	4064.0
208.90	1678.0
210.10	2376.0
210.90	4467.0
212.00	752.0
214.90	1225.0
216.00	2486.0
217.00	27696.0
218.00	3560.0
218.90	455.0
219.90	650.0
221.00	11280.0
222.00	3416.0
223.00	5439.0
224.00	51136.0
225.00	14151.0
226.10	2055.0
227.00	28320.0
227.90	3216.0
228.90	4915.0
230.00	337.0
230.90	1856.0
231.80	373.0
232.90	468.0
233.90	2018.0
235.00	1906.0
235.90	1763.0
236.90	2331.0
237.80	449.0
238.90	1180.0
239.90	801.0
241.00	1832.0
242.00	3017.0
242.90	2993.0
244.00	38472.0
245.00	3986.0
246.00	9012.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

246.90	1870.0
247.90	494.0
249.00	1094.0
250.00	412.0
251.00	511.0
252.00	369.0
253.00	1512.0
254.20	2349.0
254.90	193792.0
256.00	27624.0
257.00	2561.0
258.00	14060.0
259.00	2215.0
260.90	349.0
263.90	570.0
264.90	6829.0
266.00	1031.0
269.90	411.0
270.90	512.0
272.00	562.0
273.00	6078.0
274.00	17336.0
275.00	80224.0
276.00	9619.0
276.90	9501.0
277.90	1917.0
278.80	361.0
281.10	431.0
281.80	416.0
283.00	856.0
284.00	739.0
285.00	1441.0
288.90	535.0
289.80	371.0
292.90	1526.0
293.90	537.0
294.90	728.0
296.00	32288.0
296.90	4356.0
301.00	424.0
303.10	3024.0
304.00	1187.0
307.90	436.0
309.00	516.0
310.00	330.0
312.90	338.0
314.00	1400.0
314.90	3634.0
316.00	1984.0
316.90	425.0
321.00	591.0
321.80	415.0
323.00	8415.0
324.00	1681.0
325.90	318.0
327.00	1643.0
327.90	1020.0
332.00	766.0
332.90	756.0
334.00	5070.0
335.00	1207.0
341.00	1036.0
341.90	339.0
346.00	1548.0
346.80	401.0
352.00	2174.0
352.90	1482.0
354.00	2106.0
355.00	714.0
365.00	13748.0
366.00	1427.0
370.00	347.0
370.90	570.0
371.90	3807.0
372.90	1421.0
383.00	786.0
390.00	645.0
390.90	482.0
402.00	1878.0
402.90	1978.0
403.90	706.0
404.90	379.0
420.80	1685.0
421.90	1881.0
422.90	10859.0
424.00	3779.0
425.00	541.0
441.00	28472.0
441.90	190080.0
443.00	39712.0
444.00	3042.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

m/z	Abundance
37.00	332.0
38.00	1208.0
39.00	2932.0
39.90	558.0
44.10	807.0
49.10	489.0
50.00	13272.0
51.00	41392.0
52.00	2150.0
55.00	382.0
55.90	1320.0
57.00	2846.0
60.90	709.0
62.00	1118.0
62.90	2964.0
64.00	597.0
64.90	1123.0
67.80	1611.0
68.90	76360.0
70.00	579.0
73.00	1507.0
74.00	11561.0
74.90	13425.0
76.00	5962.0
77.00	97744.0
78.00	6629.0
78.90	8473.0
79.90	5257.0
81.00	7261.0
81.90	1701.0
82.90	1912.0
85.00	1005.0
85.90	2581.0
86.80	850.0
87.80	303.0
90.90	1512.0
92.00	2140.0
92.90	14065.0
93.90	954.0
95.10	310.0
95.90	640.0
98.00	12611.0
99.00	8886.0
100.00	721.0
101.00	3162.0
102.90	1426.0
103.90	2655.0
105.00	3261.0
106.00	560.0
107.00	33632.0
108.00	4482.0
108.90	1092.0
110.00	52584.0
111.00	9861.0
112.00	1347.0
112.90	393.0
116.00	2415.0
117.00	45336.0
117.90	3520.0
118.90	520.0
119.90	838.0
122.00	2072.0
122.90	4034.0
123.90	1714.0
125.00	1691.0
125.90	773.0
127.00	109312.0
128.00	8467.0
129.00	49344.0
130.00	4457.0
130.90	573.0
132.00	417.0
132.70	474.0
133.80	2105.0
135.00	4340.0
136.00	1667.0
137.00	2611.0
137.90	596.0
138.80	321.0
140.10	551.0
141.00	7131.0
142.00	1786.0
142.90	1947.0
144.00	548.0
145.90	2150.0
147.00	4560.0
148.00	11536.0
148.90	2037.0
150.00	501.0
151.10	1212.0
152.00	451.0
152.90	2075.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

153.90	1600.0
154.90	3570.0
156.00	5291.0
157.00	885.0
157.90	1092.0
159.10	868.0
159.90	2512.0
160.90	2513.0
161.90	1078.0
162.90	326.0
164.10	706.0
164.90	2862.0
166.00	2682.0
167.00	17544.0
168.00	12144.0
169.00	1885.0
169.90	682.0
170.90	369.0
172.00	1458.0
173.10	1360.0
173.90	3570.0
175.00	6263.0
175.80	1062.0
176.90	2524.0
177.90	1563.0
178.90	13068.0
179.90	7624.0
180.90	4143.0
182.00	696.0
184.00	994.0
185.00	6367.0
186.00	39728.0
187.00	11914.0
187.90	935.0
188.90	3191.0
189.90	519.0
191.00	1358.0
192.00	4157.0
193.00	4613.0
194.00	1003.0
195.10	363.0
196.00	7128.0
197.10	2013.0
198.00	250560.0
198.90	17448.0
200.00	1663.0
200.80	599.0
201.40	1563.0
202.00	718.0
203.00	2225.0
204.00	13806.0
205.10	19264.0
206.00	73936.0
207.00	10729.0
208.00	3401.0
208.90	1165.0
210.00	1864.0
211.00	3505.0
214.90	1105.0
216.00	1670.0
216.90	25360.0
217.90	3183.0
219.10	311.0
220.20	671.0
221.00	7580.0
221.90	3081.0
223.00	4753.0
224.00	44240.0
225.00	10957.0
226.00	1467.0
226.90	23880.0
227.90	2481.0
229.00	3956.0
230.10	646.0
231.00	1238.0
231.90	315.0
233.90	1364.0
234.90	1911.0
235.90	1488.0
237.00	2295.0
237.90	393.0
239.00	878.0
240.00	775.0
240.90	1129.0
241.90	2202.0
243.00	2324.0
244.00	35216.0
245.00	4212.0
246.00	8520.0
246.90	1581.0
248.00	582.0
249.00	1253.0
250.90	448.0
252.00	525.0
253.00	921.0
254.00	1821.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

255.00	172608.0
256.00	26064.0
256.90	2176.0
257.90	11570.0
258.90	1565.0
259.90	360.0
261.00	309.0
265.00	4561.0
266.10	872.0
271.00	736.0
271.90	612.0
273.00	4408.0
274.00	13725.0
275.00	72256.0
276.00	10022.0
276.90	7857.0
277.90	1738.0
281.10	398.0
282.90	1325.0
283.90	460.0
285.00	1617.0
289.10	303.0
291.80	439.0
292.90	1735.0
293.80	578.0
296.00	29360.0
297.00	3707.0
297.80	327.0
301.90	411.0
303.00	2072.0
303.90	747.0
308.00	441.0
308.80	403.0
310.10	308.0
314.00	1312.0
314.90	3800.0
316.00	1313.0
317.00	481.0
320.90	681.0
322.00	357.0
323.00	7357.0
324.00	1561.0
326.00	322.0
326.90	1901.0
327.80	827.0
331.90	520.0
333.10	726.0
334.00	5166.0
335.00	1373.0
341.00	918.0
341.90	382.0
345.90	1961.0
347.00	487.0
352.00	1977.0
353.00	1677.0
354.00	2340.0
354.90	855.0
364.90	16720.0
366.00	1858.0
370.90	664.0
372.00	4196.0
373.00	946.0
383.00	1132.0
383.80	457.0
390.00	411.0
400.80	313.0
402.00	1307.0
402.90	2304.0
404.00	932.0
420.90	2122.0
421.90	2034.0
423.00	10871.0
423.90	3612.0
425.00	356.0
441.00	30256.0
442.00	198208.0
443.00	35560.0
443.90	3591.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64