

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081221\
 Data File : BN015854.D
 Acq On : 12 Aug 2021 08:02
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Aug 12 10:37:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 12 10:35:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	8328	0.400	ng	0.00
7) Naphthalene-d8	10.724	136	40005	0.400	ng	0.00
13) Acenaphthene-d10	14.561	164	22344	0.400	ng	0.01
19) Phenanthrene-d10	17.310	188	45030	0.400	ng	# 0.01
29) Chrysene-d12	21.493	240	47347	0.400	ng	# 0.00
35) Perylene-d12	23.923	264	44604	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.494	112	2364	0.109	ng	0.00
5) Phenol-d6	7.080	99	3195	0.109	ng	0.00
8) Nitrobenzene-d5	9.076	82	3555	0.113	ng	0.00
11) 2-Methylnaphthalene-d10	12.314	152	6480	0.101	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	943	0.094	ng	0.00
15) 2-Fluorobiphenyl	13.190	172	10761	0.120	ng	0.01
27) Fluoranthene-d10	19.341	212	13112	0.096	ng	0.00
31) Terphenyl-d14	19.937	244	16181	0.116	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.410	88	292	0.084	ng	# 87
3) n-Nitrosodimethylamine	3.779	42	721	0.085	ng	# 94
6) bis(2-Chloroethyl)ether	7.347	93	2402	0.105	ng	98
9) Naphthalene	10.774	128	11182	0.106	ng	98
10) Hexachlorobutadiene	11.066	225	3001	0.111	ng	# 100
12) 2-Methylnaphthalene	12.385	142	7525	0.103	ng	98
16) Acenaphthylene	14.281	152	8719	0.092	ng	100
17) Acenaphthene	14.624	154	6643	0.105	ng	100
18) Fluorene	15.601	166	8239	0.104	ng	100
20) 4,6-Dinitro-2-methylph...	15.690	198	212	0.048	ng	# 25
21) 4-Bromophenyl-phenylether	16.494	248	2558	0.106	ng	# 86
22) Hexachlorobenzene	16.616	284	3432	0.110	ng	96
23) Atrazine	16.762	200	1763	0.083	ng	# 96
24) Pentachlorophenol	16.957	266	974	0.081	ng	97
25) Phenanthrene	17.346	178	13387	0.106	ng	99
26) Anthracene	17.444	178	10992	0.096	ng	99
28) Fluoranthene	19.371	202	15267	0.101	ng	100
30) Pyrene	19.728	202	15525	0.104	ng	100
32) Benzo(a)anthracene	21.482	228	15094	0.098	ng	98
33) Chrysene	21.535	228	16773	0.109	ng	99
34) Bis(2-ethylhexyl)phtha...	21.408	149	5538	0.079	ng	99
36) Indeno(1,2,3-cd)pyrene	26.438	276	14895	0.097	ng	# 95
37) Benzo(b)fluoranthene	23.231	252	16321	0.105	ng	95
38) Benzo(k)fluoranthene	23.231	252	15916	0.110	ng	# 95
39) Benzo(a)pyrene	23.817	252	13879	0.103	ng	# 89
40) Dibenzo(a,h)anthracene	26.462	278	11948	0.096	ng	# 81
41) Benzo(g,h,i)perylene	27.215	276	13414	0.099	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081221\
 Data File : BN015854.D
 Acq On : 12 Aug 2021 08:02
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Aug 12 10:37:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 12 10:35:26 2021
 Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.928 min Scan# 526
 Delta R.T. 0.009 min
 Lab File: BN015854.D
 Acq: 12 Aug 2021 08:02




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:	152	Resp:	8328
Ion	Ratio	Lower	Upper
152	100		
150	155.5	125.8	188.8
115	65.7	53.6	80.4



1,4-Dioxane
 Concen: 0.084 ng
 RT: 3.410 min Scan# 36
 Delta R.T. -0.000 min
 Lab File: BN015854.D
 Acq: 12 Aug 2021 08:02

Tgt Ion:	88	Resp:	292
Ion	Ratio	Lower	Upper
88	100		
43	53.8	29.4	44.2#
58	92.1	69.3	103.9









n-Nitrosodimethylamine
Concen: 0.085 ng
RT: 3.779 min Scan# 76
Delta R.T. 0.000 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	42	Resp:	721
Ion	Ratio	Lower	Upper
42	100		
74	101.9	86.1	129.1
44	0.0	3.8	5.8#



2-Fluorophenol
Concen: 0.109 ng
RT: 5.494 min Scan# 262
Delta R.T. 0.009 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Tgt Ion:	112	Resp:	2364
Ion	Ratio	Lower	Upper
112	100		
64	66.9	53.3	79.9
63	36.2	28.6	42.8









Phenol-d6
Concen: 0.109 ng
RT: 7.080 min Scan# 434
Delta R.T. 0.009 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:	99	Resp:	3195
Ion	Ratio	Lower	Upper
99	100		
42	23.8	18.7	28.1
71	40.9	32.4	48.6




bis(2-Chloroethyl)ether
Concen: 0.105 ng
RT: 7.347 min Scan# 463
Delta R.T. -0.000 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Tgt Ion:	93	Resp:	2402
Ion	Ratio	Lower	Upper
93	100		
63	81.5	63.8	95.6
95	33.1	26.0	39.0





Abundance Scan 758 (10.724 min): BN015825.D\data.ms (-752) (-)




Naphthalene-d8
Concen: 0.400 ng
RT: 10.724 min Scan# 758
Delta R.T. 0.000 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	136	Resp:	40005
Ion Ratio	Lower	Upper	
136	100		
137	11.1	9.0	13.4
54	7.9	5.4	8.2
68	5.5	3.8	5.8



Abundance Scan 758 (10.724 min): BN015854.D\data.ms (-742) (-)





Abundance Scan 628 (9.076 min): BN015825.D\data.ms (-625) (-)





Nitrobenzene-d5
Concen: 0.113 ng
RT: 9.076 min Scan# 628
Delta R.T. 0.000 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Tgt Ion:	82	Resp:	3555
Ion Ratio	Lower	Upper	
82	100		
128	33.4	29.2	43.8
54	57.3	41.1	61.7



Abundance Scan 628 (9.076 min): BN015854.D\data.ms (-612) (-)









Naphthalene
 Concen: 0.106 ng
 RT: 10.774 min Scan# 762
 Delta R.T. -0.000 min
 Lab File: BN015854.D
 Acq: 12 Aug 2021 08:02




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	128	Resp:	11182
Ion	Ratio	Lower	Upper
128	100		
129	11.9	8.8	13.2
127	13.6	10.5	15.7




Hexachlorobutadiene
 Concen: 0.111 ng
 RT: 11.066 min Scan# 785
 Delta R.T. -0.000 min
 Lab File: BN015854.D
 Acq: 12 Aug 2021 08:02

Tgt Ion:	225	Resp:	3001
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.1	51.4	77.0





Abundance Scan 1001 (12.314 min): BN015825.D\data.ms (-971) (-)




2-Methylnaphthalene-d10
Concen: 0.101 ng
RT: 12.314 min Scan# 1001
Delta R.T. 0.004 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 6480
Ion Ratio Lower Upper
152 100
151 20.6 16.5 24.7



Abundance Scan 1001 (12.314 min): BN015854.D\data.ms (-955) (-)





Abundance Scan 1018 (12.390 min): BN015825.D\data.ms (-1012) (-)




2-Methylnaphthalene
Concen: 0.103 ng
RT: 12.385 min Scan# 1017
Delta R.T. -0.000 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02


Tgt Ion:142 Resp: 7525
Ion Ratio Lower Upper
142 100
141 90.1 70.5 105.7
115 38.7 29.8 44.8



Abundance Scan 1017 (12.385 min): BN015854.D\data.ms (-995) (-)





Abundance Scan 1206 (14.561 min): BN015825.D\data.ms (-1189) (-)




Acenaphthene-d10
Concen: 0.400 ng
RT: 14.561 min Scan# 1206
Delta R.T. 0.013 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion	Ratio	Lower	Upper
164	100		
162	100.1	83.5	125.3
160	43.7	37.7	56.5



Abundance Scan 1206 (14.561 min): BN015854.D\data.ms (-1189) (-)





Abundance Scan 1323 (16.044 min): BN015825.D\data.ms (-1315) (-)





2,4,6-Tribromophenol
Concen: 0.094 ng
RT: 16.044 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Tgt Ion	Ratio	Lower	Upper
330	100		
332	139.8	84.1	126.1#
141	30.5	24.2	36.4



Abundance Scan 1323 (16.044 min): BN015854.D\data.ms (-1315) (-)









Scan 1098 (13.190 min): BN015825.D\data.ms (-1098) (-)
 2-Fluorobiphenyl
 Concen: 0.120 ng
 RT: 13.190 min Scan# 1098
 Delta R.T. 0.013 min
 Lab File: BN015854.D
 Acq: 12 Aug 2021 08:02




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion	Ratio	Lower	Upper
172	100		
171	33.6	28.7	43.1
170	22.3	19.0	28.4



Scan 1184 (14.281 min): BN015825.D\data.ms (-1184) (-)
 Acenaphthylene
 Concen: 0.092 ng
 RT: 14.281 min Scan# 1184
 Delta R.T. 0.013 min
 Lab File: BN015854.D
 Acq: 12 Aug 2021 08:02


Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.4	23.2
153	12.9	10.5	15.7





Abundance Scan 1211 (14.624 min): BN015825.D\data.ms (-1205) (-)

Acenaphthene
Concen: 0.105 ng
RT: 14.624 min Scan# 1211
Delta R.T. 0.013 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion:	154	Resp:	6643
Ion Ratio	Lower	Upper	
154	100		
153	112.0	90.1	135.1
152	54.2	43.5	65.3





Abundance Scan 1211 (14.624 min): BN015854.D\data.ms (-1202) (-)




Abundance Scan 1288 (15.601 min): BN015825.D\data.ms (-1214) (-)

Fluorene
Concen: 0.104 ng
RT: 15.601 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02


Tgt Ion:	166	Resp:	8239
Ion Ratio	Lower	Upper	
166	100		
165	97.5	77.8	116.6
167	13.7	11.0	16.4



Abundance Scan 1288 (15.601 min): BN015854.D\data.ms (-1272) (-)





Abundance Scan 1427 (17.310 min): BN015825.D\data.ms (-1410) (-)




Phenanthrene-d10
Concen: 0.400 ng
RT: 17.310 min Scan# 1427
Delta R.T. 0.012 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	5.7	7.2	10.8#



Abundance Scan 1427 (17.310 min): BN015854.D\data.ms (-1410) (-)





Abundance Scan 1294 (15.677 min): BN015825.D\data.ms (-1290) (-)




4,6-Dinitro-2-methylphenol
Concen: 0.048 ng
RT: 15.690 min Scan# 1295
Delta R.T. 0.013 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Tgt Ion	Ratio	Lower	Upper
198	100		
182	46.1	11.9	17.9#
77	0.0	0.0	0.0




Abundance Scan 1295 (15.690 min): BN015854.D\data.ms (-1286) (-)






Abundance Scan 1360 (16.494 min): BN015825.D\data.ms (-1352) (-)

4-Bromophenyl-phenylether
Concen: 0.106 ng
RT: 16.494 min Scan# 1360
Delta R.T. 0.000 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	248	Resp:	2558
Ion Ratio	Lower	Upper	
248	100		
250	96.2	78.6	118.0
141	79.1	43.4	65.2#






Abundance Scan 1370 (16.616 min): BN015825.D\data.ms (-1362) (-)

Hexachlorobenzene
Concen: 0.110 ng
RT: 16.616 min Scan# 1370
Delta R.T. 0.000 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02




Tgt Ion:	284	Resp:	3432
Ion Ratio	Lower	Upper	
284	100		
142	31.4	22.6	34.0
249	32.8	25.2	37.8





Abundance Scan 1382 (16.762 min): BN015825.D\data.ms (-1323) (-)

Atrazine
Concen: 0.083 ng
RT: 16.762 min Scan# 1382
Delta R.T. 0.000 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion:	200	Resp:	1763
Ion Ratio	Lower	Upper	
200	100		
173	25.4	16.4	24.6#
215	53.3	41.5	62.3





Abundance Scan 1382 (16.762 min): BN015854.D\data.ms (-1341) (-)




Abundance Scan 1398 (16.957 min): BN015825.D\data.ms (-1324) (-)

Pentachlorophenol
Concen: 0.081 ng
RT: 16.957 min Scan# 1398
Delta R.T. -0.000 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02


Tgt Ion:	266	Resp:	974
Ion Ratio	Lower	Upper	
266	100		
264	58.7	50.1	75.1
268	62.7	51.0	76.6



Abundance Scan 1398 (16.957 min): BN015854.D\data.ms (-1390) (-)





Abundance Scan 1430 (17.346 min): BN015825.D\data.ms (-1425) (-)




Phenanthrene
Concen: 0.106 ng
RT: 17.346 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	178	Resp:	13387
Ion Ratio	Lower	Upper	
178	100		
176	18.6	14.6	21.8
179	16.5	12.6	18.8



Abundance Scan 1430 (17.346 min): BN015854.D\data.ms (-1405) (-)





Abundance Scan 1437 (17.431 min): BN015825.D\data.ms (-1426) (-)




Anthracene
Concen: 0.096 ng
RT: 17.444 min Scan# 1438
Delta R.T. 0.012 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02


Tgt Ion:	178	Resp:	10992
Ion Ratio	Lower	Upper	
178	100		
176	18.4	14.3	21.5
179	15.6	12.2	18.4



Abundance Scan 1438 (17.444 min): BN015854.D\data.ms (-1429) (-)





Abundance Scan 1752 (19.341 min): BN015825.D\data.ms (-1730) (-)




Fluoranthene-d10
Concen: 0.096 ng
RT: 19.341 min Scan# 1752
Delta R.T. 0.005 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	212	Resp:	13112
Ion Ratio	Lower	Upper	
212	100		
106	9.2	7.4	11.2
104	5.3	4.3	6.5



Abundance Scan 1752 (19.341 min): BN015854.D\data.ms (-1731) (-)





Abundance Scan 1758 (19.371 min): BN015825.D\data.ms (-1730) (-)




Fluoranthene
Concen: 0.101 ng
RT: 19.371 min Scan# 1758
Delta R.T. 0.010 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Tgt Ion:	202	Resp:	15267
Ion Ratio	Lower	Upper	
202	100		
101	8.6	6.8	10.2
203	17.3	13.8	20.8





Abundance Scan 1758 (19.371 min): BN015854.D\data.ms (-1716) (-)





Abundance Scan 2026 (21.493 min): BN015825.D\data.ms (-2029) (-)

Chrysene-d12
Concen: 0.400 ng
RT: 21.493 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1





Tgt Ion:	240	Resp:	47347
Ion Ratio	Lower	Upper	
240	100		
120	6.9	3.8	5.8#
236	25.8	20.0	30.0





Abundance Scan 1831 (19.733 min): BN015825.D\data.ms (-1830) (-)

Pyrene
Concen: 0.104 ng
RT: 19.728 min Scan# 1830
Delta R.T. 0.000 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02




Tgt Ion:	202	Resp:	15525
Ion Ratio	Lower	Upper	
202	100		
200	20.1	16.2	24.2
203	17.5	14.1	21.1






Abundance Scan 1872 (19.937 min): BN015825.D\data.ms (-1841) (-)

Terphenyl-d14
Concen: 0.116 ng
RT: 19.937 min Scan# 1872
Delta R.T. 0.005 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.9	6.0	9.0
122	8.3	6.5	9.7







Abundance Scan 2025 (21.482 min): BN015825.D\data.ms (-2005) (-)

Benzo(a)anthracene
Concen: 0.098 ng
RT: 21.482 min Scan# 2025
Delta R.T. 0.011 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02




Tgt Ion	Ratio	Lower	Upper
228	100		
226	25.5	21.1	31.7
229	20.3	15.7	23.5








Chrysene
 Concen: 0.109 ng
 RT: 21.535 min Scan# 2030
 Delta R.T. 0.011 min
 Lab File: BN015854.D
 Acq: 12 Aug 2021 08:02


Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1





Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.7	23.4	35.2
229	20.4	15.8	23.6



Bis(2-ethylhexyl)phthalate
 Concen: 0.079 ng
 RT: 21.408 min Scan# 2018
 Delta R.T. 0.000 min
 Lab File: BN015854.D
 Acq: 12 Aug 2021 08:02




Tgt Ion	Ratio	Lower	Upper
149	100		
167	28.3	22.9	34.3
279	6.1	4.4	6.6





Abundance Scan 2556 (23.926 min): BN015825.D\data.ms (-2535) (-)

Perylene-d12
Concen: 0.400 ng
RT: 23.923 min Scan# 2555
Delta R.T. 0.007 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.2	19.1	28.7
265	25.6	19.8	29.8






Abundance Scan 2555 (23.923 min): BN015854.D\data.ms (-2509) (-)




Abundance Scan 3292 (26.445 min): BN015825.D\data.ms (-3236) (-)

Indeno(1,2,3-cd)pyrene
Concen: 0.097 ng
RT: 26.438 min Scan# 3290
Delta R.T. 0.010 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02


Tgt Ion	Ratio	Lower	Upper
276	100		
138	22.1	13.8	20.8#
277	25.6	20.3	30.5



Abundance Scan 3290 (26.438 min): BN015854.D\data.ms (-3258) (-)





Abundance Scan 2340 (23.187 min): BN015825.D\data.ms (-2327) (-)




Benzo(b)fluoranthene
Concen: 0.105 ng
RT: 23.231 min Scan# 2353
Delta R.T. 0.055 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Tgt Ion:	252	Resp:	16321
Ion	Ratio	Lower	Upper
252	100		
253	25.2	18.2	27.2
125	8.9	6.2	9.4


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Abundance Scan 2353 (23.231 min): BN015854.D\data.ms (-2279) (-)





Abundance Scan 2354 (23.235 min): BN015825.D\data.ms (-2328) (-)




Benzo(k)fluoranthene
Concen: 0.110 ng
RT: 23.231 min Scan# 2353
Delta R.T. 0.007 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02


Tgt Ion:	252	Resp:	15916
Ion	Ratio	Lower	Upper
252	100		
253	25.2	18.1	27.1
125	8.9	5.6	8.4#



Abundance Scan 2353 (23.231 min): BN015854.D\data.ms (-2336) (-)





Abundance Scan 2525 (23.820 min): BN015825.D\data.ms (-2509) (-)




Benzo(a)pyrene
Concen: 0.103 ng
RT: 23.817 min Scan# 2524
Delta R.T. 0.007 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	252	Resp:	13879
Ion Ratio	Lower	Upper	
252	100		
253	27.4	18.6	27.8
125	14.9	7.3	10.9#



Abundance Scan 2524 (23.817 min): BN015854.D\data.ms (-2464) (-)





Abundance Scan 3297 (26.462 min): BN015825.D\data.ms (-3274) (-)




Dibenzo(a,h)anthracene
Concen: 0.096 ng
RT: 26.462 min Scan# 3297
Delta R.T. 0.010 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02


Tgt Ion:	278	Resp:	11948
Ion Ratio	Lower	Upper	
278	100		
139	28.6	10.2	15.2#
279	28.2	19.5	29.3



Abundance Scan 3297 (26.462 min): BN015854.D\data.ms (-3265) (-)




Abundance Scan 3516 (27.212 min): BN015825.D\data.ms (-3454) (-)




Benzo(g,h,i)perylene
Concen: 0.099 ng
RT: 27.215 min Scan# 3517
Delta R.T. 0.017 min
Lab File: BN015854.D
Acq: 12 Aug 2021 08:02

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Abundance Scan 3517 (27.215 min): BN015854.D\data.ms



Tgt Ion:	276	Resp:	13414
Ion	Ratio	Lower	Upper
276	100		
277	25.2	19.5	29.3
138	15.6	11.8	17.8



Abundance Scan 3517 (27.215 min): BN015854.D\data.ms (-3454) (-)

