

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081221\
 Data File : BN015858.D
 Acq On : 12 Aug 2021 10:27
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Aug 12 11:11:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 12 10:35:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	8927	0.400	ng	0.00
7) Naphthalene-d8	10.724	136	41608	0.400	ng	0.00
13) Acenaphthene-d10	14.548	164	24199	0.400	ng	0.00
19) Phenanthrene-d10	17.298	188	49165	0.400	ng	0.00
29) Chrysene-d12	21.493	240	57018	0.400	ng	0.00
35) Perylene-d12	23.916	264	53750	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.494	112	35931	1.546	ng	0.00
5) Phenol-d6	7.071	99	49124	1.560	ng	0.00
8) Nitrobenzene-d5	9.076	82	56911	1.744	ng	0.00
11) 2-Methylnaphthalene-d10	12.310	152	110239	1.650	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	18402	1.701	ng	0.00
15) 2-Fluorobiphenyl	13.178	172	146996	1.512	ng	0.00
27) Fluoranthene-d10	19.336	212	257786	1.726	ng	0.00
31) Terphenyl-d14	19.932	244	253774	1.516	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.410	88	5980	1.601	ng	97
3) n-Nitrosodimethylamine	3.761	42	16218	1.780	ng	99
6) bis(2-Chloroethyl)ether	7.347	93	38904	1.588	ng	100
9) Naphthalene	10.774	128	173295	1.586	ng	100
10) Hexachlorobutadiene	11.066	225	47364	1.677	ng	# 100
12) 2-Methylnaphthalene	12.385	142	118939	1.568	ng	100
16) Acenaphthylene	14.269	152	168903	1.639	ng	100
17) Acenaphthene	14.611	154	111296	1.628	ng	97
18) Fluorene	15.601	166	137593	1.597	ng	100
20) 4,6-Dinitro-2-methylph...	15.677	198	8962	1.856	ng	# 82
21) 4-Bromophenyl-phenylether	16.494	248	44014	1.669	ng	99
22) Hexachlorobenzene	16.616	284	56402	1.652	ng	99
23) Atrazine	16.762	200	38625	1.665	ng	99
24) Pentachlorophenol	16.957	266	22339	1.710	ng	100
25) Phenanthrene	17.346	178	219498	1.594	ng	99
26) Anthracene	17.431	178	204196	1.639	ng	100
28) Fluoranthene	19.366	202	275106	1.667	ng	100
30) Pyrene	19.728	202	274590	1.526	ng	100
32) Benzo(a)anthracene	21.472	228	297876	1.607	ng	100
33) Chrysene	21.525	228	287662	1.549	ng	100
34) Bis(2-ethylhexyl)phtha...	21.408	149	129541	1.535	ng	99
36) Indeno(1,2,3-cd)pyrene	26.432	276	304926	1.655	ng	100
37) Benzo(b)fluoranthene	23.180	252	308364	1.641	ng	99
38) Benzo(k)fluoranthene	23.228	252	280438	1.610	ng	99
39) Benzo(a)pyrene	23.810	252	265670	1.633	ng	97
40) Dibenzo(a,h)anthracene	26.456	278	252491	1.686	ng	98
41) Benzo(g,h,i)perylene	27.205	276	258768	1.587	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081221\
 Data File : BN015858.D
 Acq On : 12 Aug 2021 10:27
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Aug 12 11:11:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 12 10:35:26 2021
 Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.928 min Scan# 526
 Delta R.T. 0.009 min
 Lab File: BN015858.D
 Acq: 12 Aug 2021 10:27




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:	152	Resp:	8927
Ion	Ratio	Lower	Upper
152	100		
150	156.2	125.8	188.8
115	64.4	53.6	80.4



1,4-Dioxane
 Concen: 1.601 ng
 RT: 3.410 min Scan# 36
 Delta R.T. -0.000 min
 Lab File: BN015858.D
 Acq: 12 Aug 2021 10:27

Tgt Ion:	88	Resp:	5980
Ion	Ratio	Lower	Upper
88	100		
43	37.5	29.4	44.2
58	83.3	69.3	103.9









n-Nitrosodimethylamine
Concen: 1.780 ng
RT: 3.761 min Scan# 74
Delta R.T. -0.018 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	42	Resp:	16218
Ion	Ratio	Lower	Upper
42	100		
74	107.1	86.1	129.1
44	5.1	3.8	5.8



2-Fluorophenol
Concen: 1.546 ng
RT: 5.494 min Scan# 262
Delta R.T. 0.009 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Tgt Ion:	112	Resp:	35931
Ion	Ratio	Lower	Upper
112	100		
64	67.1	53.3	79.9
63	35.8	28.6	42.8









Phenol-d6
 Concen: 1.560 ng
 RT: 7.071 min Scan# 433
 Delta R.T. -0.000 min
 Lab File: BN015858.D
 Acq: 12 Aug 2021 10:27




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:	99	Resp:	49124
Ion	Ratio	Lower	Upper
99	100		
42	23.0	18.7	28.1
71	40.5	32.4	48.6



bis(2-Chloroethyl)ether
 Concen: 1.588 ng
 RT: 7.347 min Scan# 463
 Delta R.T. -0.000 min
 Lab File: BN015858.D
 Acq: 12 Aug 2021 10:27




Tgt Ion:	93	Resp:	38904
Ion	Ratio	Lower	Upper
93	100		
63	79.4	63.8	95.6
95	32.8	26.0	39.0




Abundance Scan 758 (10.724 min): BN015825.D\data.ms (-757) (-)

Naphthalene-d8
Concen: 0.400 ng
RT: 10.724 min Scan# 758
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6






Tgt Ion:	136	Resp:	41608
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.0	13.4
54	7.2	5.4	8.2
68	5.0	3.8	5.8





Abundance Scan 628 (9.076 min): BN015825.D\data.ms (-625) (-)

Nitrobenzene-d5
Concen: 1.744 ng
RT: 9.076 min Scan# 628
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27



Tgt Ion:	82	Resp:	56911
Ion	Ratio	Lower	Upper
82	100		
128	37.0	29.2	43.8
54	51.9	41.1	61.7









Naphthalene
 Concen: 1.586 ng
 RT: 10.774 min Scan# 762
 Delta R.T. -0.000 min
 Lab File: BN015858.D
 Acq: 12 Aug 2021 10:27




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:	128	Resp:	173295
Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.8	13.2
127	13.1	10.5	15.7




Hexachlorobutadiene
 Concen: 1.677 ng
 RT: 11.066 min Scan# 785
 Delta R.T. -0.000 min
 Lab File: BN015858.D
 Acq: 12 Aug 2021 10:27

Tgt Ion:	225	Resp:	47364
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.3	51.4	77.0






Abundance Scan 1001 (12.314 min): BN015825.D\data.ms (-971) (-)




2-Methylnaphthalene-d10
Concen: 1.650 ng
RT: 12.310 min Scan# 1000
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:152 Resp: 110239
Ion Ratio Lower Upper
152 100
151 20.4 16.5 24.7






Abundance Scan 1018 (12.390 min): BN015825.D\data.ms (-1012) (-)




2-Methylnaphthalene
Concen: 1.568 ng
RT: 12.385 min Scan# 1017
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Tgt Ion:142 Resp: 118939
Ion Ratio Lower Upper
142 100
141 88.1 70.5 105.7
115 37.2 29.8 44.8





Abundance Scan 1206 (14.561 min): BN015825.D\data.ms (-1189) (-)




Acenaphthene-d10
Concen: 0.400 ng
RT: 14.548 min Scan# 1205
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	164	Resp:	24199
Ion	Ratio	Lower	Upper
164	100		
162	104.5	83.5	125.3
160	47.0	37.7	56.5



Abundance Scan 1205 (14.548 min): BN015858.D\data.ms (-1189) (-)





Abundance Scan 1323 (16.044 min): BN015825.D\data.ms (-1315) (-)





2,4,6-Tribromophenol
Concen: 1.701 ng
RT: 16.044 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Tgt Ion:	330	Resp:	18402
Ion	Ratio	Lower	Upper
330	100		
332	97.9	84.1	126.1
141	31.0	24.2	36.4



Abundance Scan 1323 (16.044 min): BN015858.D\data.ms (-1315) (-)









2-Fluorobiphenyl
 Concen: 1.512 ng
 RT: 13.178 min Scan# 1097
 Delta R.T. -0.000 min
 Lab File: BN015858.D
 Acq: 12 Aug 2021 10:27




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:	172	Resp:	146996
Ion	Ratio	Lower	Upper
172	100		
171	35.5	28.7	43.1
170	23.3	19.0	28.4



Acenaphthylene
 Concen: 1.639 ng
 RT: 14.269 min Scan# 1183
 Delta R.T. -0.000 min
 Lab File: BN015858.D
 Acq: 12 Aug 2021 10:27


Tgt Ion:	152	Resp:	168903
Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.4	23.2
153	13.1	10.5	15.7






Abundance Scan 1211 (14.624 min): BN015825.D\data.ms (-1205) (-)

Acenaphthene
Concen: 1.628 ng
RT: 14.611 min Scan# 1210
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6




Tgt Ion	Ratio	Lower	Upper
154	100		
153	110.2	90.1	135.1
152	51.7	43.5	65.3






Abundance Scan 1288 (15.601 min): BN015825.D\data.ms (-1214) (-)

Fluorene
Concen: 1.597 ng
RT: 15.601 min Scan# 1288
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27




Tgt Ion	Ratio	Lower	Upper
166	100		
165	97.4	77.8	116.6
167	13.5	11.0	16.4






Abundance Scan 1427 (17.310 min): BN015825.D\data.ms (-1419) (-)

Phenanthrene-d10
Concen: 0.400 ng
RT: 17.298 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6




Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	9.1	7.2	10.8






Abundance Scan 1294 (15.677 min): BN015825.D\data.ms (-1290) (-)

4,6-Dinitro-2-methylphenol
Concen: 1.856 ng
RT: 15.677 min Scan# 1294
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27




Tgt Ion	Ratio	Lower	Upper
198	100		
182	7.2	11.9	17.9#
77	0.0	0.0	0.0






Abundance Scan 1360 (16.494 min): BN015825.D\data.ms (-1352) (-)

4-Bromophenyl-phenylether
Concen: 1.669 ng
RT: 16.494 min Scan# 1360
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6






Tgt Ion:	248	Resp:	44014
Ion	Ratio	Lower	Upper
248	100		
250	98.9	78.6	118.0
141	55.5	43.4	65.2



Abundance Scan 1370 (16.616 min): BN015825.D\data.ms (-1362) (-)

Hexachlorobenzene
Concen: 1.652 ng
RT: 16.616 min Scan# 1370
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27


Tgt Ion:	284	Resp:	56402
Ion	Ratio	Lower	Upper
284	100		
142	28.2	22.6	34.0
249	32.1	25.2	37.8






Abundance Scan 1382 (16.762 min): BN015825.D\data.ms (-1328) (-)

Atrazine
Concen: 1.665 ng
RT: 16.762 min Scan# 1382
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6




Tgt Ion:	200	Resp:	38625
Ion Ratio	Lower	Upper	
200	100		
173	20.3	16.4	24.6
215	52.3	41.5	62.3






Abundance Scan 1398 (16.957 min): BN015825.D\data.ms (-1321) (-)


Pentachlorophenol
Concen: 1.710 ng
RT: 16.957 min Scan# 1398
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27



Tgt Ion:	266	Resp:	22339
Ion Ratio	Lower	Upper	
266	100		
264	62.6	50.1	75.1
268	63.9	51.0	76.6





Abundance Scan 1430 (17.346 min): BN015825.D\data.ms (-1425) (-)




Phenanthrene
Concen: 1.594 ng
RT: 17.346 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	178	Resp:	219498
Ion	Ratio	Lower	Upper
178	100		
176	18.4	14.6	21.8
179	15.4	12.6	18.8



Abundance Scan 1430 (17.346 min): BN015858.D\data.ms (-1405) (-)





Abundance Scan 1437 (17.431 min): BN015825.D\data.ms (-1426) (-)




Anthracene
Concen: 1.639 ng
RT: 17.431 min Scan# 1437
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27


Tgt Ion:	178	Resp:	204196
Ion	Ratio	Lower	Upper
178	100		
176	17.8	14.3	21.5
179	15.4	12.2	18.4



Abundance Scan 1437 (17.431 min): BN015858.D\data.ms (-1429) (-)






Abundance Scan 1752 (19.341 min): BN015825.D\data.ms (-1731) (-)




Fluoranthene-d10
Concen: 1.726 ng
RT: 19.336 min Scan# 1751
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:	212	Resp:	257786
Ion	Ratio	Lower	Upper
212	100		
106	9.3	7.4	11.2
104	5.3	4.3	6.5






Abundance Scan 1758 (19.371 min): BN015825.D\data.ms (-1731) (-)



Fluoranthene
Concen: 1.667 ng
RT: 19.366 min Scan# 1757
Delta R.T. 0.005 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27


Tgt Ion:	202	Resp:	275106
Ion	Ratio	Lower	Upper
202	100		
101	8.6	6.8	10.2
203	17.5	13.8	20.8





Abundance Scan 2026 (21.493 min): BN015825.D\data.ms (-2029) (-)

Chrysene-d12
Concen: 0.400 ng
RT: 21.493 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27


Instrument :
BNA_N
ClientSampleId :
SSTDIC1.6



Tgt Ion:	240	Resp:	57018
Ion	Ratio	Lower	Upper
240	100		
120	4.7	3.8	5.8
236	25.2	20.0	30.0





Abundance Scan 2026 (21.493 min): BN015858.D\data.ms (-2007) (-)




Abundance Scan 1831 (19.733 min): BN015825.D\data.ms (-1830) (-)

Pyrene
Concen: 1.526 ng
RT: 19.728 min Scan# 1830
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Tgt Ion:	202	Resp:	274590
Ion	Ratio	Lower	Upper
202	100		
200	20.2	16.2	24.2
203	17.6	14.1	21.1




Abundance Scan 1830 (19.728 min): BN015858.D\data.ms (-1790) (-)






Abundance Scan 1872 (19.937 min): BN015825.D\data.ms (-1841) (-)

Terphenyl-d14
Concen: 1.516 ng
RT: 19.932 min Scan# 1871
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6




Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.3	6.0	9.0
122	8.0	6.5	9.7






Abundance Scan 2025 (21.482 min): BN015825.D\data.ms (-2005) (-)


Benzo(a)anthracene
Concen: 1.607 ng
RT: 21.472 min Scan# 2024
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27



Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.2	21.1	31.7
229	19.6	15.7	23.5






Abundance Scan 2030 (21.535 min): BN015825.D\data.ms (-2029) (-)




Chrysene
Concen: 1.549 ng
RT: 21.525 min Scan# 2029
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.1	23.4	35.2
229	19.4	15.8	23.6






Abundance Scan 2019 (21.419 min): BN015825.D\data.ms (-2018) (-)



Bis(2-ethylhexyl)phthalate
Concen: 1.535 ng
RT: 21.408 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27




Tgt Ion	Ratio	Lower	Upper
149	100		
167	29.0	22.9	34.3
279	5.6	4.4	6.6




Abundance Scan 2556 (23.926 min): BN015825.D\data.ms (-2535) (-)

Perylene-d12
Concen: 0.400 ng
RT: 23.916 min Scan# 2553
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6






Tgt Ion:	264	Resp:	53750
Ion Ratio	Lower	Upper	
264	100		
260	23.8	19.1	28.7
265	24.3	19.8	29.8




Abundance Scan 3292 (26.445 min): BN015825.D\data.ms (-3236) (-)


Indeno(1,2,3-cd)pyrene
Concen: 1.655 ng
RT: 26.432 min Scan# 3288
Delta R.T. 0.003 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27



Tgt Ion:	276	Resp:	304926
Ion Ratio	Lower	Upper	
276	100		
138	17.0	13.8	20.8
277	25.3	20.3	30.5






Abundance Scan 2340 (23.187 min): BN015825.D\data.ms (-2337) (-)




Benzo(b)fluoranthene
Concen: 1.641 ng
RT: 23.180 min Scan# 2338
Delta R.T. 0.003 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Tgt Ion:	252	Resp:	308364
Ion Ratio	Lower	Upper	
252	100		
253	22.1	18.2	27.2
125	7.1	6.2	9.4

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6






Abundance Scan 2354 (23.235 min): BN015825.D\data.ms (-2338) (-)




Benzo(k)fluoranthene
Concen: 1.610 ng
RT: 23.228 min Scan# 2352
Delta R.T. 0.003 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Tgt Ion:	252	Resp:	280438
Ion Ratio	Lower	Upper	
252	100		
253	21.9	18.1	27.1
125	6.8	5.6	8.4





Abundance Scan 2525 (23.820 min): BN015825.D\data.ms (-2509) (-)




Benzo(a)pyrene
Concen: 1.633 ng
RT: 23.810 min Scan# 2522
Delta R.T. -0.000 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	252	Resp:	265670
Ion	Ratio	Lower	Upper
252	100		
253	22.0	18.6	27.8
125	7.9	7.3	10.9



Abundance Scan 2522 (23.810 min): BN015858.D\data.ms (-2464) (-)





Abundance Scan 3297 (26.462 min): BN015825.D\data.ms (-3279) (-)




Dibenzo(a,h)anthracene
Concen: 1.686 ng
RT: 26.456 min Scan# 3295
Delta R.T. 0.003 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27


Tgt Ion:	278	Resp:	252491
Ion	Ratio	Lower	Upper
278	100		
139	11.2	10.2	15.2
279	23.7	19.5	29.3



Abundance Scan 3295 (26.456 min): BN015858.D\data.ms (-3265) (-)




Abundance Scan 3516 (27.212 min): BN015825.D\data.ms (-3454) (-)




Benzo(g,h,i)perylene
Concen: 1.587 ng
RT: 27.205 min Scan# 3514
Delta R.T. 0.007 min
Lab File: BN015858.D
Acq: 12 Aug 2021 10:27

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Abundance Scan 3514 (27.205 min): BN015858.D\data.ms



Tgt Ion:	276	Resp:	258768
Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.5	29.3
138	14.3	11.8	17.8



Abundance Scan 3514 (27.205 min): BN015858.D\data.ms (-3454) (-)

