

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081221\
 Data File : BN015865.D
 Acq On : 12 Aug 2021 17:12
 Operator : CG/JU
 Sample : PB138170BL
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB138170BL

Quant Time: Aug 12 17:51:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 12 16:19:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	10102	0.400	ng	0.00
7) Naphthalene-d8	10.723	136	49484	0.400	ng	# 0.00
13) Acenaphthene-d10	14.560	164	28682	0.400	ng	0.01
19) Phenanthrene-d10	17.310	188	60799	0.400	ng	# 0.01
29) Chrysene-d12	21.503	240	66935	0.400	ng	0.01
35) Perylene-d12	23.929	264	60631	0.400	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.494	112	9409	0.377	ng	0.00
5) Phenol-d6	7.080	99	12427	0.366	ng	0.00
8) Nitrobenzene-d5	9.076	82	15300	0.370	ng	0.00
11) 2-Methylnaphthalene-d10	12.314	152	34115	0.428	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	3259	0.258	ng	0.00
15) 2-Fluorobiphenyl	13.190	172	44997	0.398	ng	0.01
27) Fluoranthene-d10	19.341	212	82920	0.443	ng	0.00
31) Terphenyl-d14	19.937	244	86737	0.457	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081221\
Data File : BN015865.D
Acq On : 12 Aug 2021 17:12
Operator : CG/JU
Sample : PB138170BL
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB138170BL

Quant Time: Aug 12 17:51:17 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 12 16:19:54 2021
Response via : Initial Calibration

