

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081221\
 Data File : BN015866.D
 Acq On : 12 Aug 2021 17:49
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 12 18:19:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 12 16:19:54 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.158	0.144	8.9	90	0.00
3	n-Nitrosodimethylamine	0.407	0.398	2.2	107	0.00
4 S	2-Fluorophenol	0.989	0.962	2.7	100	0.00
5 S	Phenol-d6	1.344	1.313	2.3	102	0.00
6	bis(2-Chloroethyl)ether	1.068	1.098	-2.8	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.334	0.307	8.1	106	0.00
9	Naphthalene	1.033	1.018	1.5	102	0.00
10	Hexachlorobutadiene	0.282	0.278	1.4	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.644	0.634	1.6	102	0.00
12	2-Methylnaphthalene	0.712	0.693	2.7	96	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.01
14 S	2,4,6-Tribromophenol	0.176	0.148	15.9	105	0.00
15 S	2-Fluorobiphenyl	1.576	1.476	6.3	103	0.00
16	Acenaphthylene	1.613	1.506	6.6	104	0.00
17	Acenaphthene	1.116	1.049	6.0	104	0.00
18	Fluorene	1.391	1.341	3.6	105	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
20	4,6-Dinitro-2-methylphenol	0.039	0.025	35.9#	104	0.00
21	4-Bromophenyl-phenylether	0.217	0.198	8.8	102	0.00
22	Hexachlorobenzene	0.283	0.269	4.9	104	0.00
23	Atrazine	0.179	0.157	12.3	108	0.00
24	Pentachlorophenol	0.103	0.080	22.3	106	0.00
25	Phenanthrene	1.097	1.053	4.0	106	0.00
26	Anthracene	0.976	0.915	6.2	108	0.00
27 SURR	Fluoranthene-d10	1.233	1.185	3.9	109	0.00
28	Fluoranthene	1.327	1.303	1.8	108	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
30	Pyrene	1.188	1.129	5.0	111	0.00
31 S	Terphenyl-d14	1.135	1.044	8.0	108	0.00
32	Benzo(a)anthracene	1.262	1.175	6.9	109	0.01
33	Chrysene	1.261	1.211	4.0	111	0.00
34	Bis(2-ethylhexyl)phthalate	0.523	0.427	18.4	115	0.00
35 I	Perylene-d12	1.000	1.000	0.0	109	0.00
36	Indeno(1,2,3-cd)pyrene	1.339	1.223	8.7	108	0.00
37	Benzo(b)fluoranthene	1.381	1.330	3.7	110	0.00
38	Benzo(k)fluoranthene	1.297	1.248	3.8	109	0.00
39 C	Benzo(a)pyrene	1.187	1.114	6.1	110	0.00
40	Dibenzo(a,h)anthracene	1.098	1.007	8.3	108	0.00
41	Benzo(g,h,i)perylene	1.157	1.043	9.9	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4