

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081221\
 Data File : BN015857.D
 Acq On : 12 Aug 2021 09:51
 Operator : CG/JU
 Sample : SSTDIC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC0.8

Quant Time: Aug 12 10:38:35 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 12 10:35:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	8644	0.400	ng	0.00
7) Naphthalene-d8	10.723	136	40680	0.400	ng	0.00
13) Acenaphthene-d10	14.548	164	23563	0.400	ng	0.00
19) Phenanthrene-d10	17.297	188	47363	0.400	ng	0.00
29) Chrysene-d12	21.493	240	54005	0.400	ng	0.00
35) Perylene-d12	23.916	264	49958	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.485	112	16766	0.745	ng	0.00
5) Phenol-d6	7.071	99	22918	0.752	ng	0.00
8) Nitrobenzene-d5	9.076	82	25517	0.800	ng	0.00
11) 2-Methylnaphthalene-d10	12.309	152	52800	0.808	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	7797	0.740	ng	0.00
15) 2-Fluorobiphenyl	13.177	172	71377	0.754	ng	0.00
27) Fluoranthene-d10	19.336	212	117177	0.815	ng	0.00
31) Terphenyl-d14	19.931	244	117342	0.740	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.410	88	2863	0.792	ng	98
3) n-Nitrosodimethylamine	3.769	42	7261	0.823	ng	# 98
6) bis(2-Chloroethyl)ether	7.347	93	18594	0.784	ng	99
9) Naphthalene	10.774	128	83535	0.782	ng	100
10) Hexachlorobutadiene	11.065	225	22958	0.832	ng	# 100
12) 2-Methylnaphthalene	12.385	142	57393	0.774	ng	100
16) Acenaphthylene	14.269	152	75706	0.754	ng	100
17) Acenaphthene	14.611	154	51988	0.781	ng	99
18) Fluorene	15.601	166	65193	0.777	ng	100
20) 4,6-Dinitro-2-methylph...	15.677	198	3203	0.689	ng	# 87
21) 4-Bromophenyl-phenylether	16.494	248	20431	0.804	ng	100
22) Hexachlorobenzene	16.616	284	26579	0.808	ng	99
23) Atrazine	16.762	200	16363	0.732	ng	99
24) Pentachlorophenol	16.956	266	9062	0.720	ng	100
25) Phenanthrene	17.346	178	102470	0.772	ng	100
26) Anthracene	17.431	178	91893	0.765	ng	100
28) Fluoranthene	19.366	202	128010	0.805	ng	100
30) Pyrene	19.728	202	126572	0.743	ng	100
32) Benzo(a)anthracene	21.471	228	136235	0.776	ng	100
33) Chrysene	21.525	228	134090	0.762	ng	99
34) Bis(2-ethylhexyl)phtha...	21.408	149	49849	0.623	ng	99
36) Indeno(1,2,3-cd)pyrene	26.431	276	135713	0.792	ng	100
37) Benzo(b)fluoranthene	23.176	252	137962	0.790	ng	99
38) Benzo(k)fluoranthene	23.224	252	131139	0.810	ng	99
39) Benzo(a)pyrene	23.806	252	117675	0.778	ng	98
40) Dibenzo(a,h)anthracene	26.452	278	112014	0.805	ng	98
41) Benzo(g,h,i)perylene	27.198	276	115696	0.764	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081221\
Data File : BN015857.D
Acq On : 12 Aug 2021 09:51
Operator : CG/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Quant Time: Aug 12 10:38:35 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 12 10:35:26 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.928 min Scan# 526
Delta R.T. 0.009 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:152 Resp: 8644
Ion Ratio Lower Upper
152 100
150 150.8 125.8 188.8
115 63.9 53.6 80.4



#2
1,4-Dioxane
Concen: 0.792 ng
RT: 3.410 min Scan# 36
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion: 88 Resp: 2863
Ion Ratio Lower Upper
88 100
43 37.0 29.4 44.2
58 83.5 69.3 103.9









#3
n-Nitrosodimethylamine
Concen: 0.823 ng
RT: 3.769 min Scan# 75
Delta R.T. -0.010 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion: 42 Resp: 7261




Ion	Ratio	Lower	Upper
42	100		
74	105.8	86.1	129.1
44	6.7	3.8	5.8




#4
2-Fluorophenol
Concen: 0.745 ng
RT: 5.485 min Scan# 261
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion: 112 Resp: 16766

Ion	Ratio	Lower	Upper
112	100		
64	67.0	53.3	79.9
63	36.3	28.6	42.8









#5
Phenol-d6
Concen: 0.752 ng
RT: 7.071 min Scan# 433
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	99	Resp:	22918
Ion	Ratio	Lower	Upper
99	100		
42	23.1	18.7	28.1
71	40.6	32.4	48.6



#6
bis(2-Chloroethyl)ether
Concen: 0.784 ng
RT: 7.347 min Scan# 463
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:	93	Resp:	18594
Ion	Ratio	Lower	Upper
93	100		
63	79.1	63.8	95.6
95	32.8	26.0	39.0









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.723 min Scan# 758
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	136	Resp:	40680
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.4
54	7.2	5.4	8.2
68	5.0	3.8	5.8



#8
Nitrobenzene-d5
Concen: 0.800 ng
RT: 9.076 min Scan# 628
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:	82	Resp:	25517
Ion	Ratio	Lower	Upper
82	100		
128	36.8	29.2	43.8
54	52.4	41.1	61.7









#9
Naphthalene
Concen: 0.782 ng
RT: 10.774 min Scan# 762
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	128	Resp:	83535
Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.8	13.2
127	13.1	10.5	15.7



#10
Hexachlorobutadiene
Concen: 0.832 ng
RT: 11.065 min Scan# 785
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:	225	Resp:	22958
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.1	51.4	77.0









#11
2-Methylnaphthalene-d10
Concen: 0.808 ng
RT: 12.309 min Scan# 1000
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 52800
Ion Ratio Lower Upper
152 100
151 20.4 16.5 24.7



#12
2-Methylnaphthalene
Concen: 0.774 ng
RT: 12.385 min Scan# 1017
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:142 Resp: 57393
Ion Ratio Lower Upper
142 100
141 88.3 70.5 105.7
115 37.6 29.8 44.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.548 min Scan# 1205
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:164 Resp: 23563
Ion Ratio Lower Upper
164 100
162 103.6 83.5 125.3
160 47.2 37.7 56.5



#14
2,4,6-Tribromophenol
Concen: 0.740 ng
RT: 16.044 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:330 Resp: 7797
Ion Ratio Lower Upper
330 100
332 101.1 84.1 126.1
141 31.0 24.2 36.4









#15
2-Fluorobiphenyl
Concen: 0.754 ng
RT: 13.177 min Scan# 1097
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:172	Resp:	71377
Ion Ratio	Lower	Upper
172	100	
171	35.7	28.7 43.1
170	23.5	19.0 28.4



#16
Acenaphthylene
Concen: 0.754 ng
RT: 14.269 min Scan# 1183
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:152	Resp:	75706
Ion Ratio	Lower	Upper
152	100	
151	19.4	15.4 23.2
153	13.1	10.5 15.7









#17
Acenaphthene
Concen: 0.781 ng
RT: 14.611 min Scan# 1210
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:154 Resp: 51988
Ion Ratio Lower Upper
154 100
153 111.6 90.1 135.1
152 52.8 43.5 65.3



#18
Fluorene
Concen: 0.777 ng
RT: 15.601 min Scan# 1288
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:166 Resp: 65193
Ion Ratio Lower Upper
166 100
165 97.4 77.8 116.6
167 13.6 11.0 16.4









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.297 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:188	Resp:	47363
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	9.5	7.2 10.8



#20
4,6-Dinitro-2-methylphenol
Concen: 0.689 ng
RT: 15.677 min Scan# 1294
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:198	Resp:	3203
Ion Ratio	Lower	Upper
198	100	
182	9.3	11.9 17.9#
77	0.0	0.0









#21
4-Bromophenyl-phenylether
Concen: 0.804 ng
RT: 16.494 min Scan# 1360
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	248	Resp:	20431
Ion	Ratio	Lower	Upper
248	100		
250	97.9	78.6	118.0
141	54.0	43.4	65.2



#22
Hexachlorobenzene
Concen: 0.808 ng
RT: 16.616 min Scan# 1370
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:	284	Resp:	26579
Ion	Ratio	Lower	Upper
284	100		
142	28.0	22.6	34.0
249	32.0	25.2	37.8









#23
Atrazine
Concen: 0.732 ng
RT: 16.762 min Scan# 1382
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:200 Resp: 16363
Ion Ratio Lower Upper
200 100
173 21.2 16.4 24.6
215 52.1 41.5 62.3



#24
Pentachlorophenol
Concen: 0.720 ng
RT: 16.956 min Scan# 1398
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:266 Resp: 9062
Ion Ratio Lower Upper
266 100
264 62.5 50.1 75.1
268 64.1 51.0 76.6









#25
Phenanthrene
Concen: 0.772 ng
RT: 17.346 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:178 Resp: 102470
Ion Ratio Lower Upper
178 100
176 18.4 14.6 21.8
179 15.7 12.6 18.8



#26
Anthracene
Concen: 0.765 ng
RT: 17.431 min Scan# 1437
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:178 Resp: 91893
Ion Ratio Lower Upper
178 100
176 17.8 14.3 21.5
179 15.4 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.815 ng
RT: 19.336 min Scan# 1751
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:212 Resp: 117177
Ion Ratio Lower Upper
212 100
106 9.2 7.4 11.2
104 5.3 4.3 6.5



#28
Fluoranthene
Concen: 0.805 ng
RT: 19.366 min Scan# 1757
Delta R.T. 0.005 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:202 Resp: 128010
Ion Ratio Lower Upper
202 100
101 8.6 6.8 10.2
203 17.5 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.493 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	240	Resp:	54005
Ion Ratio	Lower	Upper	
240	100		
120	4.4	3.8	5.8
236	25.1	20.0	30.0



#30
Pyrene
Concen: 0.743 ng
RT: 19.728 min Scan# 1830
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:	202	Resp:	126572
Ion Ratio	Lower	Upper	
202	100		
200	20.2	16.2	24.2
203	17.5	14.1	21.1









#31
Terphenyl-d14
Concen: 0.740 ng
RT: 19.931 min Scan# 1871
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:244 Resp: 117342
Ion Ratio Lower Upper
244 100
212 7.4 6.0 9.0
122 7.7 6.5 9.7



#32
Benzo(a)anthracene
Concen: 0.776 ng
RT: 21.471 min Scan# 2024
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:228 Resp: 136235
Ion Ratio Lower Upper
228 100
226 26.1 21.1 31.7
229 19.5 15.7 23.5









#33
Chrysene
Concen: 0.762 ng
RT: 21.525 min Scan# 2029
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:228 Resp: 134090
Ion Ratio Lower Upper
228 100
226 28.9 23.4 35.2
229 19.5 15.8 23.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.623 ng
RT: 21.408 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51


Tgt Ion:149 Resp: 49849
Ion Ratio Lower Upper
149 100
167 29.0 22.9 34.3
279 5.5 4.4 6.6








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.916 min Scan# 2553
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8







Tgt Ion:264 Resp: 49958
Ion Ratio Lower Upper
264 100
260 23.9 19.1 28.7
265 24.8 19.8 29.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.792 ng
RT: 26.431 min Scan# 3288
Delta R.T. 0.003 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:276 Resp: 135713
Ion Ratio Lower Upper
276 100
138 17.1 13.8 20.8
277 25.3 20.3 30.5









#37
Benzo(b)fluoranthene
Concen: 0.790 ng
RT: 23.176 min Scan# 2337
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:252 Resp: 137962
Ion Ratio Lower Upper
252 100
253 22.3 18.2 27.2
125 7.3 6.2 9.4



#38
Benzo(k)fluoranthene
Concen: 0.810 ng
RT: 23.224 min Scan# 2351
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:252 Resp: 131139
Ion Ratio Lower Upper
252 100
253 22.0 18.1 27.1
125 6.9 5.6 8.4









#39
Benzo(a)pyrene
Concen: 0.778 ng
RT: 23.806 min Scan# 2521
Delta R.T. -0.004 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:252	Resp:	117675
Ion Ratio	Lower	Upper
252	100	
253	22.4	18.6 27.8
125	8.1	7.3 10.9



#40
Dibenzo(a,h)anthracene
Concen: 0.805 ng
RT: 26.452 min Scan# 3294
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Tgt Ion:278	Resp:	112014
Ion Ratio	Lower	Upper
278	100	
139	11.8	10.2 15.2
279	23.8	19.5 29.3





#41
Benzo(g,h,i)perylene
Concen: 0.764 ng
RT: 27.198 min Scan# 3512
Delta R.T. -0.000 min
Lab File: BN015857.D
Acq: 12 Aug 2021 09:51

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Tgt Ion:276 Resp: 115696
Ion Ratio Lower Upper
276 100
277 24.1 19.5 29.3
138 14.5 11.8 17.8

