

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081221\
 Data File : BN015859.D
 Acq On : 12 Aug 2021 11:04
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Aug 12 11:36:30 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 12 10:35:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	9542	0.400	ng	0.00
7) Naphthalene-d8	10.724	136	42330	0.400	ng	0.00
13) Acenaphthene-d10	14.561	164	24781	0.400	ng	0.01
19) Phenanthrene-d10	17.298	188	50358	0.400	ng	0.00
29) Chrysene-d12	21.493	240	59229	0.400	ng	# 0.00
35) Perylene-d12	23.919	264	56789	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.494	112	73666	2.966	ng	0.00
5) Phenol-d6	7.080	99	101447	3.014	ng	0.00
8) Nitrobenzene-d5	9.076	82	120981	3.644	ng	0.00
11) 2-Methylnaphthalene-d10	12.314	152	220197	3.239	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	41086	3.708	ng	0.00
15) 2-Fluorobiphenyl	13.178	172	293561	2.948	ng	0.00
27) Fluoranthene-d10	19.336	212	533575	3.488	ng	0.00
31) Terphenyl-d14	19.932	244	506856	2.914	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.410	88	12438	3.115	ng	96
3) n-Nitrosodimethylamine	3.761	42	34724	3.566	ng	# 99
6) bis(2-Chloroethyl)ether	7.347	93	78643	3.003	ng	99
9) Naphthalene	10.774	128	345755	3.110	ng	100
10) Hexachlorobutadiene	11.066	225	94643	3.295	ng	# 100
12) 2-Methylnaphthalene	12.385	142	233639	3.028	ng	100
16) Acenaphthylene	14.269	152	345470	3.273	ng	100
17) Acenaphthene	14.611	154	224325	3.204	ng	97
18) Fluorene	15.601	166	281550	3.191	ng	100
20) 4,6-Dinitro-2-methylph...	15.677	198	23853	4.824	ng	# 77
21) 4-Bromophenyl-phenylether	16.494	248	89336	3.308	ng	98
22) Hexachlorobenzene	16.616	284	114823	3.283	ng	99
23) Atrazine	16.762	200	86249	3.630	ng	99
24) Pentachlorophenol	16.957	266	52550	3.928	ng	100
25) Phenanthrene	17.346	178	443570	3.145	ng	100
26) Anthracene	17.431	178	425615	3.334	ng	100
28) Fluoranthene	19.366	202	548436	3.244	ng	99
30) Pyrene	19.728	202	561182	3.002	ng	100
32) Benzo(a)anthracene	21.472	228	612314	3.181	ng	100
33) Chrysene	21.525	228	588822	3.052	ng	99
34) Bis(2-ethylhexyl)phtha...	21.408	149	305815	3.487	ng	99
36) Indeno(1,2,3-cd)pyrene	26.435	276	640412	3.289	ng	100
37) Benzo(b)fluoranthene	23.180	252	632725	3.188	ng	98
38) Benzo(k)fluoranthene	23.231	252	570707	3.101	ng	99
39) Benzo(a)pyrene	23.813	252	554544	3.226	ng	97
40) Dibenzo(a,h)anthracene	26.459	278	529849	3.349	ng	97
41) Benzo(g,h,i)perylene	27.209	276	543833	3.158	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081221\
Data File : BN015859.D
Acq On : 12 Aug 2021 11:04
Operator : CG/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Aug 12 11:36:30 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 12 10:35:26 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.928 min Scan# 526
Delta R.T. 0.009 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 9542
Ion Ratio Lower Upper
152 100
150 158.1 125.8 188.8
115 62.8 53.6 80.4



#2
1,4-Dioxane
Concen: 3.115 ng
RT: 3.410 min Scan# 36
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion: 88 Resp: 12438
Ion Ratio Lower Upper
88 100
43 35.5 29.4 44.2
58 82.0 69.3 103.9









#3
n-Nitrosodimethylamine
Concen: 3.566 ng
RT: 3.761 min Scan# 74
Delta R.T. -0.018 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 34724
Ion Ratio Lower Upper
42 100
74 108.5 86.1 129.1
44 5.9 3.8 5.8#



#4
2-Fluorophenol
Concen: 2.966 ng
RT: 5.494 min Scan# 262
Delta R.T. 0.009 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion: 112 Resp: 73666
Ion Ratio Lower Upper
112 100
64 67.1 53.3 79.9
63 35.8 28.6 42.8









#5
Phenol-d6
Concen: 3.014 ng
RT: 7.080 min Scan# 434
Delta R.T. 0.009 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 101447
Ion Ratio Lower Upper
99 100
42 22.9 18.7 28.1
71 40.4 32.4 48.6



#6
bis(2-Chloroethyl)ether
Concen: 3.003 ng
RT: 7.347 min Scan# 463
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion: 93 Resp: 78643
Ion Ratio Lower Upper
93 100
63 78.8 63.8 95.6
95 32.7 26.0 39.0









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.724 min Scan# 758
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	42330
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.0	13.4
54	7.3	5.4	8.2
68	5.2	3.8	5.8



#8
Nitrobenzene-d5
Concen: 3.644 ng
RT: 9.076 min Scan# 628
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion:	82	Resp:	120981
Ion	Ratio	Lower	Upper
82	100		
128	37.4	29.2	43.8
54	50.7	41.1	61.7









#9
Naphthalene
Concen: 3.110 ng
RT: 10.774 min Scan# 762
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 345755
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 13.1 10.5 15.7



#10
Hexachlorobutadiene
Concen: 3.295 ng
RT: 11.066 min Scan# 785
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion:225 Resp: 94643
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 3.239 ng
RT: 12.314 min Scan# 1001
Delta R.T. 0.004 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 220197
Ion Ratio Lower Upper
152 100
151 20.3 16.5 24.7



#12
2-Methylnaphthalene
Concen: 3.028 ng
RT: 12.385 min Scan# 1017
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion:142 Resp: 233639
Ion Ratio Lower Upper
142 100
141 88.6 70.5 105.7
115 37.2 29.8 44.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.561 min Scan# 1206
Delta R.T. 0.013 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 24781
Ion Ratio Lower Upper
164 100
162 99.1 83.5 125.3
160 43.0 37.7 56.5



#14
2,4,6-Tribromophenol
Concen: 3.708 ng
RT: 16.044 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion:330 Resp: 41086
Ion Ratio Lower Upper
330 100
332 96.6 84.1 126.1
141 30.5 24.2 36.4









#15
2-Fluorobiphenyl
Concen: 2.948 ng
RT: 13.178 min Scan# 1097
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 293561
Ion Ratio Lower Upper
172 100
171 35.6 28.7 43.1
170 23.3 19.0 28.4



#16
Acenaphthylene
Concen: 3.273 ng
RT: 14.269 min Scan# 1183
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion:152 Resp: 345470
Ion Ratio Lower Upper
152 100
151 19.5 15.4 23.2
153 13.1 10.5 15.7









#17
Acenaphthene
Concen: 3.204 ng
RT: 14.611 min Scan# 1210
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 224325
Ion Ratio Lower Upper
154 100
153 109.3 90.1 135.1
152 51.3 43.5 65.3



#18
Fluorene
Concen: 3.191 ng
RT: 15.601 min Scan# 1288
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion:166 Resp: 281550
Ion Ratio Lower Upper
166 100
165 97.6 77.8 116.6
167 13.5 11.0 16.4









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.298 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:188	Resp:	50358
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	10.0	7.2 10.8



#20
4,6-Dinitro-2-methylphenol
Concen: 4.824 ng
RT: 15.677 min Scan# 1294
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion:198	Resp:	23853
Ion Ratio	Lower	Upper
198	100	
182	5.3	11.9 17.9#
77	0.0	0.0









#21
4-Bromophenyl-phenylether
Concen: 3.308 ng
RT: 16.494 min Scan# 1360
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 89336
Ion Ratio Lower Upper
248 100
250 97.7 78.6 118.0
141 57.5 43.4 65.2



#22
Hexachlorobenzene
Concen: 3.283 ng
RT: 16.616 min Scan# 1370
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion:284 Resp: 114823
Ion Ratio Lower Upper
284 100
142 27.8 22.6 34.0
249 31.7 25.2 37.8









#23
Atrazine
Concen: 3.630 ng
RT: 16.762 min Scan# 1382
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 86249
Ion Ratio Lower Upper
200 100
173 21.4 16.4 24.6
215 51.2 41.5 62.3



#24
Pentachlorophenol
Concen: 3.928 ng
RT: 16.957 min Scan# 1398
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion:266 Resp: 52550
Ion Ratio Lower Upper
266 100
264 62.9 50.1 75.1
268 63.9 51.0 76.6









#25
Phenanthrene
Concen: 3.145 ng
RT: 17.346 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 443570
Ion Ratio Lower Upper
178 100
176 18.4 14.6 21.8
179 15.5 12.6 18.8



#26
Anthracene
Concen: 3.334 ng
RT: 17.431 min Scan# 1437
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion:178 Resp: 425615
Ion Ratio Lower Upper
178 100
176 17.8 14.3 21.5
179 15.4 12.2 18.4









#27
Fluoranthene-d10
Concen: 3.488 ng
RT: 19.336 min Scan# 1751
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 533575
Ion Ratio Lower Upper
212 100
106 9.3 7.4 11.2
104 5.3 4.3 6.5



#28
Fluoranthene
Concen: 3.244 ng
RT: 19.366 min Scan# 1757
Delta R.T. 0.005 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion:202 Resp: 548436
Ion Ratio Lower Upper
202 100
101 8.7 6.8 10.2
203 17.5 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.493 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240 Resp: 59229
Ion Ratio Lower Upper
240 100
120 6.1 3.8 5.8#
236 25.6 20.0 30.0



#30
Pyrene
Concen: 3.002 ng
RT: 19.728 min Scan# 1830
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion:202 Resp: 561182
Ion Ratio Lower Upper
202 100
200 20.1 16.2 24.2
203 17.7 14.1 21.1









#31
Terphenyl-d14
Concen: 2.914 ng
RT: 19.932 min Scan# 1871
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:244 Resp: 506856
Ion Ratio Lower Upper
244 100
212 7.4 6.0 9.0
122 8.4 6.5 9.7



#32
Benzo(a)anthracene
Concen: 3.181 ng
RT: 21.472 min Scan# 2024
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion:228 Resp: 612314
Ion Ratio Lower Upper
228 100
226 26.6 21.1 31.7
229 19.4 15.7 23.5









#33
Chrysene
Concen: 3.052 ng
RT: 21.525 min Scan# 2029
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:228 Resp: 588822
Ion Ratio Lower Upper
228 100
226 29.6 23.4 35.2
229 19.2 15.8 23.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 3.487 ng
RT: 21.408 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04


Tgt Ion:149 Resp: 305815
Ion Ratio Lower Upper
149 100
167 29.2 22.9 34.3
279 5.7 4.4 6.6








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.919 min Scan# 2554
Delta R.T. 0.003 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2





Tgt Ion:264 Resp: 56789
Ion Ratio Lower Upper
264 100
260 24.0 19.1 28.7
265 24.0 19.8 29.8




#36
Indeno(1,2,3-cd)pyrene
Concen: 3.289 ng
RT: 26.435 min Scan# 3289
Delta R.T. 0.007 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04



Tgt Ion:276 Resp: 640412
Ion Ratio Lower Upper
276 100
138 16.9 13.8 20.8
277 25.4 20.3 30.5









#37
Benzo(b)fluoranthene
Concen: 3.188 ng
RT: 23.180 min Scan# 2338
Delta R.T. 0.003 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 632725
Ion Ratio Lower Upper
252 100
253 22.0 18.2 27.2
125 7.0 6.2 9.4



#38
Benzo(k)fluoranthene
Concen: 3.101 ng
RT: 23.231 min Scan# 2353
Delta R.T. 0.007 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion:252 Resp: 570707
Ion Ratio Lower Upper
252 100
253 21.9 18.1 27.1
125 6.7 5.6 8.4









#39
Benzo(a)pyrene
Concen: 3.226 ng
RT: 23.813 min Scan# 2523
Delta R.T. 0.003 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 554544
Ion Ratio Lower Upper
252 100
253 22.0 18.6 27.8
125 7.6 7.3 10.9



#40
Dibenzo(a,h)anthracene
Concen: 3.349 ng
RT: 26.459 min Scan# 3296
Delta R.T. 0.007 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Tgt Ion:278 Resp: 529849
Ion Ratio Lower Upper
278 100
139 10.9 10.2 15.2
279 23.7 19.5 29.3





#41
Benzo(g,h,i)perylene
Concen: 3.158 ng
RT: 27.209 min Scan# 3515
Delta R.T. 0.010 min
Lab File: BN015859.D
Acq: 12 Aug 2021 11:04

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:276 Resp: 543833
Ion Ratio Lower Upper
276 100
277 23.8 19.5 29.3
138 14.4 11.8 17.8

