

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081324\
 Data File : BN033357.D
 Acq On : 13 Aug 2024 08:55
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 13 10:45:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 12 01:09:50 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	73	0.00
2	1,4-Dioxane	0.400	0.405	-1.3	77	0.00
3	n-Nitrosodimethylamine	0.400	0.427	-6.7	80	0.00
4 S	2-Fluorophenol	0.400	0.370	7.5	66	0.00
5 S	Phenol-d6	0.400	0.375	6.3	67	0.00
6	bis(2-Chloroethyl)ether	0.400	0.383	4.3	71	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	69	0.00
8 S	Nitrobenzene-d5	0.400	0.399	0.3	71	0.00
9	Naphthalene	0.400	0.400	0.0	71	0.00
10	Hexachlorobutadiene	0.400	0.425	-6.2	74	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.379	5.3	68	0.00
12	2-Methylnaphthalene	0.400	0.388	3.0	70	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	68	0.00
14 S	2,4,6-Tribromophenol	0.400	0.334	16.5	61	0.00
15 S	2-Fluorobiphenyl	0.400	0.411	-2.7	70	0.00
16	Acenaphthylene	0.400	0.378	5.5	69	0.00
17	Acenaphthene	0.400	0.399	0.3	71	-0.01
18	Fluorene	0.400	0.401	-0.3	71	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	70	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.371	7.3	76	-0.01
21	4-Bromophenyl-phenylether	0.400	0.379	5.3	70	0.00
22	Hexachlorobenzene	0.400	0.412	-3.0	74	0.00
23	Atrazine	0.400	0.322	19.5	62	-0.01
24	Pentachlorophenol	0.400	0.401	-0.3	83	0.00
25	Phenanthrene	0.400	0.398	0.5	73	0.00
26	Anthracene	0.400	0.369	7.8	69	0.00
27 SURR	Fluoranthene-d10	0.400	0.361	9.8	67	0.00
28	Fluoranthene	0.400	0.373	6.8	69	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	64	0.00
30	Pyrene	0.400	0.424	-6.0	68	0.00
31 S	Terphenyl-d14	0.400	0.403	-0.8	64	0.00
32	Benzo(a)anthracene	0.400	0.366	8.5	63	0.00
33	Chrysene	0.400	0.418	-4.5	67	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.325	18.8	60	0.00
35 I	Perylene-d12	0.400	0.400	0.0	59	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.332	17.0	54	0.00
37	Benzo(b)fluoranthene	0.400	0.392	2.0	62	0.00
38	Benzo(k)fluoranthene	0.400	0.407	-1.7	62	0.00
39 C	Benzo(a)pyrene	0.400	0.358	10.5	58	0.00
40	Dibenzo(a,h)anthracene	0.400	0.330	17.5	54	0.00
41	Benzo(g,h,i)perylene	0.400	0.343	14.2	55	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4