

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081324\
 Data File : BN033364.D
 Acq On : 13 Aug 2024 14:22
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 13 15:16:10 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 12 01:09:50 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	81	0.00
2	1,4-Dioxane	0.400	0.403	-0.8	85	0.00
3	n-Nitrosodimethylamine	0.400	0.443	-10.7	92	0.00
4 S	2-Fluorophenol	0.400	0.372	7.0	74	0.00
5 S	Phenol-d6	0.400	0.381	4.8	76	0.00
6	bis(2-Chloroethyl)ether	0.400	0.390	2.5	80	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	78	0.00
8 S	Nitrobenzene-d5	0.400	0.405	-1.3	81	0.00
9	Naphthalene	0.400	0.403	-0.8	80	-0.01
10	Hexachlorobutadiene	0.400	0.413	-3.2	81	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.383	4.3	77	0.00
12	2-Methylnaphthalene	0.400	0.391	2.3	79	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	75	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.327	18.3	66	0.00
15 S	2-Fluorobiphenyl	0.400	0.415	-3.7	78	0.00
16	Acenaphthylene	0.400	0.380	5.0	76	0.00
17	Acenaphthene	0.400	0.402	-0.5	79	-0.01
18	Fluorene	0.400	0.402	-0.5	79	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	76	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.373	6.8	83	-0.01
21	4-Bromophenyl-phenylether	0.400	0.376	6.0	75	0.00
22	Hexachlorobenzene	0.400	0.398	0.5	78	-0.01
23	Atrazine	0.400	0.332	17.0	69	-0.01
24	Pentachlorophenol	0.400	0.383	4.3	85	0.00
25	Phenanthrene	0.400	0.403	-0.8	80	0.00
26	Anthracene	0.400	0.377	5.8	77	-0.01
27 SURR	Fluoranthene-d10	0.400	0.366	8.5	73	0.00
28	Fluoranthene	0.400	0.378	5.5	76	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	72	0.00
30	Pyrene	0.400	0.412	-3.0	76	0.00
31 S	Terphenyl-d14	0.400	0.395	1.3	71	0.00
32	Benzo(a)anthracene	0.400	0.380	5.0	74	0.00
33	Chrysene	0.400	0.408	-2.0	74	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.341	14.8	71	0.00
35 I	Perylene-d12	0.400	0.400	0.0	69	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.339	15.3	64	0.00
37	Benzo(b)fluoranthene	0.400	0.396	1.0	72	0.00
38	Benzo(k)fluoranthene	0.400	0.408	-2.0	72	0.00
39 C	Benzo(a)pyrene	0.400	0.358	10.5	67	0.00
40	Dibenzo(a,h)anthracene	0.400	0.344	14.0	65	0.00
41	Benzo(g,h,i)perylene	0.400	0.351	12.3	66	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4