

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081324\
 Data File : BN033357.D
 Acq On : 13 Aug 2024 08:55
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Aug 13 10:45:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 12 01:09:50 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.573	152	3864	0.400	ng	0.00
7) Naphthalene-d8	10.338	136	10668	0.400	ng	# 0.00
13) Acenaphthene-d10	14.212	164	6200	0.400	ng	0.00
19) Phenanthrene-d10	16.958	188	15498	0.400	ng	0.00
29) Chrysene-d12	21.166	240	12286	0.400	ng	0.00
35) Perylene-d12	23.349	264	11784	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.197	112	3953	0.370	ng	0.00
5) Phenol-d6	6.749	99	5403	0.375	ng	0.00
8) Nitrobenzene-d5	8.704	82	3209	0.399	ng	0.00
11) 2-Methylnaphthalene-d10	11.936	152	6224	0.379	ng	0.00
14) 2,4,6-Tribromophenol	15.705	330	1063	0.334	ng	0.00
15) 2-Fluorobiphenyl	12.833	172	10450	0.411	ng	0.00
27) Fluoranthene-d10	18.998	212	14952	0.361	ng	0.00
31) Terphenyl-d14	19.615	244	9354	0.403	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.196	88	1738	0.405	ng	97
3) n-Nitrosodimethylamine	3.485	42	2362	0.427	ng	# 86
6) bis(2-Chloroethyl)ether	7.009	93	4474	0.383	ng	97
9) Naphthalene	10.391	128	11777	0.400	ng	99
10) Hexachlorobutadiene	10.690	225	2383	0.425	ng	# 100
12) 2-Methylnaphthalene	12.011	142	7701	0.388	ng	99
16) Acenaphthylene	13.924	152	10938	0.378	ng	99
17) Acenaphthene	14.266	154	7915	0.399	ng	98
18) Fluorene	15.260	166	10500	0.401	ng	98
20) 4,6-Dinitro-2-methylph...	15.335	198	697	0.371	ng	# 80
21) 4-Bromophenyl-phenylether	16.164	248	3516	0.379	ng	# 87
22) Hexachlorobenzene	16.276	284	4281	0.412	ng	96
23) Atrazine	16.437	200	2368	0.322	ng	# 90
24) Pentachlorophenol	16.611	266	1685	0.401	ng	99
25) Phenanthrene	16.995	178	17831	0.398	ng	100
26) Anthracene	17.095	178	14616	0.369	ng	100
28) Fluoranthene	19.030	202	20558	0.373	ng	99
30) Pyrene	19.392	202	20717	0.424	ng	100
32) Benzo(a)anthracene	21.157	228	16786	0.366	ng	99
33) Chrysene	21.201	228	19328	0.418	ng	100
34) Bis(2-ethylhexyl)phtha...	21.130	149	6789	0.325	ng	99
36) Indeno(1,2,3-cd)pyrene	25.521	276	16153	0.332	ng	98
37) Benzo(b)fluoranthene	22.703	252	17391	0.392	ng	98
38) Benzo(k)fluoranthene	22.747	252	18446	0.407	ng	97
39) Benzo(a)pyrene	23.255	252	13375	0.358	ng	# 96
40) Dibenzo(a,h)anthracene	25.542	278	12672	0.330	ng	95
41) Benzo(g,h,i)perylene	26.176	276	14642	0.343	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081324\
Data File : BN033357.D
Acq On : 13 Aug 2024 08:55
Operator : MA/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Quant Time: Aug 13 10:45:47 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 12 01:09:50 2024
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.573 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:152 Resp: 3864
Ion Ratio Lower Upper
152 100
150 155.0 122.3 183.5
115 62.8 48.8 73.2



#2
1,4-Dioxane
Concen: 0.405 ng
RT: 3.196 min Scan# 16
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Tgt Ion: 88 Resp: 1738
Ion Ratio Lower Upper
88 100
43 36.7 30.6 46.0
58 83.2 64.5 96.7









#3
 n-Nitrosodimethylamine
 Concen: 0.427 ng
 RT: 3.485 min Scan# 50
 Delta R.T. 0.000 min
 Lab File: BN033357.D
 Acq: 13 Aug 2024 08:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion: 42 Resp: 2362
 Ion Ratio Lower Upper
 42 100
 74 97.1 89.0 133.6
 44 21.2 11.0 16.4#



#4
 2-Fluorophenol
 Concen: 0.370 ng
 RT: 5.197 min Scan# 293
 Delta R.T. 0.000 min
 Lab File: BN033357.D
 Acq: 13 Aug 2024 08:55

Tgt Ion: 112 Resp: 3953
 Ion Ratio Lower Upper
 112 100
 64 60.4 48.1 72.1
 63 33.7 25.4 38.2









#5
Phenol-d6
Concen: 0.375 ng
RT: 6.749 min Scan# 508
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 5403
Ion Ratio Lower Upper
99 100
42 25.0 18.5 27.7
71 33.9 26.2 39.2



#6
bis(2-Chloroethyl)ether
Concen: 0.383 ng
RT: 7.009 min Scan# 544
Delta R.T. -0.007 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Tgt Ion: 93 Resp: 4474
Ion Ratio Lower Upper
93 100
63 85.4 65.2 97.8
95 33.5 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.338 min Scan# 914
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	10668
Ion	Ratio	Lower	Upper
136	100		
137	11.8	9.1	13.7
54	9.6	9.4	14.0
68	6.9	7.0	10.4



#8
Nitrobenzene-d5
Concen: 0.399 ng
RT: 8.704 min Scan# 761
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Tgt Ion:	82	Resp:	3209
Ion	Ratio	Lower	Upper
82	100		
128	37.7	29.0	43.4
54	60.6	50.2	75.4









#9
Naphthalene
Concen: 0.400 ng
RT: 10.391 min Scan# 919
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	128	Resp:	11777
Ion	Ratio	Lower	Upper
128	100		
129	11.7	9.2	13.8
127	13.5	10.6	15.8



#10
Hexachlorobutadiene
Concen: 0.425 ng
RT: 10.690 min Scan# 947
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Tgt Ion:	225	Resp:	2383
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.2	50.8	76.2









#11
2-Methylnaphthalene-d10
Concen: 0.379 ng
RT: 11.936 min Scan# 1136
Delta R.T. -0.004 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 6224
Ion Ratio Lower Upper
152 100
151 21.3 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.388 ng
RT: 12.011 min Scan# 1156
Delta R.T. -0.004 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Tgt Ion:142 Resp: 7701
Ion Ratio Lower Upper
142 100
141 87.8 70.2 105.4
115 37.8 29.3 43.9









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.212 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 6200
Ion Ratio Lower Upper
164 100
162 104.7 83.1 124.7
160 54.9 41.1 61.7



#14
2,4,6-Tribromophenol
Concen: 0.334 ng
RT: 15.705 min Scan# 1601
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Tgt Ion:330 Resp: 1063
Ion Ratio Lower Upper
330 100
332 93.9 77.1 115.7
141 44.5 33.4 50.0









#15
2-Fluorobiphenyl
Concen: 0.411 ng
RT: 12.833 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172 Resp: 10450
Ion Ratio Lower Upper
172 100
171 34.3 27.8 41.8
170 23.7 18.7 28.1



#16
Acenaphthylene
Concen: 0.378 ng
RT: 13.924 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Tgt Ion:152 Resp: 10938
Ion Ratio Lower Upper
152 100
151 19.3 15.7 23.5
153 13.1 10.4 15.6









#17
Acenaphthene
Concen: 0.399 ng
RT: 14.266 min Scan# 1470
Delta R.T. -0.011 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 7915
Ion Ratio Lower Upper
154 100
153 112.5 89.7 134.5
152 62.6 47.2 70.8



#18
Fluorene
Concen: 0.401 ng
RT: 15.260 min Scan# 1562
Delta R.T. -0.011 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Tgt Ion:166 Resp: 10500
Ion Ratio Lower Upper
166 100
165 100.0 78.0 117.0
167 13.5 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.958 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN033357.D

Acq: 13 Aug 2024 08:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:188 Resp: 15498

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.0 10.7 16.1



#20

4,6-Dinitro-2-methylphenol

Concen: 0.371 ng

RT: 15.335 min Scan# 1569

Delta R.T. -0.011 min

Lab File: BN033357.D

Acq: 13 Aug 2024 08:55




Tgt Ion:198 Resp: 697


Ion Ratio Lower Upper

198 100

51 95.2 62.8 94.2#

105 56.4 34.1 51.1#









#21
4-Bromophenyl-phenylether
Concen: 0.379 ng
RT: 16.164 min Scan# 1638
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 3516
Ion Ratio Lower Upper
248 100
250 95.4 73.0 109.4
141 61.4 64.6 96.8#



#22
Hexachlorobenzene
Concen: 0.412 ng
RT: 16.276 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Tgt Ion:284 Resp: 4281
Ion Ratio Lower Upper
284 100
142 43.9 32.2 48.4
249 32.2 25.3 37.9









#23
Atrazine
Concen: 0.322 ng
RT: 16.437 min Scan# 1660
Delta R.T. -0.012 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 2368
Ion Ratio Lower Upper
200 100
173 33.2 18.7 28.1#
215 45.1 38.6 58.0



#24
Pentachlorophenol
Concen: 0.401 ng
RT: 16.611 min Scan# 1674
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Tgt Ion:266 Resp: 1685
Ion Ratio Lower Upper
266 100
264 62.9 49.2 73.8
268 63.8 51.2 76.8





#25

Phenanthrene

Concen: 0.398 ng

RT: 16.995 min Scan# 1705

Delta R.T. 0.000 min

Lab File: BN033357.D

Acq: 13 Aug 2024 08:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:178 Resp: 17831

Ion Ratio Lower Upper

178 100

176 19.2 15.4 23.0

179 15.3 12.2 18.2



#26

Anthracene

Concen: 0.369 ng

RT: 17.095 min Scan# 1713

Delta R.T. 0.000 min

Lab File: BN033357.D

Acq: 13 Aug 2024 08:55




Tgt Ion:178 Resp: 14616


Ion Ratio Lower Upper

178 100

176 18.4 14.7 22.1

179 15.3 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.361 ng
RT: 18.998 min Scan# 1932
Delta R.T. -0.005 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 14952
Ion Ratio Lower Upper
212 100
106 16.6 12.8 19.2
104 9.7 7.1 10.7



#28
Fluoranthene
Concen: 0.373 ng
RT: 19.030 min Scan# 1939
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Tgt Ion:202 Resp: 20558
Ion Ratio Lower Upper
202 100
101 12.6 9.8 14.6
203 17.1 13.8 20.6





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.166 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN033357.D

Acq: 13 Aug 2024 08:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:240 Resp: 12286

Ion Ratio Lower Upper

240 100

120 18.3 13.0 19.6

236 29.3 23.2 34.8



#30

Pyrene

Concen: 0.424 ng

RT: 19.392 min Scan# 2017

Delta R.T. 0.000 min

Lab File: BN033357.D

Acq: 13 Aug 2024 08:55




Tgt Ion:202 Resp: 20717


Ion Ratio Lower Upper

202 100

200 20.8 16.7 25.1


203 17.7 14.2 21.4








#31
Terphenyl-d14
Concen: 0.403 ng
RT: 19.615 min Scan# 2065
Delta R.T. -0.005 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:244 Resp: 9354
Ion Ratio Lower Upper
244 100
212 10.5 7.7 11.5
122 20.3 14.0 21.0



#32
Benzo(a)anthracene
Concen: 0.366 ng
RT: 21.157 min Scan# 2381
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Tgt Ion:228 Resp: 16786
Ion Ratio Lower Upper
228 100
226 26.7 21.6 32.4
229 20.1 15.9 23.9









#33
 Chrysene
 Concen: 0.418 ng
 RT: 21.201 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN033357.D
 Acq: 13 Aug 2024 08:55

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion: 228 Resp: 19328




Ion	Ratio	Lower	Upper
228	100		
226	30.9	24.7	37.1
229	19.1	15.4	23.2




#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.325 ng
 RT: 21.130 min Scan# 2378
 Delta R.T. 0.000 min
 Lab File: BN033357.D
 Acq: 13 Aug 2024 08:55

Tgt Ion: 149 Resp: 6789

Ion	Ratio	Lower	Upper
149	100		
167	25.4	20.9	31.3
279	2.6	2.2	3.4









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.349 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN033357.D
 Acq: 13 Aug 2024 08:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:264 Resp: 11784
 Ion Ratio Lower Upper
 264 100
 260 27.4 21.6 32.4
 265 42.8 29.3 43.9



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.332 ng
 RT: 25.521 min Scan# 3585
 Delta R.T. -0.003 min
 Lab File: BN033357.D
 Acq: 13 Aug 2024 08:55


Tgt Ion:276 Resp: 16153
 Ion Ratio Lower Upper
 276 100
 138 33.2 24.9 37.3
 277 24.8 19.8 29.6








#37
Benzo(b)fluoranthene
Concen: 0.392 ng
RT: 22.703 min Scan# 2621
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:252 Resp: 17391
Ion Ratio Lower Upper
252 100
253 23.3 18.2 27.2
125 15.9 11.3 16.9



#38
Benzo(k)fluoranthene
Concen: 0.407 ng
RT: 22.747 min Scan# 2636
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55


Tgt Ion:252 Resp: 18446
Ion Ratio Lower Upper
252 100
253 23.0 17.9 26.9
125 15.9 11.3 16.9








#39
Benzo(a)pyrene
Concen: 0.358 ng
RT: 23.255 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:252 Resp: 13375
Ion Ratio Lower Upper
252 100
253 24.4 18.6 28.0
125 19.0 12.6 19.0#



#40
Dibenzo(a,h)anthracene
Concen: 0.330 ng
RT: 25.542 min Scan# 3592
Delta R.T. -0.006 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55


Tgt Ion:278 Resp: 12672
Ion Ratio Lower Upper
278 100
139 23.7 17.0 25.6
279 27.2 20.0 30.0





#41
Benzo(g,h,i)perylene
Concen: 0.343 ng
RT: 26.176 min Scan# 3809
Delta R.T. 0.000 min
Lab File: BN033357.D
Acq: 13 Aug 2024 08:55

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:	276	Resp:	14642
Ion Ratio	Lower	Upper	
276	100		
277	25.1	19.4	29.2
138	28.9	21.6	32.4

