

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081418\
 Data File : BN002329.D
 Acq On : 14 Aug 2018 11:22
 Operator : JU/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Quant Time: Aug 15 10:21:27 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN081318MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 13 23:38:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
2	1,4-Dioxane	0.453	0.488	-7.7	96	0.00
3 S	1,4-Dioxane-d8	0.424	0.435	-2.6	88	0.00
4	Benzaldehyde	1.260	1.298	-3.0	88	0.00
5 S	Phenol-d5	1.834	1.830	0.2	90	0.00
6	Phenol	1.868	1.887	-1.0	90	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.152	1.189	-3.2	90	0.00
8	Bis(2-Chloroethyl)ether	1.420	1.467	-3.3	90	0.00
9 S	2-Chlorophenol-d4	1.397	1.450	-3.8	90	0.00
10	2-Chlorophenol	1.406	1.448	-3.0	89	0.00
11	2-Methylphenol	1.409	1.419	-0.7	90	0.00
12	2,2'-oxybis(1-Chloropropane	2.351	2.522	-7.3	92	0.00
13 S	4-Methylphenol-d8	1.466	1.494	-1.9	90	0.00
14	Acetophenone	2.344	2.479	-5.8	91	0.00
15 P	N-Nitroso-di-n-propylamine	1.237	1.307	-5.7	90	0.00
16	4-Methylphenol	1.533	1.579	-3.0	91	0.00
17	Hexachloroethane	0.589	0.604	-2.5	90	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
19 S	Nitrobenzene-d5	0.159	0.164	-3.1	89	0.00
20	Nitrobenzene	0.381	0.398	-4.5	91	0.00
21	Isophorone	0.734	0.786	-7.1	91	0.00
22 S	2-Nitrophenol-d4	0.167	0.178	-6.6	89	0.00
23 C	2-Nitrophenol	0.176	0.186	-5.7	90	0.00
24	2,4-Dimethylphenol	0.361	0.389	-7.8	91	0.00
25	Bis(2-Chloroethoxy)methane	0.433	0.458	-5.8	90	0.00
26 S	2,4-Dichlorophenol-d3	0.308	0.329	-6.8	90	0.00
27 C	2,4-Dichlorophenol	0.298	0.321	-7.7	91	0.00
28	Naphthalene	1.029	1.096	-6.5	91	0.00
29 S	4-Chloroaniline-d4	0.333	0.329	1.2	84	0.00
30	4-Chloroaniline	0.336	0.336	0.0	84	0.00
31 C	Hexachlorobutadiene	0.188	0.199	-5.9	92	0.00
32	Caprolactam	0.112	0.115	-2.7	90	0.00
33 C	4-Chloro-3-methylphenol	0.343	0.373	-8.7	92	0.00
34	2-Methylnaphthalene	0.759	0.817	-7.6	93	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
36	1,2,4,5-Tetrachlorobenzene	0.606	0.641	-5.8	91	0.00
37	Hexachlorocyclopentadiene	0.368	0.360	2.2	89	0.00
38 C	2,4,6-Trichlorophenol	0.390	0.414	-6.2	90	0.00
39	2,4,5-Trichlorophenol	0.417	0.457	-9.6	96	0.00
40	1,1'-Biphenyl	1.601	1.721	-7.5	92	0.00
41	2-Chloronaphthalene	1.248	1.334	-6.9	92	0.00
42	2-Nitroaniline	0.398	0.438	-10.1	93	0.00
43 S	Dimethylphthalate-d6	1.658	1.798	-8.4	93	0.00
44	Dimethylphthalate	1.626	1.746	-7.4	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081418\
 Data File : BN002329.D
 Acq On : 14 Aug 2018 11:22
 Operator : JU/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Quant Time: Aug 15 10:21:27 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN081318MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 13 23:38:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.329	0.359	-9.1	92	0.00
46 S	Acenaphthylene-d8	2.034	2.223	-9.3	94	0.00
47	Acenaphthylene	1.936	2.091	-8.0	92	0.00
48	3-Nitroaniline	0.313	0.326	-4.2	89	0.00
49 C	Acenaphthene	1.366	1.469	-7.5	93	0.00
50	2,4-Dinitrophenol	0.197	0.176	10.7	87	0.00
51 S	4-Nitrophenol-d4	0.313	0.318	-1.6	89	0.00
52	4-Nitrophenol	0.238	0.256	-7.6	93	0.00
53	Dibenzofuran	1.927	2.085	-8.2	93	0.00
54	2,4-Dinitrotoluene	0.495	0.547	-10.5	94	0.00
55	2,3,4,6-Tetrachlorophenol	0.364	0.401	-10.2	94	0.00
56	Diethylphthalate	1.659	1.790	-7.9	92	0.00
57 S	Fluorene-d10	1.437	1.540	-7.2	93	0.00
58	Fluorene	1.577	1.725	-9.4	94	0.00
59	4-Chlorophenyl-phenylether	0.778	0.848	-9.0	94	0.00
60	4-Nitroaniline	0.404	0.432	-6.9	90	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.127	0.126	0.8	91	0.00
63	4,6-Dinitro-2-methylphenol	0.131	0.133	-1.5	92	0.00
64	N-Nitrosodiphenylamine	0.584	0.625	-7.0	94	0.00
65	4-Bromophenyl-phenylether	0.204	0.222	-8.8	96	0.00
66	Hexachlorobenzene	0.228	0.242	-6.1	93	0.00
67	Atrazine	0.214	0.228	-6.5	92	0.00
68 C	Pentachlorophenol	0.127	0.122	3.9	88	0.00
69	Phenanthrene	1.103	1.175	-6.5	94	0.00
70 S	Anthracene-d10	0.955	1.020	-6.8	94	0.00
71	Anthracene	1.118	1.204	-7.7	94	0.00
72	1,2,3,4-Tetrachlorobenzene	0.280	0.295	-5.4	93	0.00
73	Pentachlorobenzene	0.267	0.281	-5.2	94	0.00
74	Carbazole	0.992	1.087	-9.6	95	0.00
75	Di-n-butylphthalate	1.206	1.295	-7.4	93	0.00
76 C	Fluoranthene	1.255	1.388	-10.6	95	0.00
77 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
78 S	Pyrene-d10	0.957	1.013	-5.9	94	0.00
79	Pyrene	1.206	1.295	-7.4	96	0.00
80	Butylbenzylphthalate	0.511	0.542	-6.1	94	0.00
81	3,3'-Dichlorobenzidine	0.399	0.395	1.0	88	0.00
82	Benzo(a)anthracene	1.225	1.302	-6.3	95	0.00
83	Bis(2-ethylhexyl)phthalate	0.766	0.824	-7.6	95	0.00
84	Chrysene	1.160	1.223	-5.4	95	0.00
85 I	Perylene-d12	1.000	1.000	0.0	92	0.00
86	Di-n-octyl phthalate	1.271	1.329	-4.6	94	0.00
87	Benzo(b)fluoranthene	1.199	1.281	-6.8	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081418\
Data File : BN002329.D
Acq On : 14 Aug 2018 11:22
Operator : JU/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Aug 15 10:21:27 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN081318MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Aug 13 23:38:34 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.129	1.175	-4.1	92	0.00
89 S	Benzo(a)pyrene-d12	1.046	1.098	-5.0	96	0.00
90 C	Benzo(a)pyrene	1.142	1.204	-5.4	97	0.00
91	Indeno(1,2,3-cd)pyrene	1.359	1.441	-6.0	97	-0.01
92	Dibenzo(a,h)anthracene	1.136	1.209	-6.4	97	0.00
93	Benzo(g,h,i)perylene	1.134	1.194	-5.3	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0