


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081524\
Data File : BN033392.D
Acq On : 15 Aug 2024 09:43
Operator : MA/JU
Sample : DFTPP352
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP352

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081424.MA.M
Title : SVOA CALIBRATION
Last Update : Wed Aug 14 02:12:11 2024



AutoFind: Scans 2456, 2457, 2458; Background Corrected with Scan 2450

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.1	334836	PASS
68	69	0.00	2	1.7	6035	PASS
69	198	0.00	100	36.8	350959	PASS
70	69	0.00	2	0.5	1693	PASS
127	198	10	80	47.3	451115	PASS
197	198	0.00	2	0.3	3286	PASS
198	198	100	100	100.0	954453	PASS
199	198	5	9	6.6	62528	PASS
275	198	10	60	26.9	256363	PASS
365	198	1	100	3.5	33384	PASS
441	198	0.01	100	11.8	112869	PASS
442	198	50	100	76.2	726869	PASS
443	442	15	24	20.2	147003	PASS

m/z	Abundance
39.00	349.0
40.00	2095.0
41.00	577.0
43.00	337.0
44.00	10097.0
44.90	357.0
46.90	185.0
48.90	471.0
51.00	204.0
54.90	426.0
56.00	245.0
57.00	294.0
67.10	171.0
69.00	252.0
70.00	172.0
73.00	611.0
75.00	194.0
76.00	185.0
77.00	449.0
78.00	194.0
79.10	171.0
81.00	210.0
82.10	218.0
83.90	393.0
86.00	215.0
91.00	304.0
92.00	151.0
96.00	570.0
97.10	231.0
102.80	169.0
114.90	170.0
118.90	284.0
132.90	436.0
176.90	220.0
191.00	610.0
192.80	302.0
197.10	423.0
207.10	4736.0
208.00	1002.0
208.90	665.0
281.00	725.0
282.00	185.0

Instrument :
BNA_N
ClientSampleId :
DFTPP352

m/z	Abundance
36.00	380.0
37.00	1491.0
38.10	4421.0
39.10	25424.0
40.00	2642.0
41.00	1013.0
42.10	170.0
43.10	451.0
44.00	8002.0
45.00	857.0
46.90	232.0
48.20	263.0
49.10	2848.0
50.00	89672.0
51.00	356608.0
52.00	18832.0
53.00	1023.0
55.00	2017.0
56.00	11182.0
57.00	24128.0
57.90	1116.0
59.10	348.0
60.00	312.0
61.00	4392.0
62.00	5432.0
63.00	13705.0
64.00	2038.0
65.00	7435.0
66.00	537.0
66.90	506.0
68.00	6305.0
69.00	368000.0
70.00	1876.0
71.10	285.0
72.00	229.0
72.90	3161.0
74.00	36816.0
75.00	56656.0
76.00	20568.0
77.00	421248.0
78.00	29400.0
79.00	25568.0
80.00	20568.0
81.00	30056.0
82.00	7129.0
83.00	6869.0
83.90	842.0
85.00	4729.0
86.00	8022.0
87.00	3421.0
88.00	1694.0
88.90	654.0
89.90	222.0
91.00	7034.0
92.00	6995.0
93.00	46384.0
94.00	3617.0
95.00	847.0
96.00	2232.0
97.00	871.0
98.00	36880.0
99.00	29128.0
100.00	2726.0
101.00	16832.0
102.00	983.0
103.00	5474.0
104.00	10994.0
105.00	9391.0
106.00	3612.0
107.00	125552.0
108.00	18800.0
109.00	3316.0
110.00	237824.0
111.00	35336.0
112.00	4344.0
113.10	1232.0
114.00	363.0
114.90	742.0
116.00	7183.0
117.00	99200.0
118.00	8228.0
118.90	1022.0
119.90	1534.0
121.00	713.0
122.00	8412.0
123.00	13438.0
124.00	5705.0
125.00	5005.0
127.00	450752.0
128.00	35016.0
129.00	168640.0
130.00	14862.0

Instrument :
BNA_N
ClientSampleId :
DFTPP352

131.00	3027.0
132.00	1454.0
132.90	1355.0
133.90	5141.0
135.00	13968.0
136.00	5331.0
137.00	7956.0
137.90	1769.0
138.90	1030.0
140.00	1828.0
141.00	21152.0
142.00	6951.0
142.90	4929.0
144.00	1332.0
145.00	1003.0
146.00	3865.0
147.00	12102.0
148.00	25048.0
149.00	5424.0
150.00	1452.0
151.00	2987.0
151.80	1754.0
153.00	6654.0
154.00	5241.0
155.00	12660.0
156.00	18248.0
157.00	3745.0
157.90	4347.0
159.00	2756.0
160.00	7061.0
161.00	10187.0
162.00	3096.0
163.00	661.0
164.00	1186.0
165.00	8394.0
166.00	5950.0
167.00	45888.0
168.00	24688.0
169.00	3686.0
169.90	1393.0
171.00	1909.0
171.90	3518.0
173.00	5280.0
174.00	9576.0
175.00	17400.0
176.00	4390.0
177.00	9065.0
177.90	2874.0
178.90	34424.0
180.00	22440.0
181.00	10226.0
182.00	1521.0
182.90	988.0
184.00	2862.0
185.00	16212.0
186.00	120224.0
187.00	35792.0
188.00	3843.0
189.00	8052.0
190.00	1380.0
191.00	3867.0
192.00	9895.0
193.00	12114.0
194.00	2698.0
194.90	1447.0
196.00	24176.0
198.00	903680.0
198.90	58608.0
200.00	4766.0
201.50	4208.0
203.00	6207.0
204.00	32200.0
205.00	53912.0
206.00	235840.0
207.00	33584.0
208.00	8603.0
209.00	2959.0
210.20	3540.0
211.00	8464.0
213.00	903.0
214.00	244.0
214.90	2651.0
216.00	4916.0
217.00	61280.0
218.00	7373.0
219.00	818.0
221.10	34856.0
221.80	8171.0
223.00	13114.0
224.00	131520.0
225.00	31552.0
226.10	3473.0
227.00	54576.0
228.00	7450.0
229.00	12007.0
230.00	1649.0

Instrument :
BNA_N
ClientSampleId :
DFTPP352

231.00	4396.0
232.00	963.0
233.00	649.0
234.00	3615.0
235.00	3535.0
235.90	2434.0
237.00	5155.0
238.10	664.0
238.90	2064.0
240.00	1624.0
241.00	3281.0
242.00	6857.0
243.00	6380.0
244.00	98936.0
245.00	13226.0
246.00	19328.0
247.00	3801.0
247.90	717.0
249.00	4019.0
250.00	738.0
250.90	818.0
251.90	1073.0
253.00	2248.0
255.00	498304.0
256.00	73712.0
257.00	5466.0
258.00	28560.0
259.00	4517.0
259.90	939.0
260.90	810.0
262.90	218.0
263.90	735.0
265.00	11362.0
266.00	1695.0
267.10	226.0
268.00	202.0
269.00	150.0
270.00	432.0
270.90	1256.0
272.00	1469.0
273.00	14415.0
274.00	38064.0
275.00	222464.0
276.00	28800.0
277.00	18104.0
278.00	3164.0
279.00	836.0
281.10	1018.0
281.90	559.0
283.00	2002.0
284.00	1616.0
285.10	3458.0
285.90	484.0
287.00	154.0
287.90	235.0
288.80	653.0
290.00	625.0
290.90	376.0
292.00	703.0
293.00	3750.0
294.00	1195.0
294.90	1328.0
295.90	54592.0
297.00	8910.0
297.90	634.0
299.00	152.0
301.00	604.0
302.00	1108.0
303.00	6827.0
304.00	1944.0
304.80	304.0
307.80	749.0
309.10	564.0
310.00	771.0
311.00	198.0
311.80	186.0
312.90	524.0
314.10	2984.0
314.90	7286.0
316.00	3967.0
317.00	518.0
319.90	260.0
321.00	1975.0
322.00	903.0
323.00	19472.0
324.00	3615.0
325.00	482.0
326.00	367.0
327.00	3978.0
327.90	1831.0
328.90	399.0
332.00	1458.0
333.00	2035.0
334.00	11304.0
335.00	3142.0
335.90	429.0

Instrument :
BNA_N
ClientSampleId :
DFTPP352

339.00	244.0
340.00	263.0
341.00	2502.0
342.10	682.0
345.90	4135.0
347.00	891.0
350.00	170.0
350.90	363.0
352.00	5977.0
353.00	3890.0
354.00	5602.0
354.90	1359.0
358.90	346.0
365.00	24496.0
365.90	3419.0
366.90	197.0
369.90	650.0
371.00	1418.0
372.00	8778.0
373.00	2494.0
373.90	308.0
377.00	232.0
383.00	2818.0
384.00	931.0
385.00	230.0
389.90	1222.0
390.90	880.0
392.00	793.0
394.90	236.0
400.90	580.0
402.00	4234.0
403.00	6348.0
403.90	2209.0
405.00	436.0
415.00	209.0
421.00	5018.0
422.00	4296.0
423.00	33464.0
424.00	8158.0
425.00	1054.0
441.00	79264.0
442.10	540160.0
443.00	103600.0
444.00	9328.0
445.00	531.0

Instrument :
BNA_N
ClientSampleId :
DFTPP352

m/z	Abundance
36.10	226.0
37.00	1527.0
38.00	4693.0
39.00	26720.0
40.00	2785.0
41.10	937.0
42.00	256.0
43.00	354.0
44.00	7177.0
44.90	1036.0
46.90	182.0
47.90	312.0
49.00	3272.0
50.00	102880.0
51.00	393472.0
52.00	20072.0
53.00	922.0
55.00	2160.0
56.00	11540.0
57.00	27448.0
58.00	1262.0
59.00	337.0
60.00	349.0
61.00	5228.0
62.00	5615.0
63.00	15999.0
64.00	2363.0
65.00	7656.0
66.00	426.0
67.00	745.0
68.10	7315.0
69.00	420416.0
70.00	2177.0
70.90	322.0
71.90	252.0
73.10	3865.0
74.00	43024.0
75.00	65728.0
76.00	22632.0
77.00	477952.0
78.00	32376.0
79.00	30736.0
80.00	24088.0
81.00	35008.0
82.00	7542.0
83.00	8361.0
83.90	1009.0
85.00	5485.0
86.00	9858.0
87.00	4552.0
88.00	1552.0
88.90	851.0
91.00	7274.0
92.00	8949.0
93.00	55928.0
94.00	3826.0
95.00	793.0
96.00	2792.0
97.00	1007.0
98.00	40912.0
99.00	34648.0
100.00	3138.0
101.00	18792.0
102.00	980.0
103.00	6575.0
104.00	11537.0
105.00	11544.0
106.00	3851.0
107.00	150656.0
108.00	22688.0
109.00	4007.0
110.00	276032.0
111.00	41616.0
112.00	5321.0
113.00	1828.0
114.00	331.0
115.00	710.0
116.00	8134.0
117.00	123008.0
118.00	9201.0
119.00	1238.0
120.00	1619.0
120.90	615.0
122.00	10242.0
123.00	15902.0
124.00	7000.0
125.00	6577.0
127.00	542272.0
128.00	39512.0
129.00	207616.0
130.00	16760.0
131.00	3265.0

Instrument :
BNA_N
ClientSampleId :
DFTPP352

132.10	2030.0
133.00	1147.0
134.00	6158.0
135.00	16600.0
136.00	7116.0
137.00	9159.0
138.00	1675.0
138.90	1053.0
139.90	2550.0
141.00	26328.0
142.00	9081.0
143.00	6119.0
144.00	1744.0
144.90	1448.0
146.00	4902.0
147.00	13832.0
148.00	31720.0
149.00	6143.0
150.00	1941.0
151.00	3394.0
151.80	2292.0
153.00	9352.0
154.00	6329.0
155.00	15276.0
156.00	22904.0
157.00	3945.0
157.90	5239.0
159.00	3449.0
160.00	8725.0
161.00	12035.0
162.00	3919.0
163.00	1186.0
164.00	1590.0
164.90	10395.0
166.00	7061.0
167.00	53632.0
168.00	28824.0
169.00	4438.0
169.90	1465.0
170.90	2431.0
172.00	5115.0
173.00	6421.0
174.00	11335.0
175.00	21552.0
176.00	6272.0
177.00	10344.0
178.00	3821.0
179.00	40472.0
180.00	28432.0
181.00	13166.0
182.00	2041.0
182.80	1069.0
184.00	3475.0
185.00	20224.0
186.00	155904.0
187.00	43728.0
188.00	5210.0
189.00	9969.0
190.00	1475.0
191.00	4985.0
192.00	13088.0
193.00	14861.0
194.00	3204.0
195.00	1619.0
196.00	30640.0
197.00	6853.0
197.90	1127936.0
199.00	75192.0
200.00	6506.0
201.40	4591.0
203.00	7458.0
204.00	41264.0
205.00	70784.0
206.00	301632.0
207.00	43760.0
208.00	10385.0
209.00	3644.0
210.10	5040.0
211.00	12413.0
211.90	2333.0
213.00	975.0
213.90	458.0
215.00	3351.0
216.00	6727.0
217.00	81472.0
218.00	10156.0
219.00	1008.0
221.00	45448.0
221.80	11006.0
223.00	18288.0
224.00	170304.0
225.00	44448.0
226.00	4449.0
227.00	70464.0
228.00	10648.0
229.00	14515.0

Instrument :
BNA_N
ClientSampleId :
DFTPP352

230.00	1990.0
231.00	5445.0
232.00	1092.0
233.00	1078.0
234.00	4361.0
235.00	5177.0
236.00	2900.0
237.00	5908.0
238.00	731.0
238.90	2522.0
240.00	2170.0
241.00	3602.0
242.00	9036.0
243.00	9313.0
244.00	128344.0
245.00	18368.0
246.00	26128.0
247.00	5650.0
248.00	1214.0
249.00	4837.0
250.00	971.0
250.90	1121.0
252.00	1176.0
253.00	2739.0
255.00	672704.0
256.00	98720.0
257.00	7180.0
258.00	38536.0
259.00	6004.0
260.00	1262.0
261.00	1281.0
261.80	192.0
263.00	383.0
264.10	938.0
265.00	16648.0
265.90	2034.0
267.00	306.0
267.80	179.0
269.90	929.0
271.00	1492.0
272.10	1566.0
273.00	20312.0
274.00	50120.0
275.00	302080.0
276.00	42496.0
277.00	24752.0
278.00	4062.0
279.00	854.0
281.10	827.0
282.10	952.0
283.00	2692.0
284.00	2054.0
285.00	4949.0
286.00	968.0
287.90	307.0
289.00	1125.0
289.90	974.0
291.00	740.0
292.00	1213.0
293.00	5947.0
294.00	1549.0
295.10	1781.0
296.00	85056.0
297.00	10622.0
297.90	750.0
298.80	160.0
301.00	1178.0
302.00	1540.0
303.10	9608.0
304.00	2572.0
305.10	419.0
306.00	189.0
307.90	1057.0
309.00	786.0
310.00	1190.0
310.90	255.0
311.90	331.0
312.90	868.0
314.10	4387.0
315.00	8728.0
316.00	5604.0
317.00	1134.0
318.90	258.0
320.10	214.0
320.90	2746.0
321.90	1312.0
323.00	27232.0
324.00	4743.0
325.00	352.0
326.10	571.0
326.90	5160.0
327.90	2916.0
328.90	405.0
330.00	188.0
332.00	1924.0
333.00	2462.0

Instrument :
BNA_N
ClientSampleId :
DFTPP352

334.00	18600.0
335.00	4772.0
336.00	484.0
339.00	531.0
340.10	385.0
341.00	3427.0
342.00	1120.0
343.10	175.0
346.00	5734.0
347.00	1258.0
347.90	158.0
350.00	289.0
351.00	488.0
352.00	7934.0
353.00	6166.0
354.00	8357.0
355.00	1908.0
356.20	208.0
358.10	159.0
359.00	516.0
360.80	165.0
365.00	37904.0
365.90	5240.0
367.00	384.0
370.00	784.0
371.00	2056.0
372.00	15922.0
373.00	3700.0
373.90	547.0
377.00	378.0
382.90	4441.0
384.00	1714.0
384.90	434.0
390.00	2214.0
390.90	1485.0
392.10	1106.0
393.00	288.0
396.80	175.0
400.90	921.0
402.00	6616.0
403.00	9506.0
404.00	3253.0
405.00	549.0
410.00	331.0
414.80	331.0
419.80	167.0
421.00	8032.0
422.00	6905.0
423.00	52240.0
424.00	12449.0
425.00	1243.0
441.00	133120.0
442.00	884480.0
443.00	180416.0
444.00	14050.0
445.10	864.0

Instrument :
BNA_N
ClientSampleId :
DFTPP352

m/z	Abundance
35.00	320.0
37.00	1032.0
38.00	2984.0
39.00	17864.0
40.00	2642.0
41.00	986.0
42.00	280.0
43.00	325.0
44.00	9044.0
45.00	715.0
47.00	170.0
49.00	2129.0
50.00	64920.0
51.00	255040.0
52.00	12922.0
53.00	581.0
55.00	1516.0
56.00	7973.0
57.00	17248.0
58.00	886.0
59.00	355.0
59.90	333.0
61.00	3215.0
62.00	3477.0
63.00	9698.0
64.00	1637.0
65.00	5317.0
66.00	402.0
67.00	780.0
68.00	4486.0
69.00	265216.0
70.00	1541.0
71.00	212.0
71.90	151.0
73.00	2613.0
74.00	26568.0
75.00	43008.0
76.00	14627.0
77.00	311616.0
78.00	19888.0
79.00	19800.0
80.00	14717.0
81.00	21352.0
82.00	4912.0
83.00	5062.0
84.00	922.0
85.00	3446.0
86.00	6560.0
87.00	2920.0
87.90	1017.0
89.00	751.0
91.00	4791.0
92.00	5424.0
93.00	36432.0
94.00	2569.0
95.00	700.0
96.00	2261.0
96.90	758.0
98.00	27400.0
99.00	22128.0
100.00	1879.0
101.00	12365.0
101.90	865.0
103.00	4283.0
104.00	8846.0
105.00	7525.0
106.00	2557.0
107.00	95720.0
108.00	15412.0
109.00	2517.0
110.00	186304.0
111.00	27776.0
112.00	3291.0
113.00	783.0
114.10	195.0
115.00	676.0
116.00	5488.0
117.00	80216.0
118.00	5592.0
119.00	972.0
120.00	1435.0
121.10	474.0
122.00	6878.0
123.00	10294.0
124.00	4528.0
125.00	4394.0
126.00	1555.0
127.00	360320.0
128.00	26576.0
129.00	134208.0
130.00	11448.0
131.00	2437.0

Instrument :
BNA_N
ClientSampleId :
DFTPP352

131.90	962.0
133.00	858.0
134.00	4295.0
135.00	11181.0
136.00	4550.0
137.00	6222.0
137.90	1275.0
138.90	789.0
140.00	1680.0
141.00	18624.0
142.00	6248.0
143.00	4319.0
144.00	1163.0
144.90	1238.0
146.00	3291.0
147.00	9095.0
147.90	21968.0
149.00	4473.0
149.90	1250.0
151.10	2359.0
151.90	1351.0
152.90	5375.0
154.00	3987.0
155.00	10096.0
156.00	16512.0
157.00	2768.0
157.90	3771.0
158.90	2884.0
160.00	6492.0
161.00	9488.0
161.90	2749.0
162.90	779.0
164.10	709.0
165.00	7110.0
166.00	5436.0
167.00	39056.0
168.00	20160.0
169.00	2948.0
170.00	1078.0
170.90	1644.0
171.90	3590.0
173.00	4467.0
174.00	8664.0
175.00	16014.0
176.00	3944.0
177.00	7640.0
178.00	2424.0
179.00	29152.0
180.00	19968.0
181.00	9225.0
182.00	1681.0
183.00	851.0
183.90	2215.0
185.00	14515.0
186.00	114360.0
187.00	33416.0
188.00	3378.0
189.00	6862.0
190.00	1450.0
191.00	4204.0
192.00	9306.0
193.00	10602.0
194.00	2031.0
195.00	1162.0
196.00	22800.0
197.00	4273.0
197.90	831744.0
199.00	53784.0
200.00	3857.0
201.40	3753.0
203.00	5989.0
204.00	30904.0
205.00	51632.0
206.00	227584.0
207.00	30856.0
208.00	8369.0
209.00	2593.0
210.30	3258.0
211.00	8610.0
213.00	1015.0
213.90	280.0
215.00	2425.0
216.00	4985.0
217.00	59104.0
218.00	7526.0
219.10	641.0
221.00	33720.0
221.80	8341.0
223.00	13266.0
224.00	128888.0
225.00	32816.0
226.00	3300.0
227.00	52472.0
228.00	7573.0
229.00	11017.0
230.00	1265.0

Instrument :
BNA_N
ClientSampleId :
DFTPP352

231.00	4982.0
232.00	961.0
232.90	955.0
234.00	3608.0
234.90	3824.0
236.00	2587.0
237.00	3803.0
238.00	526.0
238.90	1879.0
240.00	1590.0
241.00	2723.0
242.00	6245.0
243.00	6859.0
244.00	98880.0
245.10	13553.0
246.00	20032.0
247.00	3585.0
247.80	764.0
249.00	3539.0
250.00	680.0
250.90	859.0
252.00	933.0
253.00	2045.0
255.00	529536.0
256.00	76808.0
257.00	6159.0
258.00	30360.0
259.00	4467.0
260.00	891.0
261.00	771.0
262.90	328.0
264.00	747.0
265.00	12681.0
266.00	1674.0
267.00	407.0
268.00	171.0
269.90	652.0
270.90	1345.0
272.00	1252.0
273.00	16392.0
274.00	41176.0
275.00	244544.0
276.00	31360.0
277.00	18280.0
277.90	2834.0
279.10	723.0
281.10	837.0
281.90	676.0
283.00	2036.0
284.10	1606.0
285.00	3945.0
286.00	743.0
287.90	164.0
288.90	878.0
290.00	743.0
291.00	571.0
292.10	868.0
293.00	4812.0
294.00	1497.0
294.90	1430.0
296.00	75384.0
297.00	9477.0
298.00	843.0
301.00	769.0
302.00	1292.0
303.00	8167.0
304.00	2147.0
305.00	303.0
308.00	1064.0
308.90	654.0
310.00	1125.0
311.00	250.0
311.90	166.0
312.90	785.0
314.00	3540.0
314.90	8438.0
316.00	4375.0
317.00	910.0
318.90	157.0
319.90	342.0
321.00	2524.0
321.90	1062.0
323.00	22968.0
324.00	4781.0
325.00	425.0
325.90	537.0
327.00	4720.0
328.00	2504.0
329.00	430.0
332.00	1663.0
333.00	2526.0
334.00	15760.0
335.10	4135.0
335.90	413.0
338.90	432.0
339.90	344.0

Instrument :
BNA_N
ClientSampleId :
DFTPP352

341.00	2620.0
341.90	857.0
345.90	5497.0
347.00	800.0
347.90	156.0
350.00	237.0
351.00	366.0
352.00	7615.0
353.00	5055.0
354.00	7709.0
355.00	1468.0
358.00	157.0
359.00	526.0
361.00	169.0
365.00	37752.0
365.90	5245.0
366.90	428.0
369.90	852.0
371.00	2584.0
372.00	14662.0
373.00	3319.0
374.00	329.0
377.00	319.0
382.90	3734.0
383.90	928.0
385.00	223.0
390.00	1985.0
391.00	1307.0
392.10	1030.0
396.80	206.0
401.00	800.0
402.00	5760.0
403.00	7633.0
404.00	3303.0
404.90	392.0
410.00	298.0
415.00	364.0
421.00	6867.0
422.10	6726.0
423.00	49704.0
424.00	10604.0
424.90	1136.0
440.10	677.0
441.00	126224.0
442.00	755968.0
443.00	156992.0
444.00	13716.0
445.00	790.0

Instrument :
BNA_N
ClientSampleId :
DFTPP352