

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081622\
 Data File : BN021271.D
 Acq On : 17 Aug 2022 02:30
 Operator : CG/JU
 Sample : N4171-01
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP-100-20220810

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 08/17/2022
 Supervised By : Jagrut Upadhyay 08/17/2022

Quant Time: Aug 17 05:18:02 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 16 09:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.798	152	3286	0.400	ng	0.01
7) Naphthalene-d8	10.592	136	10269	0.400	ng	# 0.01
13) Acenaphthene-d10	14.440	164	5373	0.400	ng	0.01
19) Phenanthrene-d10	17.193	188	12063	0.400	ng	0.01
29) Chrysene-d12	21.395	240	7240	0.400	ng	0.00
35) Perylene-d12	23.726	264	5228	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.393	112	746	0.091	ng	0.01
5) Phenol-d6	7.097	99	241m	0.025	ng	0.12
8) Nitrobenzene-d5	9.023	82	1179	0.253	ng	0.06
11) 2-Methylnaphthalene-d10	12.202	152	4773	0.267	ng	0.02
14) 2,4,6-Tribromophenol	15.952	330	320	0.182	ng	0.01
15) 2-Fluorobiphenyl	13.082	172	5132	0.240	ng	0.02
27) Fluoranthene-d10	19.227	212	12846	0.392	ng	0.00
31) Terphenyl-d14	19.826	244	8151	0.413	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.219	88	17559	3.830	ng	92
34) Bis(2-ethylhexyl)phtha...	21.306	149	863	0.054	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081622\
Data File : BN021271.D
Acq On : 17 Aug 2022 02:30
Operator : CG/JU
Sample : N4171-01
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP-100-20220810

Quant Time: Aug 17 05:18:02 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 16 09:19:59 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 08/17/2022
Supervised By :Jagrut Upadhyay 08/17/2022

