


n-Nitrosodimethylamine


Response Ratio



$R = -2.402e-002 A^2 + 8.314e-001 A - 4.395e-001$
 Coef of Det (r^2) = 0.998626 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN081622.M
 Calibration Table Last Updated: Tue Aug 16 09:19:59 2022

Nitrobenzene-d5

Response Ratio




$$\text{Response} = 3.594\text{e-}001 * \text{Amt} - 1.129\text{e-}001$$

Coef of Det (r^2) = 0.999343 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN081622.M

Calibration Table Last Updated: Tue Aug 16 09:19:59 2022

4,6-Dinitro-2-methylphenol



$\text{Response} = 6.453\text{e-}002 * \text{Amt} - 4.085\text{e-}002$
 Coef of Det (r^2) = 0.990211 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN081622.M
 Calibration Table Last Updated: Tue Aug 16 09:19:59 2022