




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081622\
Data File : BN021264.D
Acq On : 16 Aug 2022 21:01
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 17 06:35:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 17 06:35:06 2022
Response via : Initial Calibration



TIC: BN021264.D\data.ms

(70) Pentachlorophenol (C)




16.839min (0.000) 15569.65 ng

response 212727

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	66.41
264.00	61.60	61.84
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081622\
Data File : BN021264.D
Acq On : 16 Aug 2022 21:01
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 17 06:35:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 17 06:35:06 2022
Response via : Initial Calibration



TIC: BN021264.D\data.ms

(77) Benzidine

19.433min (0.000) 0.00 ng

response 780702

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.13
183.00	13.20	11.53
0.00	0.00	0.00