


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081720\
 Data File : BN011445.D
 Acq On : 17 Aug 2020 09:16
 Operator : CG/JU
 Sample : DFTPP66
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP66

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN081420MA.M
 Title : SVOA CALIBRATION
 Last Update : Wed Aug 19 02:19:14 2020



AutoFind: Scans 2433, 2434, 2435; Background Corrected with Scan 2428

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	20.6	87141	PASS
68	69	0.00	2	1.8	2702	PASS
69	198	0.00	100	36.2	153144	PASS
70	69	0.00	2	0.5	770	PASS
127	198	10	80	44.3	187541	PASS
197	198	0.00	2	0.4	1807	PASS
198	198	100	100	100.0	423616	PASS
199	198	5	9	7.1	29938	PASS
275	198	10	60	29.0	122637	PASS
365	198	1	100	5.0	21277	PASS
441	198	0.01	100	12.0	50632	PASS
442	198	50	100	69.4	294144	PASS
443	442	15	24	21.0	61853	PASS

Scan 2428 (17.269 min): BN011445.D
DFTPP66

m/z	Abundance
40.00	530.0
43.90	810.0
72.90	652.0
147.00	359.0
207.00	1188.0
281.00	673.0
340.90	303.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

m/z	Abundance
37.00	685.0
38.00	1273.0
39.00	7269.0
40.00	531.0
44.00	882.0
48.90	957.0
50.00	25544.0
51.00	88888.0
52.00	4342.0
55.00	461.0
55.90	3027.0
56.90	6489.0
58.10	392.0
61.00	1544.0
62.00	1612.0
62.90	5290.0
64.00	781.0
65.00	2046.0
67.00	387.0
68.00	2557.0
68.90	146816.0
70.00	825.0
73.00	1966.0
74.00	18384.0
75.00	27192.0
76.00	9258.0
77.00	175040.0
78.00	12347.0
79.00	13549.0
80.00	9079.0
80.90	11153.0
81.90	3219.0
82.90	2702.0
84.00	469.0
84.90	2293.0
85.90	4035.0
86.90	1595.0
87.80	580.0
88.90	348.0
91.00	3040.0
92.00	3794.0
93.00	24656.0
94.00	1802.0
95.00	507.0
96.10	785.0
97.00	399.0
97.90	22776.0
99.00	14280.0
100.00	1523.0
100.90	7522.0
102.00	437.0
102.90	2184.0
104.00	5186.0
105.00	4907.0
105.90	1348.0
107.00	54992.0
108.00	7542.0
108.90	1437.0
110.00	89208.0
111.00	14673.0
112.00	2011.0
112.90	637.0
116.00	3438.0
117.00	65544.0
117.90	4845.0
119.90	897.0
121.00	301.0
121.90	4029.0
123.00	5798.0
124.00	2581.0
125.00	2707.0
127.00	172800.0
128.00	13344.0
129.00	74336.0
129.90	7149.0
131.00	1301.0
132.00	880.0
134.00	2161.0
135.00	5668.0
135.90	3165.0
137.00	3256.0
137.90	514.0
138.90	359.0
139.80	1190.0
140.90	10544.0
142.00	3453.0
142.90	2683.0
143.80	673.0
144.20	530.0
145.00	682.0
146.00	2232.0
147.00	6222.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

148.00	15946.0
148.90	2511.0
149.90	736.0
151.10	1391.0
152.00	668.0
152.90	3240.0
154.00	2114.0
155.00	6024.0
156.00	8231.0
157.00	2244.0
157.90	2333.0
158.90	1335.0
160.00	3296.0
161.00	4144.0
161.80	1535.0
163.00	350.0
163.90	462.0
165.00	3972.0
165.90	2894.0
167.00	23856.0
168.00	15506.0
169.00	2128.0
169.90	634.0
170.90	994.0
172.00	2250.0
172.90	2333.0
174.00	4244.0
175.00	7927.0
176.00	1855.0
176.90	5503.0
177.90	1370.0
178.90	15892.0
180.00	11254.0
180.90	5129.0
181.90	773.0
182.90	365.0
184.00	1714.0
185.00	8086.0
186.00	59536.0
187.00	18272.0
187.90	1965.0
188.90	4093.0
189.90	730.0
191.00	2325.0
192.10	5346.0
192.90	6628.0
193.90	1424.0
194.90	613.0
195.90	9447.0
196.90	2439.0
197.90	355136.0
198.90	24464.0
199.90	2116.0
201.50	1738.0
202.90	2826.0
203.90	15432.0
205.00	24968.0
206.00	108504.0
207.00	16576.0
208.00	4205.0
208.90	1714.0
210.00	1624.0
210.90	4494.0
211.70	1008.0
215.00	1495.0
216.00	3001.0
216.90	28960.0
218.00	3936.0
218.90	516.0
219.90	399.0
221.00	12557.0
221.90	3834.0
222.90	7162.0
224.00	57760.0
225.00	16760.0
226.00	1896.0
227.00	28672.0
228.00	4812.0
229.00	6735.0
230.00	768.0
231.00	1766.0
232.10	529.0
233.00	740.0
234.00	1436.0
234.90	2114.0
235.90	1419.0
237.00	2525.0
237.70	674.0
238.90	627.0
240.00	852.0
241.00	1737.0
241.90	4112.0
243.00	3171.0
244.00	45000.0
245.00	5803.0
246.00	11127.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

247.00	1915.0
247.90	597.0
248.90	1444.0
250.90	538.0
253.00	923.0
254.10	1961.0
255.00	235136.0
256.00	34360.0
256.90	2768.0
258.00	14589.0
259.00	1955.0
260.00	598.0
264.00	675.0
265.00	7658.0
266.00	1333.0
269.80	565.0
270.90	518.0
271.90	767.0
272.90	6962.0
274.00	19840.0
275.00	97288.0
276.00	13942.0
276.90	8793.0
278.10	2282.0
280.90	726.0
282.00	315.0
282.90	1396.0
284.00	731.0
285.10	1840.0
289.90	477.0
290.90	530.0
292.00	422.0
292.90	2011.0
294.00	566.0
296.00	35200.0
297.00	5476.0
301.10	386.0
301.90	649.0
302.90	3433.0
304.10	678.0
304.80	463.0
307.80	534.0
309.90	493.0
313.00	503.0
314.00	1665.0
314.90	2813.0
316.00	1559.0
316.80	330.0
320.90	803.0
323.00	8171.0
324.00	1309.0
327.00	1769.0
327.90	949.0
331.90	709.0
333.00	959.0
334.00	6193.0
334.90	1483.0
336.10	393.0
340.00	310.0
341.00	1232.0
345.90	2006.0
346.80	476.0
352.00	3118.0
353.00	1748.0
354.00	3013.0
355.00	603.0
364.90	15329.0
366.00	1541.0
370.70	569.0
371.00	398.0
372.00	4092.0
373.00	1063.0
382.90	1197.0
383.80	318.0
389.90	497.0
391.10	382.0
400.80	319.0
401.90	1832.0
402.90	2764.0
404.00	1177.0
420.90	1805.0
422.00	1651.0
423.00	11463.0
424.00	3440.0
425.00	462.0
440.10	535.0
440.90	32536.0
442.00	183296.0
443.00	37688.0
444.00	3804.0
444.80	425.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

m/z	Abundance
37.00	718.0
38.10	1600.0
39.00	9065.0
39.90	592.0
44.00	1040.0
48.80	766.0
49.20	1155.0
50.00	29320.0
51.00	106792.0
52.00	6287.0
54.90	639.0
55.90	3332.0
56.90	6899.0
58.00	641.0
60.90	1789.0
61.90	2199.0
63.00	6448.0
63.90	1183.0
65.00	2793.0
65.90	358.0
68.00	2919.0
68.90	192512.0
70.00	950.0
73.10	2963.0
74.00	24192.0
75.00	32656.0
76.00	11482.0
77.00	221888.0
78.00	15040.0
78.90	17368.0
80.00	12492.0
81.00	16784.0
81.90	4067.0
83.00	3127.0
83.90	525.0
85.00	2511.0
86.00	5197.0
86.90	1692.0
87.90	1134.0
89.00	319.0
91.00	3557.0
91.90	4356.0
92.90	30040.0
94.00	2063.0
95.00	616.0
96.00	1371.0
96.90	550.0
98.00	27120.0
99.00	15796.0
99.90	1725.0
100.90	9519.0
102.00	373.0
103.00	2781.0
103.90	6315.0
105.00	6096.0
106.00	2305.0
107.00	70528.0
108.00	10599.0
109.00	1881.0
109.90	115104.0
111.00	19000.0
112.00	2507.0
113.00	1141.0
115.00	380.0
116.00	4650.0
117.00	80912.0
118.00	5595.0
118.90	817.0
120.00	983.0
121.20	614.0
121.90	4580.0
122.90	6751.0
123.90	3600.0
125.00	2802.0
127.00	229568.0
128.00	18384.0
129.00	95448.0
129.90	9471.0
130.90	1969.0
132.00	1181.0
133.00	975.0
134.00	2363.0
135.00	9089.0
136.00	3682.0
137.00	4189.0
138.00	786.0
139.00	566.0
140.00	1448.0
141.00	13264.0
142.00	4542.0
143.00	3067.0
143.90	794.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

144.90	639.0
146.00	2925.0
147.00	7401.0
148.00	20592.0
149.00	4828.0
150.00	1196.0
151.10	1847.0
151.70	1120.0
153.00	4710.0
154.00	3307.0
155.00	8403.0
156.00	10848.0
157.00	2133.0
157.90	3285.0
158.90	1693.0
160.00	4354.0
161.00	7379.0
162.00	1751.0
163.00	719.0
164.00	550.0
164.90	5110.0
166.00	4033.0
167.00	35432.0
168.00	19504.0
169.00	2157.0
170.00	1087.0
170.80	1444.0
171.90	2786.0
173.10	3650.0
174.00	5746.0
175.00	10408.0
176.00	3204.0
176.90	6422.0
177.80	2045.0
178.90	23112.0
180.00	15341.0
180.90	7658.0
182.00	1665.0
183.00	395.0
183.90	1872.0
185.00	11815.0
186.00	80440.0
187.00	25152.0
188.00	2427.0
188.90	4733.0
189.90	1091.0
191.00	3312.0
192.00	7877.0
192.90	7776.0
194.00	1478.0
194.90	635.0
195.20	653.0
196.00	15072.0
197.90	513152.0
198.90	38280.0
200.00	3739.0
201.50	1476.0
202.90	4949.0
204.00	23656.0
205.00	40456.0
206.00	161920.0
207.00	21912.0
208.00	4932.0
209.00	2423.0
209.90	3141.0
211.00	6098.0
211.80	1132.0
212.00	1202.0
212.90	604.0
214.90	1947.0
216.00	4455.0
217.00	43688.0
218.00	5785.0
219.00	499.0
219.90	913.0
221.00	16944.0
221.70	5676.0
223.00	9574.0
224.00	90176.0
225.00	25184.0
226.00	2630.0
227.00	44784.0
228.00	6173.0
229.00	8021.0
230.10	1099.0
231.00	2508.0
232.20	450.0
233.00	865.0
233.90	2473.0
234.90	2927.0
236.00	2028.0
236.90	4123.0
238.00	713.0
239.00	1552.0
240.00	1257.0
241.00	2744.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

242.00	4862.0
243.00	5172.0
244.00	66512.0
245.00	8982.0
246.00	16584.0
247.00	2870.0
248.00	710.0
249.00	2442.0
250.00	555.0
250.90	611.0
252.00	811.0
253.00	1673.0
255.00	361088.0
256.00	50160.0
256.90	3506.0
258.00	23256.0
258.90	4223.0
259.90	619.0
261.00	596.0
264.10	753.0
264.90	9610.0
265.90	1866.0
270.80	730.0
271.90	1357.0
273.00	10394.0
274.00	26664.0
275.00	151680.0
276.00	16952.0
276.90	16720.0
277.90	2432.0
279.00	741.0
281.10	513.0
281.90	533.0
282.90	1627.0
283.90	985.0
285.00	2121.0
289.00	548.0
289.80	308.0
291.00	569.0
292.10	645.0
292.90	2810.0
293.90	879.0
294.90	918.0
296.00	54888.0
297.00	6961.0
298.10	588.0
300.90	659.0
301.90	931.0
303.00	5259.0
304.00	1496.0
307.90	1144.0
308.80	585.0
309.90	759.0
313.00	556.0
313.90	2631.0
314.90	5056.0
316.00	2782.0
317.00	561.0
320.90	1374.0
322.00	666.0
323.00	14775.0
324.00	2160.0
325.90	622.0
327.00	2572.0
327.90	1643.0
331.90	1555.0
333.00	1809.0
334.00	9378.0
335.00	2142.0
335.90	465.0
338.80	389.0
339.90	395.0
340.90	1840.0
342.00	317.0
345.90	2957.0
347.00	801.0
351.00	423.0
352.00	4460.0
353.00	3793.0
354.00	4430.0
355.00	1295.0
364.90	23784.0
365.90	2715.0
367.20	343.0
370.90	1253.0
371.90	7744.0
373.00	2227.0
374.00	307.0
383.00	1641.0
383.90	582.0
384.90	375.0
390.00	1052.0
391.00	518.0
391.90	404.0
400.90	692.0
401.90	2243.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

402.90	4425.0
404.00	1872.0
405.00	349.0
420.90	3495.0
421.90	3395.0
422.90	23112.0
423.90	6918.0
425.00	762.0
441.00	61272.0
442.00	353856.0
443.00	72984.0
444.00	6522.0
444.90	548.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

m/z	Abundance
37.10	486.0
38.10	972.0
39.10	5768.0
39.90	568.0
44.00	825.0
45.00	385.0
49.10	734.0
50.00	20248.0
51.00	65744.0
52.10	3897.0
54.80	416.0
56.00	2697.0
56.90	5208.0
61.00	1384.0
62.00	1940.0
62.90	3602.0
64.10	595.0
64.90	1651.0
68.10	2631.0
68.90	120104.0
70.00	536.0
73.00	2452.0
74.00	15006.0
75.00	22496.0
76.00	7656.0
77.00	144704.0
78.00	10648.0
79.00	10961.0
79.90	7013.0
81.00	11587.0
82.00	2190.0
83.00	2088.0
84.10	544.0
84.90	1505.0
86.00	3683.0
86.90	1409.0
88.00	398.0
89.00	399.0
90.90	2582.0
92.00	3037.0
93.00	22512.0
94.00	1114.0
94.90	565.0
96.00	1099.0
97.90	17664.0
99.00	12707.0
99.80	892.0
101.00	6049.0
102.80	1544.0
103.90	4438.0
104.90	3664.0
106.00	1157.0
106.90	49128.0
108.00	7261.0
110.00	77872.0
111.00	12580.0
111.90	1217.0
112.90	580.0
116.00	3115.0
116.90	57984.0
118.00	4110.0
118.80	415.0
119.10	348.0
119.90	756.0
120.90	342.0
121.90	2803.0
123.00	5991.0
124.00	2670.0
125.00	2502.0
127.00	160256.0
128.00	12982.0
129.00	70184.0
130.00	5751.0
130.90	1168.0
132.00	628.0
132.30	634.0
132.70	441.0
133.90	2629.0
134.90	5503.0
136.00	2848.0
137.00	3690.0
137.90	732.0
139.00	348.0
140.00	917.0
141.00	10981.0
141.90	2434.0
142.90	1793.0
143.80	463.0
145.00	339.0
146.00	1645.0
147.00	5141.0
147.90	14516.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

149.00	2338.0
150.00	753.0
150.90	1634.0
151.50	885.0
152.00	629.0
153.00	3149.0
154.00	2961.0
155.00	5785.0
156.00	8480.0
157.10	1086.0
158.00	2039.0
158.80	1206.0
160.00	3697.0
161.00	5235.0
162.00	1545.0
162.80	508.0
164.10	768.0
164.90	3957.0
165.90	3560.0
167.00	27992.0
168.00	15393.0
168.90	1622.0
169.80	458.0
171.00	728.0
171.90	2066.0
172.90	2026.0
174.00	3569.0
174.90	8697.0
176.00	2625.0
176.90	3598.0
178.10	1200.0
178.90	19072.0
180.00	12921.0
181.10	5287.0
182.10	1133.0
182.80	466.0
184.00	1607.0
185.00	7822.0
186.00	58792.0
187.00	17352.0
188.00	1330.0
188.90	4084.0
189.90	706.0
191.00	2502.0
192.00	5463.0
193.00	7497.0
193.90	1579.0
195.00	795.0
196.00	13248.0
197.00	2984.0
197.90	402560.0
198.90	27072.0
199.90	2493.0
201.40	1618.0
202.90	3765.0
204.00	18728.0
205.00	30232.0
206.00	123576.0
207.00	16664.0
207.90	3902.0
209.00	1908.0
210.10	2959.0
211.00	5240.0
211.70	1146.0
212.90	371.0
215.00	2101.0
216.00	3293.0
216.90	35176.0
218.00	4724.0
219.00	580.0
221.00	14718.0
222.00	4817.0
223.00	8012.0
224.00	64280.0
225.00	17856.0
226.00	2342.0
227.00	31048.0
228.00	4162.0
229.00	6806.0
229.90	1047.0
231.00	2555.0
232.00	568.0
233.00	631.0
233.90	2716.0
234.90	1950.0
236.00	1965.0
237.00	2565.0
238.90	1317.0
239.80	1093.0
240.90	2353.0
242.00	4046.0
243.00	3557.0
244.00	54464.0
244.90	5982.0
246.00	13380.0
247.00	3459.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

248.00	661.0
248.90	2429.0
249.80	610.0
251.00	709.0
251.80	406.0
252.90	1232.0
254.00	2584.0
255.00	285440.0
256.00	40648.0
257.00	2955.0
258.00	18320.0
259.00	2703.0
259.80	451.0
261.00	451.0
264.10	632.0
265.00	8102.0
265.80	935.0
269.90	344.0
271.00	897.0
271.80	612.0
273.00	7626.0
274.00	22616.0
275.00	118944.0
276.00	15622.0
277.00	13063.0
278.00	2565.0
279.00	528.0
281.00	649.0
282.00	702.0
283.00	1043.0
284.00	1103.0
284.90	2310.0
286.00	451.0
288.80	576.0
289.90	371.0
292.10	669.0
293.00	3052.0
294.00	671.0
295.00	1072.0
295.90	41168.0
296.90	6343.0
300.90	675.0
301.90	1032.0
303.00	5140.0
304.00	1329.0
305.00	339.0
308.10	567.0
309.00	442.0
310.00	326.0
312.80	321.0
313.90	1853.0
315.00	5508.0
316.00	2778.0
317.00	507.0
321.00	1599.0
321.90	665.0
323.00	12213.0
324.00	2604.0
326.90	2698.0
328.00	1222.0
331.90	675.0
333.00	1616.0
334.00	7659.0
335.00	2140.0
338.80	359.0
341.00	1766.0
341.90	506.0
346.00	2795.0
346.80	325.0
352.00	4680.0
352.90	2545.0
354.10	5407.0
355.00	1341.0
359.10	338.0
365.00	24720.0
366.00	3360.0
369.90	702.0
371.00	1372.0
371.90	7433.0
372.90	1975.0
373.80	373.0
377.00	310.0
383.00	2155.0
383.80	648.0
389.90	933.0
390.90	611.0
401.00	566.0
401.90	3806.0
402.90	4603.0
403.90	1473.0
404.90	324.0
414.90	341.0
421.00	4642.0
422.00	3120.0
423.00	25256.0
423.90	5500.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

425.00	681.0
439.20	373.0
441.00	58088.0
442.00	345280.0
443.00	74888.0
444.00	7010.0
444.90	504.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66