

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081722\
 Data File : BN021276.D
 Acq On : 17 Aug 2022 09:37
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 18 06:40:15 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 16 09:19:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	125	0.00
2	1,4-Dioxane	0.400	0.377	5.8	132	0.00
3	n-Nitrosodimethylamine	0.400	0.398	0.5	133	0.00
4 S	2-Fluorophenol	0.400	0.389	2.8	136	0.00
5 S	Phenol-d6	0.400	0.402	-0.5	137	0.00
6	bis(2-Chloroethyl)ether	0.400	0.443	-10.7	141	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	112	-0.01
8 S	Nitrobenzene-d5	0.400	0.420	-5.0	140	-0.01
9	Naphthalene	0.400	0.397	0.8	117	0.00
10	Hexachlorobutadiene	0.400	0.378	5.5	116	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.360	10.0	113	0.00
12	2-Methylnaphthalene	0.400	0.348	13.0	110	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	115	0.00
14 S	2,4,6-Tribromophenol	0.400	0.286	28.5#	105	0.00
15 S	2-Fluorobiphenyl	0.400	0.392	2.0	119	0.00
16	Acenaphthylene	0.400	0.328	18.0	110	0.00
17	Acenaphthene	0.400	0.352	12.0	110	0.00
18	Fluorene	0.400	0.354	11.5	113	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	127	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.371	7.3	80	0.00
21	4-Bromophenyl-phenylether	0.400	0.377	5.8	143	0.00
22	Hexachlorobenzene	0.400	0.384	4.0	143	0.00
23	Atrazine	0.400	0.331	17.3	135	0.00
24	Pentachlorophenol	0.400	0.343	14.2	140	-0.01
25	Phenanthrene	0.400	0.379	5.3	144	0.00
26	Anthracene	0.400	0.353	11.8	138	-0.01
27 SURR	Fluoranthene-d10	0.400	0.371	7.3	138	0.00
28	Fluoranthene	0.400	0.372	7.0	142	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	125	0.00
30	Pyrene	0.400	0.420	-5.0	141	0.00
31 S	Terphenyl-d14	0.400	0.404	-1.0	136	0.00
32	Benzo(a)anthracene	0.400	0.366	8.5	131	0.00
33	Chrysene	0.400	0.405	-1.3	140	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.356	11.0	124	0.00
35 I	Perylene-d12	0.400	0.400	0.0	107	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.359	10.3	109	0.00
37	Benzo(b)fluoranthene	0.400	0.381	4.8	110	0.00
38	Benzo(k)fluoranthene	0.400	0.386	3.5	115	0.00
39 C	Benzo(a)pyrene	0.400	0.376	6.0	107	0.00
40	Dibenzo(a,h)anthracene	0.400	0.361	9.8	109	0.00
41	Benzo(g,h,i)perylene	0.400	0.387	3.3	115	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4