

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081722\
 Data File : BN021282.D
 Acq On : 17 Aug 2022 17:08
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 18 06:42:53 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 16 09:19:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	107	0.00
2	1,4-Dioxane	0.400	0.378	5.5	113	0.00
3	n-Nitrosodimethylamine	0.400	0.413	-3.2	122	0.00
4 S	2-Fluorophenol	0.400	0.341	14.8	102	0.00
5 S	Phenol-d6	0.400	0.421	-5.2	123	0.00
6	bis(2-Chloroethyl)ether	0.400	0.411	-2.7	111	0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	106	0.00
8 S	Nitrobenzene-d5	0.400	0.373	6.8	112	0.00
9	Naphthalene	0.400	0.386	3.5	108	0.00
10	Hexachlorobutadiene	0.400	0.354	11.5	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.309	22.8	92	0.00
12	2-Methylnaphthalene	0.400	0.308	23.0	92	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	91	0.00
14 S	2,4,6-Tribromophenol	0.400	0.345	13.8	100	0.00
15 S	2-Fluorobiphenyl	0.400	0.386	3.5	93	0.00
16	Acenaphthylene	0.400	0.364	9.0	96	0.00
17	Acenaphthene	0.400	0.380	5.0	93	0.00
18	Fluorene	0.400	0.370	7.5	93	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	100	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.427	-6.7	94	0.01
21	4-Bromophenyl-phenylether	0.400	0.357	10.8	107	0.00
22	Hexachlorobenzene	0.400	0.381	4.8	112	0.00
23	Atrazine	0.400	0.388	3.0	125	0.00
24	Pentachlorophenol	0.400	0.343	14.2	111	0.00
25	Phenanthrene	0.400	0.371	7.3	111	0.00
26	Anthracene	0.400	0.360	10.0	111	0.00
27 SURR	Fluoranthene-d10	0.400	0.382	4.5	112	0.00
28	Fluoranthene	0.400	0.381	4.8	114	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	112	0.00
30	Pyrene	0.400	0.383	4.3	115	0.00
31 S	Terphenyl-d14	0.400	0.394	1.5	119	0.00
32	Benzo(a)anthracene	0.400	0.359	10.3	115	0.00
33	Chrysene	0.400	0.371	7.3	115	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.418	-4.5	130	0.00
35 I	Perylene-d12	0.400	0.400	0.0	103	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.382	4.5	111	0.00
37	Benzo(b)fluoranthene	0.400	0.359	10.3	99	0.00
38	Benzo(k)fluoranthene	0.400	0.343	14.2	98	0.00
39 C	Benzo(a)pyrene	0.400	0.343	14.2	94	0.00
40	Dibenzo(a,h)anthracene	0.400	0.373	6.8	108	0.00
41	Benzo(g,h,i)perylene	0.400	0.386	3.5	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4