

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081822\
 Data File : BN021288.D
 Acq On : 18 Aug 2022 20:01
 Operator : CG/JU
 Sample : SSTD08027
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD080228

Quant Time: Aug 18 21:13:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 18 19:40:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.099	152	112110	20.000	ng/u1	0.00
20) Naphthalene-d8	10.934	136	459979	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.757	164	259270	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.510	188	520824	20.000	ng/u1	0.00
79) Chrysene-d12	21.698	240	489165	20.000	ng/u1	0.00
88) Perylene-d12	24.239	264	509412	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	103795	40.100	ng/uL	0.00
4) Pyridine-d5	3.817	84	691019	97.674	ng/u1	0.00
7) Phenol-d5	7.246	99	808027	87.102	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.422	67	474713	81.904	ng/u1	0.00
11) 2-Chlorophenol-d4	7.622	132	636341	84.422	ng/u1	0.00
15) 4-Methylphenol-d8	8.816	113	601195	74.906	ng/u1	0.00
21) Nitrobenzene-d5	9.281	128	285913	81.703	ng/u1	0.00
24) 2-Nitrophenol-d4	10.005	143	267612	72.058	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.552	165	555034	85.609	ng/u1	0.00
31) 4-Chloroaniline-d4	11.075	131	803043	76.487	ng/u1	0.00
46) Dimethylphthalate-d6	14.169	166	1443485	76.517	ng/u1	0.00
49) Acenaphthylene-d8	14.451	160	1806904	82.592	ng/u1	0.00
54) 4-Nitrophenol-d4	14.951	143	277702	73.461	ng/u1	0.00
60) Fluorene-d10	15.751	176	1261925	80.837	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.863	200	179012	59.336	ng/u1	0.00
73) Anthracene-d10	17.610	188	1886198	82.982	ng/u1	0.00
81) Pyrene-d10	19.898	212	2055299	77.504	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.080	264	2172074	87.147	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.423	88	110912	40.590	ng/uL	95
5) Pyridine	3.840	79	678700	91.509	ng/u1	86
6) Benzaldehyde	7.222	77	330453	75.600	ng/u1	92
8) Phenol	7.275	94	821910	83.204	ng/u1	92
10) Bis(2-Chloroethyl)ether	7.516	93	661581	85.569	ng/u1	95
12) 2-Chlorophenol	7.658	128	662883	84.135	ng/u1	96
13) 2-Methylphenol	8.546	108	611099	80.470	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.640	45	808075	66.652	ng/u1	96
16) Acetophenone	8.940	105	909120	74.484	ng/u1	94
17) N-Nitroso-di-n-propyla...	8.928	70	457582	73.388	ng/u1	92
18) 4-Methylphenol	8.881	108	642202	76.620	ng/u1	97
19) Hexachloroethane	9.199	117	272996	86.112	ng/u1	93
22) Nitrobenzene	9.322	77	675388	83.239	ng/u1	94
23) Isophorone	9.858	82	1322143	82.529	ng/u1	98
25) 2-Nitrophenol	10.040	139	312783	76.218	ng/u1	92
26) 2,4-Dimethylphenol	10.099	107	675853	79.494	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.340	93	819478	83.588	ng/u1	98
29) 2,4-Dichlorophenol	10.575	162	567912	84.681	ng/u1	96
30) Naphthalene	10.987	128	1929883	80.954	ng/u1	99
32) 4-Chloroaniline	11.099	127	827827	78.829	ng/u1	99
33) Hexachlorobutadiene	11.269	225	326582	84.817	ng/u1	98
34) Caprolactam	11.869	113	117862	45.004	ng/u1	82
35) 4-Chloro-3-methylphenol	12.216	107	594816	75.114	ng/u1	98
36) 2-Methylnaphthalene	12.593	142	1272581	78.960	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081822\
 Data File : BN021288.D
 Acq On : 18 Aug 2022 20:01
 Operator : CG/JU
 Sample : SSTD08027
 Misc :
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD080228


Quant Time: Aug 18 21:13:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 18 19:40:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.810	142	1257876	76.482	ng/ul	92
39) 1,2,4,5-Tetrachloroben...	12.951	216	608731	92.057	ng/ul	99
40) Hexachlorocyclopentadiene	12.934	237	386575	96.104	ng/ul	99
41) 2,4,6-Trichlorophenol	13.193	196	386440	84.755	ng/ul	98
42) 2,4,5-Trichlorophenol	13.263	196	432719	85.691	ng/ul	96
43) 1,1'-Biphenyl	13.599	154	1617654	84.260	ng/ul	97
44) 2-Chloronaphthalene	13.640	162	1243926	85.409	ng/ul	98
45) 2-Nitroaniline	13.840	65	334382	73.798	ng/ul	91
47) Dimethylphthalate	14.216	163	1482262	78.050	ng/ul	98
48) 2,6-Dinitrotoluene	14.334	165	291608	78.535	ng/ul	95
50) Acenaphthylene	14.481	152	1984602	81.934	ng/ul	99
51) 3-Nitroaniline	14.663	138	290996	74.131	ng/ul#	98
52) Acenaphthene	14.822	153	1272977	80.458	ng/ul	99
53) 2,4-Dinitrophenol	14.863	184	120280	50.595	ng/ul	98
55) 4-Nitrophenol	14.969	109	221538	68.656	ng/ul	91
56) Dibenzofuran	15.157	168	1805945	81.020	ng/ul	99
57) 2,4-Dinitrotoluene	15.116	165	405800	73.811	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.381	232	325919	82.220	ng/ul	96
59) Diethylphthalate	15.581	149	1482152	75.804	ng/ul	97
61) Fluorene	15.810	166	1374098	79.223	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.798	204	634952	79.492	ng/ul	99
63) 4-Nitroaniline	15.828	138	257243	70.130	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.875	198	189919	62.091	ng/ul	95
67) N-Nitrosodiphenylamine	16.010	169	1230639	82.421	ng/ul	95
68) 4-Bromophenyl-phenylether	16.698	248	403689	89.004	ng/ul	99
69) Hexachlorobenzene	16.804	284	458415	87.926	ng/ul	95
70) Atrazine	16.969	200	429346	82.413	ng/ul	97
71) Pentachlorophenol	17.151	266	272799	84.009	ng/ul	98
72) Phenanthrene	17.551	178	2175653	79.156	ng/ul	99
74) Anthracene	17.645	178	2199231	79.574	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.557	216	626912	89.889	ng/ul	98
76) Pentachlorobenzene	15.069	250	557343	90.419	ng/ul	100
77) Carbazole	17.916	167	2060687	81.958	ng/ul	96
78) Di-n-butylphthalate	18.481	149	2601393	83.438	ng/ul	99
80) Fluoranthene	19.563	202	2451329	76.537	ng/ul#	93
82) Pyrene	19.928	202	2497725	76.014	ng/ul#	92
83) Butylbenzylphthalate	20.828	149	1149707	77.747	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.616	252	790308	84.589	ng/ul	99
85) Benzo(a)anthracene	21.680	228	2415164	78.188	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.616	149	1600315	76.700	ng/ul	100
87) Chrysene	21.739	228	2282430	75.528	ng/ul	95
89) Di-n-octyl phthalate	22.592	149	2974120	71.732	ng/ul	100
90) Benzo(b)fluoranthene	23.463	252	2610768	81.659	ng/ul	97
91) Benzo(k)fluoranthene	23.522	252	2413858	79.312	ng/ul	98
93) Benzo(a)pyrene	24.139	252	2507477	81.023	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.939	276	3085090	96.891	ng/ul	97
95) Dibenzo(a,h)anthracene	26.974	278	2643227	97.190	ng/ul	98
96) Benzo(g,h,i)perylene	27.774	276	2716402	101.023	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 18 21:13:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081822.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Aug 18 19:40:32 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 40.590 ng/uL
 RT: 3.423 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BN021288.D
 Acq: 18 Aug 2022 20:01




Instrument :
 BNA_N
 ClientSampleId :
 SSTD080228


Tgt Ion: 88 Resp: 110912
 Ion Ratio Lower Upper
 88 100
 43 27.5 22.4 33.6
 58 70.9 61.0 91.4



#5
 Pyridine
 Concen: 91.509 ng/uL
 RT: 3.840 min Scan# 94
 Delta R.T. -0.000 min
 Lab File: BN021288.D
 Acq: 18 Aug 2022 20:01

Tgt Ion: 79 Resp: 678700
 Ion Ratio Lower Upper
 79 100
 52 67.7 65.0 97.4
 51 32.2 31.6 47.4









#6
Benzaldehyde
Concen: 75.600 ng/ul
RT: 7.222 min Scan# 669
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion: 77 Resp: 330453
Ion Ratio Lower Upper
77 100
105 99.6 74.7 112.1
106 96.3 70.2 105.4



#8
Phenol
Concen: 83.204 ng/ul
RT: 7.275 min Scan# 678
Delta R.T. 0.006 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion: 94 Resp: 821910
Ion Ratio Lower Upper
94 100
65 26.1 23.6 35.4
66 33.7 31.2 46.8









#10
Bis(2-Chloroethyl)ether
Concen: 85.569 ng/ul
RT: 7.516 min Scan# 719
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion: 93 Resp: 661581
Ion Ratio Lower Upper
93 100
63 72.8 63.4 95.0
95 32.7 26.2 39.2



#12
2-Chlorophenol
Concen: 84.135 ng/ul
RT: 7.658 min Scan# 743
Delta R.T. 0.006 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:128 Resp: 662883
Ion Ratio Lower Upper
128 100
64 44.8 38.6 57.8
130 32.8 25.0 37.4









#13
2-Methylphenol
Concen: 80.470 ng/ul
RT: 8.546 min Scan# 894
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:108 Resp: 611099
Ion Ratio Lower Upper
108 100
107 91.0 71.5 107.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 66.652 ng/ul
RT: 8.640 min Scan# 910
Delta R.T. 0.006 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion: 45 Resp: 808075
Ion Ratio Lower Upper
45 100
77 15.5 10.9 16.3
79 11.8 8.6 13.0









#16
Acetophenone
Concen: 74.484 ng/ul
RT: 8.940 min Scan# 90
Delta R.T. 0.006 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:105 Resp: 909120
Ion Ratio Lower Upper
105 100
77 75.4 63.5 95.3
51 24.1 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 73.388 ng/ul
RT: 8.928 min Scan# 959
Delta R.T. 0.012 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion: 70 Resp: 457582
Ion Ratio Lower Upper
70 100
42 52.2 47.8 71.6
101 9.8 7.8 11.6
130 21.8 15.8 23.8





#18

4-Methylphenol

Concen: 76.620 ng/ul

RT: 8.881 min Scan# 91

Delta R.T. 0.006 min

Lab File: BN021288.D

Acq: 18 Aug 2022 20:01

Instrument :

BNA_N

ClientSampleId :





SSTD080228

Tgt Ion:108 Resp: 642202

Ion Ratio Lower Upper

108 100

107 113.5 88.0 132.0



#19

Hexachloroethane

Concen: 86.112 ng/ul

RT: 9.199 min Scan# 1005

Delta R.T. -0.000 min

Lab File: BN021288.D

Acq: 18 Aug 2022 20:01




Tgt Ion:117 Resp: 272996


Ion Ratio Lower Upper

117 100

201 95.7 83.2 124.8

199 59.3 50.2 75.4









#22
Nitrobenzene
Concen: 83.239 ng/ul
RT: 9.322 min Scan# 1026
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion: 77 Resp: 675388
Ion Ratio Lower Upper
77 100
123 47.4 34.5 51.7
65 13.0 11.1 16.7



#23
Isophorone
Concen: 82.529 ng/ul
RT: 9.858 min Scan# 1117
Delta R.T. 0.012 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion: 82 Resp: 1322143
Ion Ratio Lower Upper
82 100
95 6.9 5.8 8.8
138 18.1 13.8 20.8









#25
2-Nitrophenol
Concen: 76.218 ng/ul
RT: 10.040 min Scan# 1148
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:139 Resp: 312783
Ion Ratio Lower Upper
139 100
65 48.5 43.7 65.5
109 28.6 25.8 38.6



#26
2,4-Dimethylphenol
Concen: 79.494 ng/ul
RT: 10.099 min Scan# 1158
Delta R.T. 0.006 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:107 Resp: 675853
Ion Ratio Lower Upper
107 100
121 51.8 38.3 57.5
122 86.2 66.3 99.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 83.588 ng/ul
 RT: 10.340 min Scan# 1199
 Delta R.T. -0.000 min
 Lab File: BN021288.D
 Acq: 18 Aug 2022 20:01




Instrument :
 BNA_N
 ClientSampleId :
 SSTD080228


Tgt Ion: 93 Resp: 819478
 Ion Ratio Lower Upper
 93 100
 95 30.8 25.4 38.0
 123 11.5 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 84.681 ng/ul
 RT: 10.575 min Scan# 1239
 Delta R.T. -0.000 min
 Lab File: BN021288.D
 Acq: 18 Aug 2022 20:01

Tgt Ion:162 Resp: 567912
 Ion Ratio Lower Upper
 162 100
 164 62.4 52.9 79.3
 98 35.9 29.8 44.8









#30
Naphthalene
Concen: 80.954 ng/ul
RT: 10.987 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:128 Resp: 1929883
Ion Ratio Lower Upper
128 100
129 11.3 8.7 13.1
127 13.1 10.6 16.0



#32
4-Chloroaniline
Concen: 78.829 ng/ul
RT: 11.099 min Scan# 1328
Delta R.T. 0.006 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:127 Resp: 827827
Ion Ratio Lower Upper
127 100
129 31.8 26.1 39.1





#33

Hexachlorobutadiene

Concen: 84.817 ng/ul

RT: 11.269 min Scan# 11

Delta R.T. -0.006 min

Lab File: BN021288.D

Acq: 18 Aug 2022 20:01

Instrument :

BNA_N

ClientSampleId :

SSTD080228





Tgt Ion:225 Resp: 326582

Ion Ratio Lower Upper

225 100

223 63.3 51.8 77.6

227 64.7 50.6 76.0



#34

Caprolactam

Concen: 45.004 ng/ul

RT: 11.869 min Scan# 1459

Delta R.T. 0.023 min

Lab File: BN021288.D

Acq: 18 Aug 2022 20:01




Tgt Ion:113 Resp: 117862


Ion Ratio Lower Upper

113 100

55 156.6 145.5 218.3

56 111.4 106.5 159.7









#35
4-Chloro-3-methylphenol
Concen: 75.114 ng/ul
RT: 12.216 min Scan# 1518
Delta R.T. 0.006 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:107 Resp: 594816
Ion Ratio Lower Upper
107 100
144 25.8 20.6 31.0
142 79.7 62.2 93.4



#36
2-Methylnaphthalene
Concen: 78.960 ng/ul
RT: 12.593 min Scan# 1582
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:142 Resp: 1272581
Ion Ratio Lower Upper
142 100
141 90.6 70.6 105.8









#37
1-Methylnaphthalene
Concen: 76.482 ng/ul
RT: 12.810 min Scan# 1619
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:142 Resp: 1257876
Ion Ratio Lower Upper
142 100
141 96.0 70.9 106.3
116 3.9 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 92.057 ng/ul
RT: 12.951 min Scan# 1643
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:216 Resp: 608731
Ion Ratio Lower Upper
216 100
214 78.6 63.4 95.0
179 21.1 17.8 26.8
108 19.4 15.1 22.7









#40
Hexachlorocyclopentadiene
Concen: 96.104 ng/ul
RT: 12.934 min Scan# 1640
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:237 Resp: 386575
Ion Ratio Lower Upper
237 100
235 60.4 48.8 73.2
272 12.9 9.3 13.9



#41
2,4,6-Trichlorophenol
Concen: 84.755 ng/ul
RT: 13.193 min Scan# 1684
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:196 Resp: 386440
Ion Ratio Lower Upper
196 100
198 98.5 79.8 119.8
200 30.0 25.0 37.6









#42
 2,4,5-Trichlorophenol
 Concen: 85.691 ng/ul
 RT: 13.263 min Scan# 1696
 Delta R.T. -0.000 min
 Lab File: BN021288.D
 Acq: 18 Aug 2022 20:01

Instrument :
 BNA_N
 ClientSampleId :
 SST080228

Tgt Ion:196 Resp: 432719




Ion	Ratio	Lower	Upper
196	100		
198	93.3	77.6	116.4
200	27.2	24.4	36.6




#43
 1,1'-Biphenyl
 Concen: 84.260 ng/ul
 RT: 13.599 min Scan# 1753
 Delta R.T. -0.000 min
 Lab File: BN021288.D
 Acq: 18 Aug 2022 20:01

Tgt Ion:154 Resp: 1617654

Ion	Ratio	Lower	Upper
154	100		
153	41.2	31.5	47.3
76	13.7	10.5	15.7









#44
2-Chloronaphthalene
Concen: 85.409 ng/ul
RT: 13.640 min Scan# 1759
Delta R.T. 0.006 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:162 Resp: 1243926
Ion Ratio Lower Upper
162 100
164 33.1 25.8 38.6
127 40.7 31.4 47.2



#45
2-Nitroaniline
Concen: 73.798 ng/ul
RT: 13.840 min Scan# 1794
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion: 65 Resp: 334382
Ion Ratio Lower Upper
65 100
92 71.1 54.1 81.1
138 108.3 77.0 115.4









#47
Dimethylphthalate
Concen: 78.050 ng/ul
RT: 14.216 min Scan# 1858
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:163 Resp: 1482262
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.0
164 10.7 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 78.535 ng/ul
RT: 14.334 min Scan# 1878
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:165 Resp: 291608
Ion Ratio Lower Upper
165 100
89 57.7 50.6 75.8
121 22.4 17.9 26.9









#50
 Acenaphthylene
 Concen: 81.934 ng/ul
 RT: 14.481 min Scan# 1903
 Delta R.T. -0.000 min
 Lab File: BN021288.D
 Acq: 18 Aug 2022 20:01

Instrument :
 BNA_N
 ClientSampleId :
 SSTD080228

Tgt Ion:152 Resp: 1984602




Ion	Ratio	Lower	Upper
152	100		
151	20.3	16.0	24.0
153	13.7	10.7	16.1




#51
 3-Nitroaniline
 Concen: 74.131 ng/ul
 RT: 14.663 min Scan# 1934
 Delta R.T. 0.006 min
 Lab File: BN021288.D
 Acq: 18 Aug 2022 20:01

Tgt Ion:138 Resp: 290996

Ion	Ratio	Lower	Upper
138	100		
108	13.3	8.6	12.8
92	118.1	92.9	139.3









#52
Acenaphthene
Concen: 80.458 ng/ul
RT: 14.822 min Scan# 1961
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:153 Resp: 1272977
Ion Ratio Lower Upper
153 100
152 46.6 38.4 57.6
154 87.2 69.7 104.5



#53
2,4-Dinitrophenol
Concen: 50.595 ng/ul
RT: 14.863 min Scan# 1968
Delta R.T. 0.006 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:184 Resp: 120280
Ion Ratio Lower Upper
184 100
63 66.8 55.9 83.9
154 65.6 53.1 79.7









#55
4-Nitrophenol
Concen: 68.656 ng/ul
RT: 14.969 min Scan# 1986
Delta R.T. 0.012 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:109 Resp: 221538
Ion Ratio Lower Upper
109 100
139 142.6 100.1 150.1
65 127.8 104.6 157.0



#56
Dibenzofuran
Concen: 81.020 ng/ul
RT: 15.157 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:168 Resp: 1805945
Ion Ratio Lower Upper
168 100
139 39.4 31.1 46.7









#57
2,4-Dinitrotoluene
Concen: 73.811 ng/ul
RT: 15.116 min Scan# 2011
Delta R.T. 0.006 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:165 Resp: 405800
Ion Ratio Lower Upper
165 100
63 43.1 36.3 54.5
182 2.8 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 82.220 ng/ul
RT: 15.381 min Scan# 2056
Delta R.T. 0.006 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:232 Resp: 325919
Ion Ratio Lower Upper
232 100
131 46.2 38.5 57.7
130 2.3 2.2 3.2
166 29.4 25.8 38.8





#59

Diethylphthalate

Concen: 75.804 ng/ul

RT: 15.581 min Scan# 2089

Delta R.T. 0.006 min

Lab File: BN021288.D

Acq: 18 Aug 2022 20:01

Instrument :

BNA_N

ClientSampleId :

SSTD080228





Tgt Ion:149 Resp: 1482152

Ion Ratio Lower Upper

149 100

177 23.8 17.8 26.8

150 12.2 10.5 15.7



#61

Fluorene

Concen: 79.223 ng/ul

RT: 15.810 min Scan# 2129

Delta R.T. 0.006 min

Lab File: BN021288.D

Acq: 18 Aug 2022 20:01




Tgt Ion:166 Resp: 1374098


Ion Ratio Lower Upper

166 100

165 95.3 77.7 116.5

167 12.7 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 79.492 ng/ul
RT: 15.798 min Scan# 2127
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:204 Resp: 634952
Ion Ratio Lower Upper
204 100
206 33.1 26.7 40.1
141 64.1 50.2 75.2



#63
4-Nitroaniline
Concen: 70.130 ng/ul
RT: 15.828 min Scan# 2132
Delta R.T. 0.012 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:138 Resp: 257243
Ion Ratio Lower Upper
138 100
92 54.1 41.0 61.6
108 78.8 65.3 97.9









#66
 4,6-Dinitro-2-methylphenol
 Concen: 62.091 ng/ul
 RT: 15.875 min Scan# 2140
 Delta R.T. 0.006 min
 Lab File: BN021288.D
 Acq: 18 Aug 2022 20:01




Instrument :
 BNA_N
 ClientSampleId :
 SSTD080228


Tgt Ion:198 Resp: 189919
 Ion Ratio Lower Upper
 198 100
 182 4.3 3.0 4.4
 77 28.1 20.1 30.1



#67
 N-Nitrosodiphenylamine
 Concen: 82.421 ng/ul
 RT: 16.010 min Scan# 2163
 Delta R.T. -0.000 min
 Lab File: BN021288.D
 Acq: 18 Aug 2022 20:01

Tgt Ion:169 Resp: 1230639
 Ion Ratio Lower Upper
 169 100
 168 68.9 52.4 78.6
 167 37.8 27.5 41.3









#68
4-Bromophenyl-phenylether
Concen: 89.004 ng/ul
RT: 16.698 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:248 Resp: 403689
Ion Ratio Lower Upper
248 100
250 98.3 77.8 116.8
141 77.1 62.6 93.8



#69
Hexachlorobenzene
Concen: 87.926 ng/ul
RT: 16.804 min Scan# 2298
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:284 Resp: 458415
Ion Ratio Lower Upper
284 100
142 37.7 27.8 41.8
249 29.1 25.7 38.5









#70
 Atrazine
 Concen: 82.413 ng/ul
 RT: 16.969 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: BN021288.D
 Acq: 18 Aug 2022 20:01




Instrument :
 BNA_N
 ClientSampleId :
 SST080228


Tgt Ion:200 Resp: 429346
 Ion Ratio Lower Upper
 200 100
 173 26.8 20.5 30.7
 215 46.6 38.9 58.3



#71
 Pentachlorophenol
 Concen: 84.009 ng/ul
 RT: 17.151 min Scan# 2357
 Delta R.T. -0.000 min
 Lab File: BN021288.D
 Acq: 18 Aug 2022 20:01

Tgt Ion:266 Resp: 272799
 Ion Ratio Lower Upper
 266 100
 264 64.3 50.2 75.2
 268 60.3 49.3 73.9









#72
Phenanthrene
Concen: 79.156 ng/uI
RT: 17.551 min Scan# 2425
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:178 Resp: 2175653
Ion Ratio Lower Upper
178 100
179 15.7 12.2 18.4
176 18.9 15.2 22.8



#74
Anthracene
Concen: 79.574 ng/uI
RT: 17.645 min Scan# 2441
Delta R.T. 0.006 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:178 Resp: 2199231
Ion Ratio Lower Upper
178 100
179 16.1 13.0 19.4
176 18.9 15.5 23.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 89.889 ng/uL

RT: 13.557 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN021288.D

Acq: 18 Aug 2022 20:01

Instrument :

BNA_N

ClientSampleId :

SSTD080228





Tgt Ion:216 Resp: 626912

Ion Ratio Lower Upper

216 100

214 75.9 63.2 94.8

218 47.8 37.9 56.9



#76

Pentachlorobenzene

Concen: 90.419 ng/uL

RT: 15.069 min Scan# 2003

Delta R.T. -0.000 min

Lab File: BN021288.D

Acq: 18 Aug 2022 20:01




Tgt Ion:250 Resp: 557343


Ion Ratio Lower Upper

250 100

248 66.0 53.0 79.4

252 66.4 53.4 80.2









#77
 Carbazole
 Concen: 81.958 ng/ul
 RT: 17.916 min Scan# 2487
 Delta R.T. 0.006 min
 Lab File: BN021288.D
 Acq: 18 Aug 2022 20:01

Instrument :
 BNA_N
 ClientSampleId :
 SSTD080228

Tgt Ion:167 Resp: 2060687




Ion	Ratio	Lower	Upper
167	100		
166	22.8	16.1	24.1
139	13.5	10.2	15.2




#78
 Di-n-butylphthalate
 Concen: 83.438 ng/ul
 RT: 18.481 min Scan# 2583
 Delta R.T. -0.000 min
 Lab File: BN021288.D
 Acq: 18 Aug 2022 20:01

Tgt Ion:149 Resp: 2601393

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.1	10.7
104	5.3	4.1	6.1









#80
Fluoranthene
Concen: 76.537 ng/u1
RT: 19.563 min Scan# 2767
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:202 Resp: 2451329
Ion Ratio Lower Upper
202 100
101 13.9 8.9 13.3#
100 10.4 6.7 10.1#



#82
Pyrene
Concen: 76.014 ng/u1
RT: 19.928 min Scan# 2829
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:202 Resp: 2497725
Ion Ratio Lower Upper
202 100
101 15.3 9.7 14.5#
100 12.6 8.0 12.0#









#83
Butylbenzylphthalate
Concen: 77.747 ng/ul
RT: 20.828 min Scan# 2982
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:149 Resp: 1149707
Ion Ratio Lower Upper
149 100
91 74.5 58.8 88.2
206 20.0 18.1 27.1



#84
3,3'-Dichlorobenzidine
Concen: 84.589 ng/ul
RT: 21.616 min Scan# 3116
Delta R.T. 0.006 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:252 Resp: 790308
Ion Ratio Lower Upper
252 100
254 63.8 50.8 76.2
126 11.9 8.2 12.4





#85

Benzo(a)anthracene

Concen: 78.188 ng/ul

RT: 21.680 min Scan# 3127

Delta R.T. -0.000 min

Lab File: BN021288.D

Acq: 18 Aug 2022 20:01

Instrument :

BNA_N

ClientSampleId :

SSTD080228





Tgt Ion:228 Resp: 2415164

Ion Ratio Lower Upper

228 100

229 19.8 16.3 24.5

226 26.7 23.2 34.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 76.700 ng/ul

RT: 21.616 min Scan# 3116

Delta R.T. -0.000 min

Lab File: BN021288.D

Acq: 18 Aug 2022 20:01




Tgt Ion:149 Resp: 1600315


Ion Ratio Lower Upper

149 100

167 28.0 22.5 33.7

279 4.4 3.5 5.3









#87
Chrysene
Concen: 75.528 ng/ul
RT: 21.739 min Scan# 3137
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:228 Resp: 2282430
Ion Ratio Lower Upper
228 100
226 33.8 24.2 36.2
229 21.7 16.2 24.2



#89
Di-n-octyl phthalate
Concen: 71.732 ng/ul
RT: 22.592 min Scan# 3282
Delta R.T. -0.000 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:149 Resp: 2974120









#90
Benzo(b)fluoranthene
Concen: 81.659 ng/ul
RT: 23.463 min Scan# 3429
Delta R.T. 0.006 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:252 Resp: 2610768
Ion Ratio Lower Upper
252 100
253 21.5 16.5 24.7
125 11.5 7.8 11.8



#91
Benzo(k)fluoranthene
Concen: 79.312 ng/ul
RT: 23.522 min Scan# 3440
Delta R.T. 0.012 min
Lab File: BN021288.D
Acq: 18 Aug 2022 20:01

Tgt Ion:252 Resp: 2413858
Ion Ratio Lower Upper
252 100
253 21.7 18.2 27.4
125 11.4 8.2 12.4





#93

Benzo(a)pyrene

Concen: 81.023 ng/u1

RT: 24.139 min Scan# 3544

Delta R.T. 0.006 min

Lab File: BN021288.D

Acq: 18 Aug 2022 20:01

Instrument :

BNA_N

ClientSampleId :

SSTD080228





Tgt Ion:252 Resp: 2507477

Ion Ratio Lower Upper

252 100

253 22.6 18.7 28.1

125 12.8 9.2 13.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 96.891 ng/u1

RT: 26.939 min Scan# 4021

Delta R.T. 0.012 min

Lab File: BN021288.D

Acq: 18 Aug 2022 20:01




Tgt Ion:276 Resp: 3085090


Ion Ratio Lower Upper

276 100

138 25.7 19.5 29.3

277 24.8 21.4 32.0









#95
 Dibenzo(a,h)anthracene
 Concen: 97.190 ng/ul
 RT: 26.974 min Scan# 4027
 Delta R.T. 0.012 min
 Lab File: BN021288.D
 Acq: 18 Aug 2022 20:01

Instrument :
 BNA_N
 ClientSampleId :
 SSTD080228

Tgt Ion:278 Resp: 2643227
 Ion Ratio Lower Upper
 278 100
 139 17.7 13.8 20.8
 279 23.8 18.3 27.5



#96
 Benzo(g,h,i)perylene
 Concen: 101.023 ng/ul
 RT: 27.774 min Scan# 4163
 Delta R.T. 0.018 min
 Lab File: BN021288.D
 Acq: 18 Aug 2022 20:01

Tgt Ion:276 Resp: 2716402
 Ion Ratio Lower Upper
 276 100
 138 24.6 17.0 25.6
 277 21.4 17.8 26.6

