


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081822\
 Data File : BN021283.D
 Acq On : 18 Aug 2022 16:57
 Operator : CG/JU
 Sample : DFTPP247
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP247

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081622.M
 Title : SVOA CALIBRATION
 Last Update : Wed Aug 17 01:58:37 2022



AutoFind: Scans 2492, 2493, 2494; Background Corrected with Scan 2486

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.1	67955	PASS
68	69	0.00	2	1.5	1086	PASS
69	198	0.00	100	39.8	70958	PASS
70	69	0.00	2	0.3	221	PASS
127	198	10	80	51.9	92411	PASS
197	198	0.00	2	0.3	591	PASS
198	198	100	100	100.0	178176	PASS
199	198	5	9	6.9	12316	PASS
275	198	10	60	25.0	44549	PASS
365	198	1	100	3.1	5545	PASS
441	198	0.01	100	12.9	23011	PASS
442	198	50	100	76.0	135389	PASS
443	442	15	24	20.5	27789	PASS

m/z	Abundance
35.90	308.0
37.90	249.0
39.00	729.0
40.00	976.0
41.00	998.0
42.00	546.0
43.00	1187.0
44.00	8176.0
45.00	393.0
48.90	302.0
51.00	325.0
53.00	250.0
54.00	255.0
55.00	956.0
56.10	421.0
57.00	664.0
58.00	232.0
58.90	180.0
60.00	165.0
64.90	222.0
67.00	501.0
68.00	275.0
69.00	503.0
70.00	416.0
71.10	361.0
73.00	1175.0
74.00	222.0
75.00	189.0
77.00	333.0
77.90	299.0
79.00	416.0
81.10	417.0
82.00	378.0
83.00	423.0
83.90	329.0
85.00	257.0
85.90	243.0
91.10	505.0
93.00	288.0
94.00	174.0
95.00	429.0
96.00	710.0
96.90	328.0
98.00	423.0
103.00	242.0
105.00	315.0
107.00	190.0
115.00	266.0
117.00	185.0
119.10	248.0
128.00	186.0
129.10	176.0
133.00	587.0
134.90	379.0
147.00	572.0
149.00	154.0
162.90	243.0
164.90	162.0
176.90	250.0
191.00	660.0
192.00	171.0
192.90	392.0
207.00	5151.0
208.00	1084.0
208.90	767.0
221.00	285.0
267.00	167.0
277.00	257.0
281.00	815.0
282.10	251.0
283.00	170.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

m/z	Abundance
36.00	460.0
37.00	315.0
38.00	1089.0
39.00	5463.0
40.00	1139.0
41.00	1083.0
42.00	555.0
43.00	1004.0
44.00	6916.0
45.00	445.0
49.00	733.0
50.00	18176.0
51.00	71320.0
52.00	3486.0
53.00	454.0
54.00	194.0
55.00	1401.0
56.00	2286.0
57.00	5236.0
58.00	369.0
59.00	224.0
60.00	214.0
61.00	958.0
62.00	1169.0
63.00	2895.0
63.90	460.0
65.00	1517.0
66.00	226.0
67.00	611.0
68.00	1321.0
69.00	72840.0
70.00	654.0
71.00	332.0
73.00	1698.0
74.00	7963.0
75.00	11791.0
76.00	4311.0
77.00	88936.0
78.00	6200.0
79.00	5573.0
80.00	4373.0
81.00	6528.0
82.00	1717.0
83.00	1952.0
84.00	656.0
85.00	1227.0
86.00	1821.0
87.00	791.0
87.90	431.0
88.90	223.0
91.00	1486.0
92.00	1545.0
93.00	9211.0
94.00	916.0
95.00	458.0
96.00	1199.0
97.00	562.0
98.00	7824.0
99.00	5756.0
99.90	466.0
101.00	3293.0
101.90	263.0
102.90	1116.0
104.00	2101.0
105.00	2118.0
106.00	745.0
107.00	26272.0
108.00	4006.0
109.00	700.0
110.00	46360.0
111.00	6891.0
112.00	835.0
113.00	333.0
115.00	280.0
116.00	1300.0
117.00	23000.0
118.00	1643.0
119.00	330.0
120.00	384.0
121.00	242.0
122.00	1709.0
123.00	2671.0
123.90	1174.0
125.00	1153.0
126.10	501.0
127.00	89440.0
128.00	7266.0
129.00	32760.0
130.00	2864.0
131.00	590.0
131.90	460.0
133.00	658.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

134.00	1051.0
135.00	3290.0
136.00	1188.0
137.00	1752.0
138.10	258.0
138.90	195.0
140.00	485.0
141.00	4607.0
142.00	1698.0
143.00	1004.0
144.00	278.0
145.00	284.0
146.00	805.0
147.00	2858.0
148.00	5678.0
149.00	1301.0
150.10	422.0
151.00	549.0
151.90	488.0
153.00	1241.0
154.00	1048.0
155.00	2408.0
156.00	3678.0
157.00	638.0
158.00	724.0
159.00	660.0
160.00	1500.0
161.10	2022.0
162.00	652.0
163.00	279.0
164.00	226.0
165.00	1683.0
166.00	1537.0
167.00	9656.0
168.00	5540.0
169.00	961.0
170.00	323.0
170.90	325.0
172.00	719.0
173.00	1144.0
174.00	1752.0
175.00	3134.0
175.90	877.0
177.00	1738.0
177.90	624.0
179.00	6006.0
180.00	4696.0
181.00	1907.0
181.90	352.0
183.00	261.0
183.90	348.0
185.00	2837.0
186.00	21864.0
187.00	6541.0
188.00	741.0
189.00	1388.0
190.00	338.0
191.00	1179.0
192.00	2364.0
193.00	2874.0
193.90	721.0
195.00	409.0
196.00	4576.0
197.00	755.0
198.00	171520.0
199.00	11500.0
199.90	965.0
201.40	840.0
203.00	1170.0
204.00	5330.0
205.10	9751.0
206.00	41064.0
207.00	11441.0
208.00	2520.0
209.00	1198.0
210.10	730.0
211.00	1926.0
211.80	362.0
213.00	186.0
214.90	463.0
216.00	1060.0
217.00	10809.0
218.00	1567.0
219.00	223.0
221.00	6641.0
221.80	1721.0
223.00	2278.0
224.00	22608.0
225.00	6256.0
226.00	714.0
227.00	9696.0
228.00	1278.0
228.90	2082.0
229.90	234.0
231.00	746.0
232.00	155.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

234.00	576.0
235.00	761.0
235.90	445.0
236.90	769.0
238.90	389.0
240.00	266.0
240.90	526.0
242.00	1235.0
243.00	1494.0
244.00	18112.0
245.10	2634.0
246.00	3414.0
247.00	852.0
248.00	168.0
249.00	724.0
251.00	198.0
252.90	510.0
254.10	629.0
255.00	76080.0
256.00	13854.0
257.00	1050.0
258.00	5052.0
259.00	673.0
261.00	220.0
265.00	2221.0
265.90	213.0
267.00	202.0
270.90	199.0
272.00	179.0
273.10	2429.0
274.00	7196.0
275.10	39544.0
276.00	5356.0
277.00	3955.0
278.00	666.0
281.00	836.0
282.00	274.0
282.90	483.0
284.10	343.0
284.90	651.0
293.00	810.0
294.00	270.0
294.90	270.0
296.00	11338.0
297.00	1628.0
298.00	170.0
302.00	207.0
303.00	1188.0
304.00	307.0
308.00	185.0
310.00	151.0
314.00	529.0
315.00	1192.0
315.90	507.0
321.00	292.0
322.00	155.0
323.00	2976.0
324.00	579.0
326.90	735.0
328.00	500.0
332.10	204.0
333.00	302.0
334.00	1970.0
335.00	591.0
341.00	538.0
346.00	768.0
352.00	1004.0
353.00	701.0
354.00	1064.0
355.10	380.0
365.00	3913.0
365.90	748.0
371.10	226.0
372.10	1958.0
373.00	420.0
383.00	664.0
390.10	168.0
390.90	168.0
402.10	534.0
403.00	1062.0
404.00	357.0
421.00	613.0
422.00	763.0
423.00	5330.0
424.00	1284.0
441.10	16992.0
442.10	102288.0
443.10	17704.0
444.00	2087.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

m/z	Abundance
36.00	399.0
37.00	332.0
38.00	1049.0
39.10	5982.0
40.00	1100.0
41.00	1143.0
41.90	381.0
43.00	1267.0
44.00	7265.0
45.00	482.0
48.00	159.0
49.00	747.0
50.00	20432.0
51.00	77928.0
52.00	4271.0
53.00	440.0
54.00	277.0
55.00	1359.0
56.00	2554.0
57.00	5623.0
57.90	409.0
59.00	166.0
59.90	192.0
61.00	763.0
62.00	1155.0
63.00	3198.0
64.00	552.0
65.00	1963.0
65.90	276.0
67.00	479.0
68.00	1597.0
69.00	82040.0
70.00	689.0
71.00	434.0
73.00	1907.0
74.00	8507.0
75.00	13626.0
76.00	4629.0
77.10	100024.0
78.00	7179.0
79.00	6234.0
80.00	4810.0
81.00	6703.0
82.00	1915.0
83.00	1992.0
84.00	476.0
85.00	1301.0
86.00	2055.0
87.00	1085.0
88.00	436.0
88.80	227.0
91.00	1742.0
92.00	1741.0
93.00	10887.0
94.00	855.0
95.00	482.0
96.00	1329.0
97.00	555.0
98.00	8831.0
99.00	6772.0
100.00	681.0
101.00	3819.0
101.90	360.0
103.00	1333.0
104.00	2273.0
105.00	2584.0
106.00	800.0
107.00	29216.0
108.00	4614.0
109.00	832.0
110.00	53744.0
111.00	8843.0
112.00	1133.0
113.00	388.0
115.00	421.0
116.00	1486.0
117.00	26224.0
117.90	2069.0
118.90	465.0
119.90	409.0
121.00	288.0
121.90	1913.0
123.00	3111.0
124.00	1467.0
125.00	1279.0
127.00	105912.0
128.00	8068.0
129.00	39808.0
130.00	3234.0
131.00	683.0
132.00	465.0
132.90	807.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

134.00	1336.0
135.00	3108.0
136.00	1369.0
137.00	1566.0
138.00	382.0
138.90	218.0
139.90	507.0
141.00	4815.0
142.00	1833.0
143.00	1505.0
144.00	345.0
144.90	420.0
145.90	1001.0
147.00	2933.0
148.00	6842.0
148.90	1607.0
149.90	407.0
151.10	707.0
151.90	470.0
153.00	1554.0
154.00	1330.0
155.00	3131.0
156.00	4615.0
157.10	872.0
158.00	1076.0
158.90	717.0
160.00	1625.0
161.00	2694.0
161.90	702.0
163.00	344.0
163.80	348.0
165.00	1980.0
166.00	1706.0
167.00	11179.0
168.00	6233.0
169.00	996.0
170.10	341.0
171.00	498.0
171.90	953.0
173.00	1299.0
174.00	2347.0
175.00	3912.0
176.00	1412.0
177.00	1981.0
178.00	731.0
179.00	8157.0
180.00	5815.0
181.00	2402.0
182.00	384.0
182.90	339.0
184.00	678.0
185.00	3751.0
186.00	29752.0
187.00	8290.0
188.00	813.0
189.00	1945.0
190.00	348.0
191.00	1262.0
192.00	2443.0
193.00	3247.0
194.00	551.0
195.00	490.0
196.00	6006.0
197.00	1018.0
198.00	202688.0
199.00	14164.0
200.00	1085.0
201.50	920.0
202.90	1481.0
204.00	7468.0
205.10	12291.0
206.10	53936.0
207.10	12097.0
208.00	2858.0
209.00	1284.0
210.10	1054.0
211.00	2458.0
211.70	413.0
212.90	174.0
215.00	655.0
216.00	991.0
217.00	13671.0
218.00	1996.0
218.90	208.0
221.10	9197.0
221.80	2431.0
223.00	3124.0
224.10	30120.0
225.00	7983.0
226.00	881.0
227.00	12915.0
228.00	1814.0
229.00	2572.0
230.00	334.0
231.00	1097.0
231.90	243.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

232.90	192.0
234.00	897.0
235.00	949.0
236.00	768.0
237.10	1114.0
237.90	185.0
239.00	504.0
240.00	353.0
241.00	594.0
242.00	1558.0
243.10	1699.0
244.00	23192.0
245.00	3505.0
246.00	5017.0
247.00	826.0
247.90	200.0
249.00	932.0
250.10	249.0
250.90	337.0
252.00	229.0
253.00	430.0
255.00	127272.0
256.00	17312.0
257.00	1312.0
258.00	6209.0
259.00	925.0
260.00	185.0
261.00	157.0
264.00	169.0
265.00	2626.0
266.00	372.0
267.00	278.0
270.90	188.0
272.00	305.0
273.10	3982.0
274.00	8813.0
275.00	55536.0
276.00	7855.0
277.00	4994.0
277.90	774.0
279.00	249.0
281.00	741.0
282.00	300.0
283.00	597.0
284.00	373.0
285.00	875.0
286.00	170.0
289.00	171.0
292.00	220.0
293.00	1187.0
294.00	263.0
295.10	359.0
296.00	15363.0
297.00	2096.0
301.00	205.0
302.00	304.0
303.00	1586.0
304.00	410.0
307.90	173.0
310.00	209.0
314.00	792.0
315.00	1528.0
316.00	959.0
317.00	243.0
321.00	591.0
322.00	280.0
323.00	4924.0
324.00	841.0
327.00	1196.0
327.90	430.0
332.00	366.0
333.00	432.0
334.00	3146.0
335.00	795.0
341.10	742.0
341.90	163.0
346.00	1123.0
346.80	161.0
350.90	169.0
352.00	1383.0
353.00	949.0
354.00	1598.0
355.10	406.0
365.00	6970.0
366.00	1104.0
370.90	289.0
372.10	2374.0
373.00	628.0
383.00	680.0
383.90	198.0
390.00	382.0
391.00	253.0
392.00	240.0
401.00	259.0
402.00	1127.0
403.00	1520.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

404.00	518.0
404.90	175.0
421.00	1289.0
422.10	1290.0
423.00	9070.0
424.10	2163.0
424.90	172.0
441.10	27400.0
442.10	130120.0
443.10	34520.0
444.00	2982.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

m/z	Abundance
35.00	412.0
37.00	377.0
38.00	1047.0
39.00	4422.0
40.00	1090.0
41.00	1191.0
42.00	548.0
43.00	1324.0
44.00	7481.0
44.90	467.0
49.00	596.0
50.00	14796.0
51.00	55592.0
52.00	2952.0
53.00	321.0
54.00	288.0
55.00	1209.0
56.00	1870.0
57.00	4411.0
58.00	382.0
59.00	209.0
60.90	664.0
62.00	856.0
63.00	2430.0
64.00	425.0
65.00	1155.0
67.00	598.0
68.00	1165.0
69.00	59504.0
70.00	569.0
70.90	389.0
73.00	1567.0
74.00	6697.0
75.00	10543.0
76.00	3373.0
77.00	74704.0
78.00	5463.0
79.00	4935.0
80.00	3543.0
81.00	5247.0
82.00	1300.0
83.00	1597.0
84.00	523.0
85.00	1161.0
86.00	1602.0
87.00	693.0
87.90	333.0
88.90	275.0
91.00	1557.0
92.00	1252.0
93.00	8497.0
94.00	592.0
95.00	366.0
96.10	892.0
97.10	531.0
98.00	6747.0
99.00	5607.0
99.90	533.0
101.00	2764.0
101.90	163.0
103.00	1148.0
103.90	1814.0
105.00	2080.0
106.00	604.0
107.00	23128.0
108.00	3580.0
109.00	709.0
110.00	40696.0
111.00	6550.0
112.00	991.0
112.90	420.0
114.90	300.0
116.00	1367.0
117.00	19472.0
117.90	1749.0
119.00	496.0
119.90	382.0
121.00	263.0
122.00	1606.0
123.00	2234.0
123.90	1166.0
125.00	1084.0
126.20	459.0
127.00	81880.0
128.00	6119.0
129.00	30320.0
130.00	2714.0
130.90	726.0
132.00	451.0
133.00	673.0
134.00	812.0
135.00	2576.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

136.00	1124.0
137.00	1443.0
138.00	285.0
138.90	210.0
140.00	517.0
141.00	4155.0
142.00	1486.0
142.90	1072.0
144.00	296.0
145.10	230.0
145.90	739.0
147.00	2278.0
148.00	5341.0
149.00	1109.0
149.90	425.0
151.10	653.0
151.90	346.0
153.00	1187.0
154.00	1129.0
155.00	2419.0
156.00	3825.0
157.00	688.0
157.90	700.0
159.00	716.0
160.00	1294.0
161.00	2205.0
162.00	607.0
162.90	311.0
163.90	254.0
165.00	1586.0
166.00	1131.0
167.00	9424.0
168.00	5066.0
169.00	853.0
170.00	270.0
170.80	315.0
172.00	693.0
173.00	981.0
174.00	1689.0
175.00	3336.0
176.00	992.0
176.90	1620.0
178.00	644.0
179.00	5797.0
180.00	3822.0
181.00	2099.0
182.00	335.0
182.90	326.0
184.00	504.0
185.00	3028.0
186.00	23992.0
187.00	6945.0
188.00	856.0
189.00	1564.0
190.10	310.0
190.90	1217.0
192.00	2304.0
193.00	2804.0
194.00	695.0
195.00	411.0
196.00	5472.0
198.00	160320.0
199.00	11283.0
200.00	859.0
201.30	906.0
203.00	1162.0
204.00	6121.0
205.00	10624.0
206.00	44232.0
207.10	10877.0
208.00	2264.0
209.00	1039.0
210.10	822.0
211.00	1904.0
215.00	477.0
216.10	911.0
217.00	11103.0
218.00	1529.0
221.00	7557.0
221.80	2043.0
223.00	2596.0
224.10	23968.0
225.00	6419.0
226.00	621.0
227.00	9190.0
228.00	1436.0
229.00	2319.0
229.90	210.0
230.90	894.0
232.10	175.0
234.00	602.0
235.00	706.0
236.00	505.0
237.00	888.0
238.00	178.0
239.00	413.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

239.90	407.0
241.00	704.0
242.00	1471.0
243.00	1393.0
244.00	18664.0
245.10	2771.0
246.00	3603.0
246.90	717.0
248.00	209.0
249.00	745.0
251.00	183.0
252.00	183.0
253.00	399.0
254.00	605.0
255.00	101384.0
256.00	14977.0
257.10	1043.0
258.00	6097.0
259.00	873.0
260.00	178.0
260.90	220.0
264.00	155.0
265.00	2509.0
266.00	372.0
267.00	314.0
271.00	199.0
271.90	276.0
273.00	3103.0
274.00	7906.0
275.00	38568.0
276.00	6037.0
277.00	4081.0
278.00	691.0
279.00	164.0
281.10	794.0
282.10	289.0
283.10	633.0
284.00	295.0
285.00	720.0
288.90	152.0
292.10	172.0
293.00	970.0
293.90	170.0
295.10	441.0
296.00	13605.0
297.00	1869.0
298.00	178.0
301.00	237.0
302.00	251.0
303.00	1582.0
304.00	371.0
310.10	163.0
314.00	668.0
315.00	1265.0
316.10	727.0
321.00	535.0
321.90	204.0
323.10	4239.0
324.00	702.0
326.90	742.0
328.00	576.0
332.00	227.0
333.00	396.0
334.00	3260.0
335.00	748.0
341.00	606.0
342.00	179.0
346.00	1147.0
347.00	226.0
352.00	1412.0
353.00	974.0
354.00	1345.0
355.00	361.0
365.00	5751.0
366.00	851.0
369.90	173.0
371.00	289.0
372.00	2211.0
373.10	660.0
383.10	603.0
384.10	181.0
390.00	369.0
391.00	226.0
391.90	157.0
401.00	177.0
401.90	1125.0
403.00	1301.0
404.00	494.0
421.10	1447.0
422.00	1087.0
423.00	9640.0
424.00	2045.0
441.10	24640.0
442.10	173760.0
443.10	31144.0
444.10	2888.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247