

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081922\
 Data File : BN021298.D
 Acq On : 19 Aug 2022 17:01
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV237

Quant Time: Aug 22 11:23:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 19 16:13:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.099	152	96198	20.000	ng/u1	0.00
20) Naphthalene-d8	10.934	136	461756	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.751	164	308792	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.504	188	654876	20.000	ng/u1	0.00
79) Chrysene-d12	21.692	240	662002	20.000	ng/u1	0.00
88) Perylene-d12	24.227	264	763724	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	17767	6.934	ng/uL	0.00
4) Pyridine-d5	3.817	84	121351	17.378	ng/u1	0.00
7) Phenol-d5	7.240	99	144778	16.594	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.416	67	97270	18.418	ng/u1	0.00
11) 2-Chlorophenol-d4	7.616	132	113887	17.941	ng/u1	0.00
15) 4-Methylphenol-d8	8.810	113	122447	17.886	ng/u1	0.00
21) Nitrobenzene-d5	9.275	128	59155	18.803	ng/u1	0.00
24) 2-Nitrophenol-d4	10.004	143	53343	18.983	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.546	165	116261	18.682	ng/u1	0.00
31) 4-Chloroaniline-d4	11.063	131	188001	17.788	ng/u1	0.00
46) Dimethylphthalate-d6	14.163	166	384707	18.228	ng/u1	0.00
49) Acenaphthylene-d8	14.445	160	462000	18.776	ng/u1	0.00
54) 4-Nitrophenol-d4	14.939	143	73881	17.591	ng/u1	0.00
60) Fluorene-d10	15.745	176	334893	18.236	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.851	200	43885	16.754	ng/u1	0.00
73) Anthracene-d10	17.598	188	521963	18.086	ng/u1	0.00
81) Pyrene-d10	19.892	212	615133	19.539	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.057	264	666489	18.412	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.417	88	18030	6.833	ng/uL	97
5) Pyridine	3.834	79	118186	16.820	ng/u1	93
6) Benzaldehyde	7.222	77	88609	23.374	ng/u1	92
8) Phenol	7.269	94	155312	17.226	ng/u1	94
10) Bis(2-Chloroethyl)ether	7.510	93	125893	17.512	ng/u1	96
12) 2-Chlorophenol	7.652	128	121594	17.855	ng/u1	97
13) 2-Methylphenol	8.540	108	117864	17.206	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.634	45	161838	17.848	ng/u1	96
16) Acetophenone	8.928	105	203287	18.775	ng/u1	94
17) N-Nitroso-di-n-propyla...	8.910	70	101140	18.760	ng/u1	94
18) 4-Methylphenol	8.869	108	131786	17.513	ng/u1	92
19) Hexachloroethane	9.193	117	48489	17.864	ng/u1	91
22) Nitrobenzene	9.316	77	141354	18.233	ng/u1	95
23) Isophorone	9.840	82	291166	18.239	ng/u1	100
25) 2-Nitrophenol	10.034	139	64511	19.599	ng/u1	90
26) 2,4-Dimethylphenol	10.093	107	146465	17.969	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.334	93	180224	18.027	ng/u1	98
29) 2,4-Dichlorophenol	10.569	162	118858	18.229	ng/u1	99
30) Naphthalene	10.981	128	426453	17.737	ng/u1	99
32) 4-Chloroaniline	11.093	127	195117	17.814	ng/u1	99
33) Hexachlorobutadiene	11.263	225	67418	17.948	ng/u1	97
34) Caprolactam	11.845	113	49899	20.007	ng/u1	86
35) 4-Chloro-3-methylphenol	12.204	107	136344	18.554	ng/u1	98
36) 2-Methylnaphthalene	12.587	142	294631	18.115	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081922\
 Data File : BN021298.D
 Acq On : 19 Aug 2022 17:01
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SICV237


Quant Time: Aug 22 11:23:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 19 16:13:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.804	142	295771	18.094	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	12.951	216	137072	17.311	ng/ul	96
40) Hexachlorocyclopentadiene	12.928	237	94076	19.956	ng/ul	98
41) 2,4,6-Trichlorophenol	13.187	196	89985	18.315	ng/ul	94
42) 2,4,5-Trichlorophenol	13.257	196	99992	18.549	ng/ul	95
43) 1,1'-Biphenyl	13.592	154	405467	17.968	ng/ul	99
44) 2-Chloronaphthalene	13.634	162	303851	17.588	ng/ul	99
45) 2-Nitroaniline	13.834	65	83260	19.215	ng/ul	90
47) Dimethylphthalate	14.204	163	387474	17.948	ng/ul	98
48) 2,6-Dinitrotoluene	14.322	165	78959	21.319	ng/ul	94
50) Acenaphthylene	14.475	152	506538	17.833	ng/ul	98
51) 3-Nitroaniline	14.651	138	90126	21.843	ng/ul	97
52) Acenaphthene	14.816	153	328839	17.917	ng/ul	97
53) 2,4-Dinitrophenol	14.857	184	31642	17.451	ng/ul	98
55) 4-Nitrophenol	14.951	109	63610	18.894	ng/ul	92
56) Dibenzofuran	15.151	168	461891	17.516	ng/ul	98
57) 2,4-Dinitrotoluene	15.110	165	103348	19.213	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.369	232	83130	20.091	ng/ul	97
59) Diethylphthalate	15.569	149	400330	18.388	ng/ul	97
61) Fluorene	15.798	166	383280	18.429	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.792	204	171865	17.926	ng/ul	96
63) 4-Nitroaniline	15.816	138	98851	23.873	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.863	198	47817	17.231	ng/ul#	95
67) N-Nitrosodiphenylamine	16.004	169	338461	18.440	ng/ul	99
68) 4-Bromophenyl-phenylether	16.692	248	105617	18.134	ng/ul	92
69) Hexachlorobenzene	16.798	284	119854	17.287	ng/ul	93
70) Atrazine	16.951	200	115208	17.738	ng/ul	96
71) Pentachlorophenol	17.145	266	74243	19.234	ng/ul	94
72) Phenanthrene	17.545	178	604669	17.480	ng/ul	99
74) Anthracene	17.633	178	605986	17.593	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.551	216	142710	17.122	ng/ul	99
76) Pentachlorobenzene	15.063	250	143313	18.049	ng/ul	92
77) Carbazole	17.904	167	602570	18.512	ng/ul	97
78) Di-n-butylphthalate	18.475	149	663249	18.073	ng/ul	99
80) Fluoranthene	19.557	202	716765	18.875	ng/ul	96
82) Pyrene	19.922	202	729319	18.427	ng/ul#	92
83) Butylbenzylphthalate	20.822	149	286339	19.114	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.604	252	251341	19.912	ng/ul#	90
85) Benzo(a)anthracene	21.674	228	741009	18.591	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.604	149	447269	19.501	ng/ul	99
87) Chrysene	21.727	228	722203	18.714	ng/ul	99
89) Di-n-octyl phthalate	22.580	149	787264	17.598	ng/ul	100
90) Benzo(b)fluoranthene	23.445	252	832520	17.918	ng/ul#	97
91) Benzo(k)fluoranthene	23.498	252	785386	18.896	ng/ul	97
93) Benzo(a)pyrene	24.110	252	767757	17.270	ng/ul#	96
94) Indeno(1,2,3-cd)pyrene	26.904	276	946841	17.349	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.939	278	836206	18.155	ng/ul	98
96) Benzo(g,h,i)perylene	27.727	276	836008	17.554	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 22 11:23:07 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081922.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 19 16:13:51 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.833 ng/uL
RT: 3.417 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion: 88 Resp: 18030
Ion Ratio Lower Upper
88 100
43 26.5 22.4 33.6
58 73.9 61.0 91.4



#5
Pyridine
Concen: 16.820 ng/uL
RT: 3.834 min Scan# 93
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion: 79 Resp: 118186
Ion Ratio Lower Upper
79 100
52 75.5 65.0 97.4
51 34.6 31.6 47.4









#6
Benzaldehyde
Concen: 23.374 ng/ul
RT: 7.222 min Scan# 669
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion: 77 Resp: 88609
Ion Ratio Lower Upper
77 100
105 100.3 74.7 112.1
106 96.9 70.2 105.4



#8
Phenol
Concen: 17.226 ng/ul
RT: 7.269 min Scan# 677
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion: 94 Resp: 155312
Ion Ratio Lower Upper
94 100
65 26.6 23.6 35.4
66 35.3 31.2 46.8









#10
Bis(2-Chloroethyl)ether
Concen: 17.512 ng/ul
RT: 7.510 min Scan# 718
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion: 93 Resp: 125893
Ion Ratio Lower Upper
93 100
63 75.4 63.4 95.0
95 34.2 26.2 39.2



#12
2-Chlorophenol
Concen: 17.855 ng/ul
RT: 7.652 min Scan# 742
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:128 Resp: 121594
Ion Ratio Lower Upper
128 100
64 45.9 38.6 57.8
130 32.3 25.0 37.4









#13
2-Methylphenol
Concen: 17.206 ng/ul
RT: 8.540 min Scan# 893
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:108 Resp: 117864
Ion Ratio Lower Upper
108 100
107 90.3 71.5 107.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 17.848 ng/ul
RT: 8.634 min Scan# 909
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion: 45 Resp: 161838
Ion Ratio Lower Upper
45 100
77 15.3 10.9 16.3
79 12.0 8.6 13.0









#16
 Acetophenone
 Concen: 18.775 ng/ul
 RT: 8.928 min Scan# 91
 Delta R.T. -0.000 min
 Lab File: BN021298.D
 Acq: 19 Aug 2022 17:01




Instrument :
 BNA_N
 ClientSampleId :
 SICV237


Tgt Ion:105 Resp: 203287
 Ion Ratio Lower Upper
 105 100
 77 74.8 63.5 95.3
 51 25.1 23.5 35.3



#17
 N-Nitroso-di-n-propylamine
 Concen: 18.760 ng/ul
 RT: 8.910 min Scan# 956
 Delta R.T. -0.000 min
 Lab File: BN021298.D
 Acq: 19 Aug 2022 17:01

Tgt Ion: 70 Resp: 101140
 Ion Ratio Lower Upper
 70 100
 42 54.1 47.8 71.6
 101 9.8 7.8 11.6
 130 21.7 15.8 23.8









#18
4-Methylphenol
Concen: 17.513 ng/ul
RT: 8.869 min Scan# 949
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:108 Resp: 131786
Ion Ratio Lower Upper
108 100
107 118.3 88.0 132.0



#19
Hexachloroethane
Concen: 17.864 ng/ul
RT: 9.193 min Scan# 1004
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:117 Resp: 48489
Ion Ratio Lower Upper
117 100
201 92.1 83.2 124.8
199 58.2 50.2 75.4









#22
Nitrobenzene
Concen: 18.233 ng/ul
RT: 9.316 min Scan# 1025
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion: 77 Resp: 141354
Ion Ratio Lower Upper
77 100
123 47.1 34.5 51.7
65 12.8 11.1 16.7



#23
Isophorone
Concen: 18.239 ng/ul
RT: 9.840 min Scan# 1114
Delta R.T. -0.006 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion: 82 Resp: 291166
Ion Ratio Lower Upper
82 100
95 6.9 5.8 8.8
138 17.2 13.8 20.8









#25
2-Nitrophenol
Concen: 19.599 ng/ul
RT: 10.034 min Scan# 1147
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:139 Resp: 64511
Ion Ratio Lower Upper
139 100
65 45.6 43.7 65.5
109 28.3 25.8 38.6



#26
2,4-Dimethylphenol
Concen: 17.969 ng/ul
RT: 10.093 min Scan# 1157
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:107 Resp: 146465
Ion Ratio Lower Upper
107 100
121 51.8 38.3 57.5
122 81.1 66.3 99.5









#27
Bis(2-Chloroethoxy)methane
Concen: 18.027 ng/ul
RT: 10.334 min Scan# 1198
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion: 93 Resp: 180224
Ion Ratio Lower Upper
93 100
95 32.5 25.4 38.0
123 11.5 8.3 12.5



#29
2,4-Dichlorophenol
Concen: 18.229 ng/ul
RT: 10.569 min Scan# 1238
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion: 162 Resp: 118858
Ion Ratio Lower Upper
162 100
164 65.5 52.9 79.3
98 36.6 29.8 44.8









#30
Naphthalene
Concen: 17.737 ng/ul
RT: 10.981 min Scan# 1308
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:128 Resp: 426453
Ion Ratio Lower Upper
128 100
129 11.2 8.7 13.1
127 12.7 10.6 16.0



#32
4-Chloroaniline
Concen: 17.814 ng/ul
RT: 11.093 min Scan# 1327
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:127 Resp: 195117
Ion Ratio Lower Upper
127 100
129 33.2 26.1 39.1









#33
Hexachlorobutadiene
Concen: 17.948 ng/ul
RT: 11.263 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:225 Resp: 67418
Ion Ratio Lower Upper
225 100
223 60.7 51.8 77.6
227 63.9 50.6 76.0



#34
Caprolactam
Concen: 20.007 ng/ul
RT: 11.845 min Scan# 1455
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:113 Resp: 49899
Ion Ratio Lower Upper
113 100
55 158.3 145.5 218.3
56 120.3 106.5 159.7









#35
 4-Chloro-3-methylphenol
 Concen: 18.554 ng/ul
 RT: 12.204 min Scan# 1516
 Delta R.T. -0.000 min
 Lab File: BN021298.D
 Acq: 19 Aug 2022 17:01




Instrument :
 BNA_N
 ClientSampleId :
 SICV237


Tgt Ion:107 Resp: 136344
 Ion Ratio Lower Upper
 107 100
 144 27.0 20.6 31.0
 142 80.0 62.2 93.4



#36
 2-Methylnaphthalene
 Concen: 18.115 ng/ul
 RT: 12.587 min Scan# 1581
 Delta R.T. -0.000 min
 Lab File: BN021298.D
 Acq: 19 Aug 2022 17:01

Tgt Ion:142 Resp: 294631
 Ion Ratio Lower Upper
 142 100
 141 91.3 70.6 105.8





#37

1-Methylnaphthalene

Concen: 18.094 ng/ul

RT: 12.804 min Scan# 1618

Delta R.T. -0.000 min

Lab File: BN021298.D

Acq: 19 Aug 2022 17:01

Instrument :

BNA_N

ClientSampleId :

SICV237





Tgt Ion:142 Resp: 295771

Ion Ratio Lower Upper

142 100

141 92.8 70.9 106.3

116 3.6 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 17.311 ng/ul

RT: 12.951 min Scan# 1643

Delta R.T. -0.000 min

Lab File: BN021298.D

Acq: 19 Aug 2022 17:01

Tgt Ion:216 Resp: 137072




Ion Ratio Lower Upper


216 100

214 74.5 63.4 95.0

179 21.3 17.8 26.8

108 17.5 15.1 22.7









#40
Hexachlorocyclopentadiene
Concen: 19.956 ng/ul
RT: 12.928 min Scan# 1639
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:237 Resp: 94076
Ion Ratio Lower Upper
237 100
235 62.3 48.8 73.2
272 12.0 9.3 13.9



#41
2,4,6-Trichlorophenol
Concen: 18.315 ng/ul
RT: 13.187 min Scan# 1683
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:196 Resp: 89985
Ion Ratio Lower Upper
196 100
198 91.5 79.8 119.8
200 31.6 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 18.549 ng/ul
RT: 13.257 min Scan# 1695
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:196	Resp:	99992
Ion Ratio	Lower	Upper
196	100	
198	91.7	77.6 116.4
200	29.1	24.4 36.6



#43
1,1'-Biphenyl
Concen: 17.968 ng/ul
RT: 13.592 min Scan# 1752
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:154	Resp:	405467
Ion Ratio	Lower	Upper
154	100	
153	38.7	31.5 47.3
76	12.2	10.5 15.7









#44
2-Chloronaphthalene
Concen: 17.588 ng/ul
RT: 13.634 min Scan# 1759
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:162 Resp: 303851
Ion Ratio Lower Upper
162 100
164 33.0 25.8 38.6
127 39.3 31.4 47.2



#45
2-Nitroaniline
Concen: 19.215 ng/ul
RT: 13.834 min Scan# 1793
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion: 65 Resp: 83260
Ion Ratio Lower Upper
65 100
92 68.9 54.1 81.1
138 111.0 77.0 115.4









#47
Dimethylphthalate
Concen: 17.948 ng/ul
RT: 14.204 min Scan# 1856
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:163 Resp: 387474
Ion Ratio Lower Upper
163 100
194 3.7 3.4 5.0
164 10.3 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 21.319 ng/ul
RT: 14.322 min Scan# 1876
Delta R.T. -0.006 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:165 Resp: 78959
Ion Ratio Lower Upper
165 100
89 57.9 50.6 75.8
121 21.4 17.9 26.9









#50
Acenaphthylene
Concen: 17.833 ng/ul
RT: 14.475 min Scan# 1902
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:152 Resp: 506538
Ion Ratio Lower Upper
152 100
151 19.1 16.0 24.0
153 12.8 10.7 16.1



#51
3-Nitroaniline
Concen: 21.843 ng/ul
RT: 14.651 min Scan# 1932
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:138 Resp: 90126
Ion Ratio Lower Upper
138 100
108 11.2 8.6 12.8
92 120.1 92.9 139.3









#52
Acenaphthene
Concen: 17.917 ng/ul
RT: 14.816 min Scan# 1960
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:153 Resp: 328839
Ion Ratio Lower Upper
153 100
152 48.0 38.4 57.6
154 91.6 69.7 104.5



#53
2,4-Dinitrophenol
Concen: 17.451 ng/ul
RT: 14.857 min Scan# 1967
Delta R.T. 0.006 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:184 Resp: 31642
Ion Ratio Lower Upper
184 100
63 68.8 55.9 83.9
154 64.7 53.1 79.7









#55
4-Nitrophenol
Concen: 18.894 ng/ul
RT: 14.951 min Scan# 1983
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:109 Resp: 63610
Ion Ratio Lower Upper
109 100
139 140.8 100.1 150.1
65 134.2 104.6 157.0



#56
Dibenzofuran
Concen: 17.516 ng/ul
RT: 15.151 min Scan# 1717
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:168 Resp: 461891
Ion Ratio Lower Upper
168 100
139 37.6 31.1 46.7









#57
2,4-Dinitrotoluene
Concen: 19.213 ng/ul
RT: 15.110 min Scan# 2009
Delta R.T. 0.006 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:165 Resp: 103348
Ion Ratio Lower Upper
165 100
63 41.9 36.3 54.5
182 2.7 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.091 ng/ul
RT: 15.369 min Scan# 2054
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:232 Resp: 83130
Ion Ratio Lower Upper
232 100
131 50.3 38.5 57.7
130 2.5 2.2 3.2
166 33.9 25.8 38.8









#59
Diethylphthalate
Concen: 18.388 ng/ul
RT: 15.569 min Scan# 2088
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:149 Resp: 400330
Ion Ratio Lower Upper
149 100
177 21.0 17.8 26.8
150 11.7 10.5 15.7



#61
Fluorene
Concen: 18.429 ng/ul
RT: 15.798 min Scan# 2127
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:166 Resp: 383280
Ion Ratio Lower Upper
166 100
165 98.0 77.7 116.5
167 13.3 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 17.926 ng/ul
RT: 15.792 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:204 Resp: 171865
Ion Ratio Lower Upper
204 100
206 31.5 26.7 40.1
141 66.3 50.2 75.2



#63
4-Nitroaniline
Concen: 23.873 ng/ul
RT: 15.816 min Scan# 2130
Delta R.T. 0.006 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:138 Resp: 98851
Ion Ratio Lower Upper
138 100
92 49.1 41.0 61.6
108 74.5 65.3 97.9









#66
4,6-Dinitro-2-methylphenol
Concen: 17.231 ng/ul
RT: 15.863 min Scan# 2138
Delta R.T. -0.006 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:198 Resp: 47817
Ion Ratio Lower Upper
198 100
182 4.4 3.0 4.4#
77 27.9 20.1 30.1



#67
N-Nitrosodiphenylamine
Concen: 18.440 ng/ul
RT: 16.004 min Scan# 2162
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:169 Resp: 338461
Ion Ratio Lower Upper
169 100
168 63.9 52.4 78.6
167 34.1 27.5 41.3









#68
4-Bromophenyl-phenylether
Concen: 18.134 ng/ul
RT: 16.692 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:248 Resp: 105617
Ion Ratio Lower Upper
248 100
250 89.1 77.8 116.8
141 70.8 62.6 93.8



#69
Hexachlorobenzene
Concen: 17.287 ng/ul
RT: 16.798 min Scan# 2297
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:284 Resp: 119854
Ion Ratio Lower Upper
284 100
142 40.9 27.8 41.8
249 29.9 25.7 38.5









#70
Atrazine
Concen: 17.738 ng/ul
RT: 16.951 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:200 Resp: 115208
Ion Ratio Lower Upper
200 100
173 29.0 20.5 30.7
215 46.6 38.9 58.3



#71
Pentachlorophenol
Concen: 19.234 ng/ul
RT: 17.145 min Scan# 2356
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:266 Resp: 74243
Ion Ratio Lower Upper
266 100
264 63.9 50.2 75.2
268 69.3 49.3 73.9









#72
Phenanthrene
Concen: 17.480 ng/u1
RT: 17.545 min Scan# 2424
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:178 Resp: 604669
Ion Ratio Lower Upper
178 100
179 14.6 12.2 18.4
176 19.1 15.2 22.8



#74
Anthracene
Concen: 17.593 ng/u1
RT: 17.633 min Scan# 2439
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:178 Resp: 605986
Ion Ratio Lower Upper
178 100
179 16.0 13.0 19.4
176 20.2 15.5 23.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 17.122 ng/uL
RT: 13.551 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:216 Resp: 142710
Ion Ratio Lower Upper
216 100
214 77.7 63.2 94.8
218 47.9 37.9 56.9



#76
Pentachlorobenzene
Concen: 18.049 ng/uL
RT: 15.063 min Scan# 2002
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:250 Resp: 143313
Ion Ratio Lower Upper
250 100
248 61.5 53.0 79.4
252 58.0 53.4 80.2









#77
 Carbazole
 Concen: 18.512 ng/ul
 RT: 17.904 min Scan# 2485
 Delta R.T. -0.000 min
 Lab File: BN021298.D
 Acq: 19 Aug 2022 17:01




Instrument :
 BNA_N
 ClientSampleId :
 SICV237


Tgt Ion:167 Resp: 602570
 Ion Ratio Lower Upper
 167 100
 166 21.8 16.1 24.1
 139 13.4 10.2 15.2



#78
 Di-n-butylphthalate
 Concen: 18.073 ng/ul
 RT: 18.475 min Scan# 2582
 Delta R.T. -0.000 min
 Lab File: BN021298.D
 Acq: 19 Aug 2022 17:01

Tgt Ion:149 Resp: 663249
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.1 10.7
 104 4.7 4.1 6.1









#80
 Fluoranthene
 Concen: 18.875 ng/u1
 RT: 19.557 min Scan# 2766
 Delta R.T. -0.000 min
 Lab File: BN021298.D
 Acq: 19 Aug 2022 17:01




Instrument :
 BNA_N
 ClientSampleId :
 SICV237


Tgt Ion:202 Resp: 716765
 Ion Ratio Lower Upper
 202 100
 101 12.7 8.9 13.3
 100 9.4 6.7 10.1



#82
 Pyrene
 Concen: 18.427 ng/u1
 RT: 19.922 min Scan# 2828
 Delta R.T. -0.000 min
 Lab File: BN021298.D
 Acq: 19 Aug 2022 17:01

Tgt Ion:202 Resp: 729319
 Ion Ratio Lower Upper
 202 100
 101 15.2 9.7 14.5#
 100 12.6 8.0 12.0#





#83

Butylbenzylphthalate

Concen: 19.114 ng/ul

RT: 20.822 min Scan# 2981

Delta R.T. -0.000 min

Lab File: BN021298.D

Acq: 19 Aug 2022 17:01

Instrument :

BNA_N

ClientSampleId :

SICV237





Tgt Ion:149 Resp: 286339

Ion Ratio Lower Upper

149 100

91 70.2 58.8 88.2

206 20.6 18.1 27.1



#84

3,3'-Dichlorobenzidine

Concen: 19.912 ng/ul

RT: 21.604 min Scan# 3114

Delta R.T. -0.006 min

Lab File: BN021298.D

Acq: 19 Aug 2022 17:01




Tgt Ion:252 Resp: 251341


Ion Ratio Lower Upper

252 100

254 71.6 50.8 76.2

126 12.6 8.2 12.4#









#85
Benzo(a)anthracene
Concen: 18.591 ng/ul
RT: 21.674 min Scan# 3127
Delta R.T. -0.006 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:228 Resp: 741009
Ion Ratio Lower Upper
228 100
229 19.4 16.3 24.5
226 26.9 23.2 34.8



#86
Bis(2-ethylhexyl)phthalate
Concen: 19.501 ng/ul
RT: 21.604 min Scan# 3114
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:149 Resp: 447269
Ion Ratio Lower Upper
149 100
167 27.8 22.5 33.7
279 4.1 3.5 5.3









#87
Chrysene
Concen: 18.714 ng/ul
RT: 21.727 min Scan# 3136
Delta R.T. -0.006 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:228 Resp: 722203
Ion Ratio Lower Upper
228 100
226 29.8 24.2 36.2
229 21.1 16.2 24.2



#89
Di-n-octyl phthalate
Concen: 17.598 ng/ul
RT: 22.580 min Scan# 3280
Delta R.T. -0.000 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:149 Resp: 787264





#90

Benzo(b)fluoranthene

Concen: 17.918 ng/ul

RT: 23.445 min Scan# 3427

Delta R.T. -0.006 min

Lab File: BN021298.D

Acq: 19 Aug 2022 17:01

Instrument :

BNA_N

ClientSampleId :

SICV237





Tgt Ion:252 Resp: 832520

Ion Ratio Lower Upper

252 100

253 21.0 16.5 24.7

125 12.7 7.8 11.8#



#91

Benzo(k)fluoranthene

Concen: 18.896 ng/ul

RT: 23.498 min Scan# 3436

Delta R.T. -0.006 min

Lab File: BN021298.D

Acq: 19 Aug 2022 17:01




Tgt Ion:252 Resp: 785386


Ion Ratio Lower Upper

252 100

253 21.8 18.2 27.4

125 12.0 8.2 12.4









#93
Benzo(a)pyrene
Concen: 17.270 ng/ul
RT: 24.110 min Scan# 3540
Delta R.T. -0.012 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01




Instrument :
BNA_N
ClientSampleId :
SICV237


Tgt Ion:252 Resp: 767757
Ion Ratio Lower Upper
252 100
253 22.0 18.7 28.1
125 14.0 9.2 13.8#



#94
Indeno(1,2,3-cd)pyrene
Concen: 17.349 ng/ul
RT: 26.904 min Scan# 4015
Delta R.T. -0.006 min
Lab File: BN021298.D
Acq: 19 Aug 2022 17:01

Tgt Ion:276 Resp: 946841
Ion Ratio Lower Upper
276 100
138 25.5 19.5 29.3
277 20.9 21.4 32.0#









#95
 Dibenzo(a,h)anthracene
 Concen: 18.155 ng/ul
 RT: 26.939 min Scan# 4021
 Delta R.T. -0.006 min
 Lab File: BN021298.D
 Acq: 19 Aug 2022 17:01

Instrument :
 BNA_N
 ClientSampleId :
 SICV237

Tgt Ion:278 Resp: 836206
 Ion Ratio Lower Upper
 278 100
 139 18.1 13.8 20.8
 279 22.0 18.3 27.5



#96
 Benzo(g,h,i)perylene
 Concen: 17.554 ng/ul
 RT: 27.727 min Scan# 4155
 Delta R.T. -0.012 min
 Lab File: BN021298.D
 Acq: 19 Aug 2022 17:01

Tgt Ion:276 Resp: 836008
 Ion Ratio Lower Upper
 276 100
 138 24.7 17.0 25.6
 277 22.8 17.8 26.6

