

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081924\
 Data File : BN033500.D
 Acq On : 20 Aug 2024 11:35
 Operator : MA/JU
 Sample : P3651-01
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP201-20240815

Quant Time: Aug 20 12:19:34 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 19 23:32:18 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.552	152	7417	0.400	ng	0.00
7) Naphthalene-d8	10.314	136	19845	0.400	ng	# 0.00
13) Acenaphthene-d10	14.189	164	10125	0.400	ng	0.00
19) Phenanthrene-d10	16.942	188	21657	0.400	ng	0.00
29) Chrysene-d12	21.148	240	12812	0.400	ng	0.00
35) Perylene-d12	23.320	264	7238	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.183	112	2837	0.120	ng	0.00
5) Phenol-d6	6.736	99	2532	0.090	ng	0.00
8) Nitrobenzene-d5	8.681	82	5033	0.306	ng	-0.01
11) 2-Methylnaphthalene-d10	11.911	152	8990	0.317	ng	0.00
14) 2,4,6-Tribromophenol	15.688	330	1049	0.193	ng	0.00
15) 2-Fluorobiphenyl	12.809	172	13951	0.337	ng	0.00
27) Fluoranthene-d10	18.979	212	20141	0.387	ng	0.00
31) Terphenyl-d14	19.593	244	16479	0.566	ng	0.00
Target Compounds						
						Qvalue
32) Benzo(a)anthracene	21.130	228	1411	0.030	ng	# 91
33) Chrysene	21.184	228	2167	0.047	ng	93
34) Bis(2-ethylhexyl)phtha...	21.094	149	4185	0.143	ng	94
36) Indeno(1,2,3-cd)pyrene	25.478	276	1751	0.058	ng	99
37) Benzo(b)fluoranthene	22.677	252	2016	0.075	ng	# 36
38) Benzo(k)fluoranthene	22.718	252	1933	0.073	ng	# 34
40) Dibenzo(a,h)anthracene	25.493	278	1381	0.057	ng	# 52
41) Benzo(g,h,i)perylene	26.124	276	1512	0.059	ng	# 71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081924\
Data File : BN033500.D
Acq On : 20 Aug 2024 11:35
Operator : MA/JU
Sample : P3651-01
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20240815

Quant Time: Aug 20 12:19:34 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 19 23:32:18 2024
Response via : Initial Calibration

