

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081924\
 Data File : BN033504.D
 Acq On : 20 Aug 2024 13:59
 Operator : MA/JU
 Sample : SP6605
 Misc : 8270 SIM SURROGATE
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP6605

Quant Time: Aug 20 14:27:48 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 19 23:32:18 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.552	152	9694	0.400	ng	0.00
7) Naphthalene-d8	10.314	136	24835	0.400	ng	# 0.00
13) Acenaphthene-d10	14.189	164	11184	0.400	ng	0.00
19) Phenanthrene-d10	16.942	188	20155	0.400	ng	0.00
29) Chrysene-d12	21.148	240	11312	0.400	ng	0.00
35) Perylene-d12	23.317	264	12280	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.183	112	9364	0.304	ng	0.00
5) Phenol-d6	6.736	99	10782	0.294	ng	0.00
8) Nitrobenzene-d5	8.681	82	7689	0.373	ng	-0.01
11) 2-Methylnaphthalene-d10	11.911	152	12994	0.366	ng	0.00
14) 2,4,6-Tribromophenol	15.688	330	1137	0.189	ng	0.00
15) 2-Fluorobiphenyl	12.809	172	18679	0.409	ng	0.00
27) Fluoranthene-d10	18.979	212	16513	0.341	ng	0.00
31) Terphenyl-d14	19.593	244	10909	0.424	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081924\
Data File : BN033504.D
Acq On : 20 Aug 2024 13:59
Operator : MA/JU
Sample : SP6605
Misc : 8270 SIM SURROGATE
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP6605

Quant Time: Aug 20 14:27:48 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 19 23:32:18 2024
Response via : Initial Calibration

