




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081924\
Data File : BN033488.D
Acq On : 20 Aug 2024 04:04
Operator : MA/JU
Sample : DFTTP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 20 04:49:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 20 04:49:48 2024
Response via : Initial Calibration






TIC: BN033488.D\data.ms

(70) Pentachlorophenol (C)**16.598min (0.000) 23305.39 ng****response 721703**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.39
264.00	61.60	61.61
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081924\
Data File : BN033488.D
Acq On : 20 Aug 2024 04:04
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 20 04:49:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 20 04:49:48 2024
Response via : Initial Calibration



TIC: BN033488.D\data.ms

(77) Benzidine

19.210min (0.000) 0.00 ng

response 4312132

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.14
183.00	13.20	11.75
0.00	0.00	0.00