



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081924\
Data File : BN033488.D
Acq On : 20 Aug 2024 04:04
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081924.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Aug 19 23:32:18 2024






AutoFind: Scans 2453, 2454, 2455; Background Corrected with Scan 2445

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.7	225531	PASS
68	69	0.00	2	1.7	3948	PASS
69	198	0.00	100	38.0	233941	PASS
70	69	0.00	2	0.5	1218	PASS
127	198	10	80	48.9	300736	PASS
197	198	0.00	2	0.5	3029	PASS
198	198	100	100	100.0	615275	PASS
199	198	5	9	6.8	41555	PASS
275	198	10	60	26.9	165448	PASS
365	198	1	100	3.8	23208	PASS
441	198	0.01	100	10.6	65525	PASS
442	442	50	100	100.0	436715	PASS
443	442	15	24	17.9	78141	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081924\
Data File : BN033488.D
Acq On : 20 Aug 2024 04:04
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Aug 20 04:49:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 20 04:49:48 2024
Response via : Initial Calibration



TIC: BN033488.D\data.ms

(70) Pentachlorophenol (C)

16.598min (0.000) 23305.39 ng




response 721703

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.39
264.00	61.60	61.61
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081924\
Data File : BN033488.D
Acq On : 20 Aug 2024 04:04
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Aug 20 04:49:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 20 04:49:48 2024
Response via : Initial Calibration



TIC: BN033488.D\data.ms

(77) Benzidine

19.210min (0.000) 0.00 ng

response 4312132

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.14
183.00	13.20	11.75
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/19/2024	BNA_N	<u>BN033488.D</u>
Compound Name	Response	Retention Time
DDT	2112901	20.457
DDD	33939	20.01
DDE	827	19.504
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
34766	2147667	1.62

Instrument :
BNA_N
ClientSampleId :
DFTPP