

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081925\
 Data File : BN037621.D
 Acq On : 20 Aug 2025 06:03
 Operator : RC/JU
 Sample : Q2842-02DL 10X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RMW-02B-66.3-081225DL

Quant Time: Aug 20 06:52:31 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 13 05:00:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.710	152	1938	0.400	ng	0.00
7) Naphthalene-d8	10.498	136	4467	0.400	ng	0.00
13) Acenaphthene-d10	14.345	164	2275	0.400	ng	0.00
19) Phenanthrene-d10	17.086	188	4809	0.400	ng	0.00
29) Chrysene-d12	21.268	240	3968	0.400	ng	# 0.00
35) Perylene-d12	23.501	264	3640	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.305	112	72	0.016	ng	0.00
5) Phenol-d6	6.879	99	50	0.009	ng	0.00
8) Nitrobenzene-d5	8.854	82	113	0.036	ng	0.00
11) 2-Methylnaphthalene-d10	12.090	152	192	0.032	ng	0.00
14) 2,4,6-Tribromophenol	15.833	330	41	0.041	ng	0.00
15) 2-Fluorobiphenyl	12.973	172	478	0.036	ng	0.00
27) Fluoranthene-d10	19.118	212	570	0.045	ng	0.00
31) Terphenyl-d14	19.717	244	449	0.055	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.232	88	4850	2.616	ng	98
34) Bis(2-ethylhexyl)phtha...	21.196	149	315	0.058	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081925\
Data File : BN037621.D
Acq On : 20 Aug 2025 06:03
Operator : RC/JU
Sample : Q2842-02DL 10X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RMW-02B-66.3-081225DL

Quant Time: Aug 20 06:52:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 13 05:00:58 2025
Response via : Initial Calibration

