


Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN081925\  
Data File : BN037617.D  
Acq On : 20 Aug 2025 03:35  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN081225.M  
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
Last Update : Wed Aug 13 05:00:58 2025



AutoFind: Scans 2457, 2458, 2459; Background Corrected with Scan 2451

| Target | Rel. to | Lower  | Upper  | Rel.  | Raw    | Result    |
|--------|---------|--------|--------|-------|--------|-----------|
| Mass   | Mass    | Limit% | Limit% | Abn%  | Abn    | Pass/Fail |
| 68     | 69      | 0.00   | 2      | 0.9   | 485    | PASS      |
| 69     | 69      | 100    | 100    | 100.0 | 53448  | PASS      |
| 70     | 69      | 0.00   | 2      | 0.6   | 308    | PASS      |
| 197    | 198     | 0.00   | 2      | 0.0   | 0      | PASS      |
| 198    | 198     | 100    | 100    | 100.0 | 136293 | PASS      |
| 199    | 198     | 5      | 9      | 6.8   | 9235   | PASS      |
| 365    | 198     | 1      | 100    | 3.6   | 4939   | PASS      |
| 441    | 443     | 0.01   | 150    | 79.0  | 16038  | PASS      |
| 442    | 442     | 100    | 100    | 100.0 | 106744 | PASS      |
| 443    | 442     | 15     | 24     | 19.0  | 20305  | PASS      |

# DDT Breakdown




| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 8/19/2025     | BNA_N            | <u>BN037617.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 466043           | 20.575             |
| DDD           | 8008             | 20.133             |
| DDE           | 521              | 19.628             |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 8529          | 474572           | 1.80               |
|               |                  |                    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN081925\  
Data File : BN037617.D  
Acq On : 20 Aug 2025 03:35  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: Aug 20 07:01:46 2025  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Aug 06 01:49:15 2025  
Response via : Initial Calibration



TIC: BN037617.D\data.ms

**(70) Pentachlorophenol (C)**

16.739min (-0.006) 24577.65 ng




response 150060

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 64.52  |
| 264.00 | 61.60  | 64.16  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN081925\  
Data File : BN037617.D  
Acq On : 20 Aug 2025 03:35  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: Aug 20 07:01:46 2025  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Aug 06 01:49:15 2025  
Response via : Initial Calibration



TIC: BN037617.D\data.ms

(77) Benzidine

19.328min (-0.006) 0.00 ng

response 801660

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 14.02  |
| 183.00 | 13.20  | 12.34  |
| 0.00   | 0.00   | 0.00   |