

Data File : BN007271.D
 Acq On : 20 Aug 2019 19:03
 Operator : HP/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN082019

Quant Time: Aug 20 19:38:18 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 20 18:52:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.568	0.576	-1.4	97	0.00
3	n-Nitrosodimethylamine	0.380	0.366	3.7	96	0.00
4 S	2-Fluorophenol	1.219	1.116	8.4	95	0.00
5 S	Phenol-d6	1.567	1.304	16.8	93	0.00
6	bis(2-Chloroethyl)ether	1.181	1.210	-2.5	96	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.307	0.306	0.3	95	0.00
9	Naphthalene	1.134	1.154	-1.8	94	0.00
10	Hexachlorobutadiene	0.250	0.260	-4.0	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.691	0.706	-2.2	96	0.00
12	2-Methylnaphthalene	0.785	0.801	-2.0	96	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.225	0.207	8.0	95	0.00
15 S	2-Fluorobiphenyl	1.715	1.780	-3.8	96	0.00
16	Acenaphthylene	2.121	2.091	1.4	96	0.00
17	Acenaphthene	1.352	1.378	-1.9	96	0.00
18	Fluorene	1.766	1.821	-3.1	98	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4-Bromophenyl-phenylether	0.229	0.248	-8.3	125	0.00
21	Hexachlorobenzene	0.265	0.273	-3.0	97	0.00
22	Pentachlorophenol	0.106	0.101	4.7	101	0.00
23	Phenanthrene	1.254	1.270	-1.3	96	0.00
24	Anthracene	1.129	1.119	0.9	97	0.00
25 SURR	Fluoranthene-d10	1.268	1.265	0.2	96	0.00
26	Fluoranthene	1.519	1.522	-0.2	96	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
28	Pyrene	1.496	1.548	-3.5	97	0.00
29 S	Terphenyl-d14	1.001	1.035	-3.4	97	0.00
30	Benzo(a)anthracene	1.487	1.514	-1.8	98	0.00
31	Chrysene	1.518	1.614	-6.3	100	0.00
32	Bis(2-ethylhexyl)phthalate	0.594	0.606	-2.0	111	0.00
33	Indeno(1,2,3-cd)pyrene	1.826	1.913	-4.8	98	0.00
34 I	Perylene-d12	1.000	1.000	0.0	98	0.00
35	Benzo(b)fluoranthene	1.491	1.512	-1.4	98	0.00
36	Benzo(k)fluoranthene	1.461	1.480	-1.3	96	0.00
37 C	Benzo(a)pyrene	1.344	1.338	0.4	97	0.00
38	Dibenzo(a,h)anthracene	1.412	1.444	-2.3	97	0.00
39	Benzo(g,h,i)perylene	1.427	1.461	-2.4	97	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN082019\
Evaluate Continuing Calibration Report

Data File : BN007271.D
Acq On : 20 Aug 2019 19:03
Operator : HP/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN082019

Quant Time: Aug 20 19:38:18 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 20 18:52:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			