

Data File : BN007276.D
Acq On : 21 Aug 2019 03:01
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 21 05:48:35 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 20 18:52:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.568	0.540	4.9	102	0.00
3	n-Nitrosodimethylamine	0.380	0.352	7.4	103	0.00
4 S	2-Fluorophenol	1.219	1.136	6.8	108	0.00
5 S	Phenol-d6	1.567	1.379	12.0	110	0.00
6	bis(2-Chloroethyl)ether	1.181	1.229	-4.1	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
8 S	Nitrobenzene-d5	0.307	0.309	-0.7	110	0.00
9	Naphthalene	1.134	1.168	-3.0	109	0.00
10	Hexachlorobutadiene	0.250	0.254	-1.6	107	0.00
11 SURR	2-Methylnaphthalene-d10	0.691	0.711	-2.9	111	0.00
12	2-Methylnaphthalene	0.785	0.820	-4.5	113	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
14 S	2,4,6-Tribromophenol	0.225	0.209	7.1	114	0.00
15 S	2-Fluorobiphenyl	1.715	1.741	-1.5	111	0.00
16	Acenaphthylene	2.121	2.127	-0.3	116	0.00
17	Acenaphthene	1.352	1.349	0.2	111	0.00
18	Fluorene	1.766	1.762	0.2	112	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
20	4-Bromophenyl-phenylether	0.229	0.242	-5.7	141	0.00
21	Hexachlorobenzene	0.265	0.272	-2.6	112	0.00
22	Pentachlorophenol	0.106	0.102	3.8	119	0.00
23	Phenanthrene	1.254	1.279	-2.0	112	0.00
24	Anthracene	1.129	1.137	-0.7	114	0.00
25 SURR	Fluoranthene-d10	1.268	1.268	0.0	112	0.00
26	Fluoranthene	1.519	1.524	-0.3	112	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
28	Pyrene	1.496	1.571	-5.0	114	0.00
29 S	Terphenyl-d14	1.001	1.028	-2.7	111	0.00
30	Benzo(a)anthracene	1.487	1.535	-3.2	115	-0.01
31	Chrysene	1.518	1.582	-4.2	113	0.00
32	Bis(2-ethylhexyl)phthalate	0.594	0.599	-0.8	127	0.00
33	Indeno(1,2,3-cd)pyrene	1.826	1.953	-7.0	115	0.00
34 I	Perylene-d12	1.000	1.000	0.0	117	0.00
35	Benzo(b)fluoranthene	1.491	1.420	4.8	110	0.00
36	Benzo(k)fluoranthene	1.461	1.470	-0.6	113	0.00
37 C	Benzo(a)pyrene	1.344	1.334	0.7	116	0.00
38	Dibenzo(a,h)anthracene	1.412	1.413	-0.1	114	0.00
39	Benzo(g,h,i)perylene	1.427	1.390	2.6	111	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN082019\
Evaluate Continuing Calibration Report

Data File : BN007276.D
Acq On : 21 Aug 2019 03:01
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 21 05:48:35 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 20 18:52:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			