

Data File : BN007282.D
Acq On : 21 Aug 2019 06:38
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 21 07:10:23 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 20 18:52:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	116	0.00
2	1,4-Dioxane	0.400	0.391	2.3	113	0.00
3	n-Nitrosodimethylamine	0.400	0.376	6.0	113	0.00
4 S	2-Fluorophenol	0.400	0.382	4.5	120	0.00
5 S	Phenol-d6	0.400	0.406	-1.5	123	0.00
6	bis(2-Chloroethyl)ether	0.400	0.410	-2.5	116	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	125	0.00
8 S	Nitrobenzene-d5	0.400	0.402	-0.5	123	0.00
9	Naphthalene	0.400	0.412	-3.0	123	0.00
10	Hexachlorobutadiene	0.400	0.404	-1.0	120	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.404	-1.0	123	0.00
12	2-Methylnaphthalene	0.400	0.412	-3.0	126	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	127	0.00
14 S	2,4,6-Tribromophenol	0.400	0.367	8.3	124	0.00
15 S	2-Fluorobiphenyl	0.400	0.406	-1.5	123	0.00
16	Acenaphthylene	0.400	0.416	-4.0	133	0.00
17	Acenaphthene	0.400	0.402	-0.5	124	0.00
18	Fluorene	0.400	0.404	-1.0	126	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	123	0.00
20	4-Bromophenyl-phenylether	0.400	0.424	-6.0	155	0.00
21	Hexachlorobenzene	0.400	0.417	-4.2	124	0.00
22	Pentachlorophenol	0.400	0.393	1.8	133	0.00
23	Phenanthrene	0.400	0.409	-2.2	124	0.00
24	Anthracene	0.400	0.405	-1.3	125	0.00
25 SURR	Fluoranthene-d10	0.400	0.397	0.8	122	0.00
26	Fluoranthene	0.400	0.399	0.3	122	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	120	0.00
28	Pyrene	0.400	0.424	-6.0	124	0.00
29 S	Terphenyl-d14	0.400	0.414	-3.5	121	0.00
30	Benzo(a)anthracene	0.400	0.413	-3.2	124	0.00
31	Chrysene	0.400	0.417	-4.2	122	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.437	-9.2	148	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.441	-10.2	128	0.00
34 I	Perylene-d12	0.400	0.400	0.0	130	0.00
35	Benzo(b)fluoranthene	0.400	0.381	4.8	122	0.00
36	Benzo(k)fluoranthene	0.400	0.397	0.8	123	0.00
37 C	Benzo(a)pyrene	0.400	0.398	0.5	129	0.00
38	Dibenzo(a,h)anthracene	0.400	0.403	-0.8	127	0.00
39	Benzo(g,h,i)perylene	0.400	0.392	2.0	123	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN082019\
Evaluate Continuing Calibration Report

Data File : BN007282.D
Acq On : 21 Aug 2019 06:38
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 21 07:10:23 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 20 18:52:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				